



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/31/2022

OVENTEMP IN Celsius(°C): 108
Time IN: 14:25
In Date: 10/28/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108
Time OUT: 07:35
Out Date: 10/29/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122658

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N5325-01	EY796	1	0.97	8.92	9.89	8.66	86.2	
N5325-02	EY797	2	0.99	8.81	9.8	8.62	86.6	
N5325-03	EY798	3	0.97	8.94	9.91	8.96	89.4	
N5325-04	EY799	4	0.97	8.99	9.96	9.04	89.8	
N5325-05	EY7A0	5	0.96	8.91	9.87	8.77	87.7	
N5325-06	EY7A1	6	0.98	8.85	9.83	9.03	91.0	
N5325-07	EY7A2	7	0.94	9.02	9.96	8.99	89.2	
N5325-08	EY7A3	8	0.97	8.97	9.94	8.72	86.4	
N5325-09	EY7A3D	9	0.97	8.97	9.94	8.72	86.4	
N5325-10	EY7A3S	10	0.97	8.97	9.94	8.72	86.4	
N5325-11	EY7A4	11	0.96	9.02	9.98	8.86	87.6	
N5325-12	EY7A5	12	0.93	8.97	9.9	8.46	83.9	
N5325-13	EY7A6	13	0.94	8.97	9.91	8.51	84.4	
N5325-14	EY7A7	14	0.93	9.03	9.96	8.57	84.6	
N5325-15	EY7A8	15	0.91	9.04	9.95	8.78	87.1	
N5325-16	EY7A9	16	0.94	9.03	9.97	8.46	83.3	
N5325-17	EY7B0	17	0.94	8.93	9.87	8.35	83.0	
N5325-18	EY7B1	18	0.96	8.96	9.92	8.79	87.4	
N5325-19	EY7B2	19	0.92	9.00	9.92	8.65	85.9	
N5325-20	EY7B3	20	0.95	8.99	9.94	8.57	84.8	
N5325-21	EY7B4	21	0.93	8.98	9.91	8.34	82.5	
N5325-22	EY7B5	22	0.95	9.01	9.96	8.39	82.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$