



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122660

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	10/28/2022	13:00
2	N5317-06		1	15	46.00	114.8	114.1	10/28/2022	13:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 122660

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Flash-PT-1028

WorkList ID : 164479

Department : Wet-Chemistry

Date : 10-28-2022 12:53:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/07/2022	Solid	N5317-06	Flash Point	Cool 4 deg C	CHEM02	QA Of	PT-FP-SOIL	10/24/2022	1010B

Date/Time 10/28/22 13:00
Raw Sample Received by: JP (wc)
Raw Sample Relinquished by: SL (QA-office)

Date/Time 10/28/22 15:00
Raw Sample Received by: SL (QA-office)
Raw Sample Relinquished by: JP (wc)