



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/31/2022

OVENTEMP IN Celsius(°C): 106
Time IN: 14:00
In Date: 10/29/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 10/30/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122669

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N5364-01	EY7D4	1	0.92	8.79	9.71	8.5	86.2	
N5364-02	EY7D5	2	0.91	8.70	9.61	8.9	91.8	
N5364-03	EY7D6	3	0.92	8.73	9.65	8.81	90.4	
N5364-04	EY7D7	4	0.92	8.80	9.72	8.91	90.8	
N5364-05	EY7D8	5	0.91	8.80	9.71	8.98	91.7	
N5364-06	EY7D9	6	0.9	8.84	9.74	8.9	90.5	
N5364-07	EY7E0	7	0.92	9.07	9.99	9.25	91.8	
N5364-08	EY7E1	8	0.91	8.91	9.82	8.87	89.3	
N5364-09	EY7E2	9	0.91	8.82	9.73	8.63	87.5	
N5364-10	EY7E3	10	0.92	8.87	9.79	8.68	87.5	
N5364-11	EY7E4	11	0.91	8.72	9.63	8.57	87.8	
N5364-12	EY7E5	12	0.93	8.75	9.68	8.9	91.1	
N5364-13	EY7E6	13	0.92	8.94	9.86	8.7	87.0	
N5364-14	EY7E7	14	0.9	8.86	9.76	8.87	90.0	
N5364-15	EY7E8	15	0.9	8.78	9.68	8.79	89.9	
N5364-16	EY7E9	16	0.91	8.74	9.65	8.77	89.9	
N5364-17	EY7F0	17	0.92	9.06	9.98	9.18	91.2	
N5364-18	EY7F1	18	0.91	8.87	9.78	8.68	87.6	
N5364-19	EY7F2	19	0.91	8.75	9.66	8.38	85.4	
N5364-20	EY7F6	20	0.91	8.96	9.87	8.97	90.0	
N5364-21	EY7F6D	21	0.91	8.96	9.87	8.97	90.0	
N5364-22	EY7F6S	22	0.91	8.96	9.87	8.97	90.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$