



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/31/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 15:10
In Date: 10/29/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 2.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:40
Out Date: 10/30/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122674

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N5370-01	EY7N6	1	0.88	9.04	9.92	9.02	90.0	
N5370-02	EY7N8	2	0.93	8.98	9.91	8.72	86.7	
N5370-03	EY7N9	3	0.93	9.03	9.96	8.84	87.6	
N5370-04	EY7P0	4	0.89	8.92	9.81	9.03	91.3	
N5370-05	EY7P1	5	0.87	9.11	9.98	8.97	88.9	
N5370-06	EY7P2	6	0.91	9.07	9.98	8.29	81.4	
N5370-07	EY7P3	7	0.91	8.95	9.86	8.54	85.3	
N5370-08	EY7P4	8	0.9	9.08	9.98	8.62	85.0	
N5370-09	EY7P5	9	0.92	8.99	9.91	8.36	82.8	
N5370-10	EY7P6	10	0.9	9.07	9.97	8.92	88.4	
N5370-11	EY7P7	11	0.9	9.05	9.95	8.73	86.5	
N5370-12	EY7P8	12	0.93	8.89	9.82	8.64	86.7	
N5370-13	EY7P9	13	0.89	9.07	9.96	8.7	86.1	
N5370-14	EY7Q0	14	0.91	9.01	9.92	8.83	87.9	
N5370-15	EY7Q1	15	0.89	9.05	9.94	8.85	88.0	
N5370-16	EY7Q2	16	0.93	8.89	9.82	8.65	86.8	
N5370-17	EY7Q3	17	0.91	9.05	9.96	8.95	88.8	
N5370-18	EY7Q4	18	0.92	9.03	9.95	8.7	86.2	
N5370-19	EY7Q5	19	0.92	9.02	9.94	8.54	84.5	
N5370-20	EY7R1	20	0.91	8.98	9.89	8.47	84.2	
N5370-21	EY7R1D	21	0.91	8.98	9.89	8.47	84.2	
N5370-22	EY7R1S	22	0.91	8.98	9.89	8.47	84.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$