



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB122679

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/30/2022	07:15
2	CAL2	1	Water	NA	NA	20.7	7.00	10/30/2022	07:16
3	CAL3	1	Water	NA	NA	20.7	10.03	10/30/2022	07:17
4	ICV	1	Water	NA	NA	20.8	6.98	10/30/2022	07:20
5	CCV1	1	Water	NA	NA	20.7	2.01	10/30/2022	07:22
6	N5377-02	1	Water	NA	NA	20.2	9.36	10/30/2022	07:50
7	N5377-02DUP	1	Water	NA	NA	20.3	9.37	10/30/2022	07:51
8	CCV2	1	Water	NA	NA	20.8	12.03	10/30/2022	08:00

WORKLIST(Hardcopy Internal Chain)

122679

WorkList Name : CORROSION n5377 w WorkList ID : 164523 Department : Wet-Chemistry Date : 10-30-2022 07:04:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/28/2022	Water	N5377-02	Corrosivity	Cool 4 deg C	PARS02	F11	LIQUID-WC-20221028	10/28/2022	9040C

Date/Time 10/30/22 07:11
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: JDC

Date/Time 10/30/22 09:20
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: JDC