



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB122680

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/30/2022	07:15
2	CAL2	1	Water	NA	NA	20.7	7.00	10/30/2022	07:16
3	CAL3	1	Water	NA	NA	20.7	10.03	10/30/2022	07:17
4	ICV	1	Water	NA	NA	20.8	6.98	10/30/2022	07:20
5	CCV1	1	Water	NA	NA	20.7	2.03	10/30/2022	08:10
6	N5352-01	1	Water	NA	NA	20.1	8.30	10/30/2022	08:15
7	N5355-01	1	Water	NA	NA	20.4	7.56	10/30/2022	08:20
8	N5355-02	1	Water	NA	NA	20.5	7.57	10/30/2022	08:22
9	N5357-01	1	Water	NA	NA	20.3	7.82	10/30/2022	08:25
10	N5357-02	1	Water	NA	NA	20.7	7.79	10/30/2022	08:26
11	N5357-02DUP	1	Water	NA	NA	20.8	7.80	10/30/2022	08:27
12	CCV2	1	Water	NA	NA	20.7	12.03	10/30/2022	08:30

WORKLIST(Hardcopy Internal Chain)

VB 122680

WorkList Name : ph n5333

WorkList ID : 164524

Department : Wet-Chemistry

Date : 10-30-2022 07:05:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/28/2022	Water	N5352-01	pH A	Cool 4 deg C	PSEG03	F11	CMS-WATERPIT-1028	10/28/2022	9040C
10/27/2022	Water	N5355-01	pH E	Cool 4 deg C	ENV170	F11	5518-1	10/27/2022	9040C
10/27/2022	Water	N5355-02	pH E	Cool 4 deg C	ENV170	F11	5518-2	10/27/2022	9040C
10/27/2022	Water	N5357-01	pH E	Cool 4 deg C	ENV170	F11	5051-1	10/27/2022	9040C
10/27/2022	Water	N5357-02	pH E	Cool 4 deg C	ENV170	F11	5051-2	10/27/2022	9040C

Date/Time 10/30/22 07:20
 Raw Sample Received by: JOC
 Raw Sample Relinquished by: RS (Ext 646)

Date/Time 10/30/22 10:00
 Raw Sample Received by: RS (Ext-646)
 Raw Sample Relinquished by: JOC