

LB122682


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43962



A= 0.0000e+000

B= 2.2812e-004 *slope*C= 1.5400e-002 *intercept*

Rho= 0.9999652

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.020	0.020	21	0.000	21					=
0.05	0.050					185	0				LB
0.20	0.200	0.225	0.025	920	0.0 %	920					-1
2.50	2.500	2.467	-0.033	10749	0.0 %	10749					0
5.00	5.000	4.977	-0.023	21750	0.0 %	21750					0
7.50	7.500	7.466	-0.034	32663	0.0 %	32663					0
10.0	10.000	10.044	0.044	43962	0.0 %	43962					0

LB122682

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	21	-		0 SFAM01.1	PPB	10/30/2022 14:13	Std
	0.2 S01	920	-		0.2 SFAM01.1	PPB	10/30/2022 14:18	Std
	2.5 S02	10749	-		2.5 SFAM01.1	PPB	10/30/2022 14:21	Std
	5 S03	21750	-		5 SFAM01.1	PPB	10/30/2022 14:23	Std
	7.5 S04	32663	-		7.5 SFAM01.1	PPB	10/30/2022 14:25	Std
	10 S05	43962	-		10 SFAM01.1	PPB	10/30/2022 14:30	Std
ICV096	ICV096	16877	3.8653 -		SFAM01.1	PPB	10/30/2022 14:33	SMPL
ICB096	ICB096	90	0.0359 -		SFAM01.1	PPB	10/30/2022 14:35	SMPL
CCV098	CCV098	21850	4.9997 -		SFAM01.1	PPB	10/30/2022 14:37	SMPL
CCB098	CCB098	-41	0.006 -		SFAM01.1	PPB	10/30/2022 14:40	SMPL
PB148678BL	PBW678	-6	0.014 -		SFAM01.1	PPB	10/30/2022 14:42	SMPL
N5138-07	X3C98	4255	0.986 -		SFAM01.1	PPB	10/30/2022 14:44	SMPL
N5138-08	X3C98D	4326	1.0022 -		SFAM01.1	PPB	10/30/2022 14:49	SMPL
N5138-09	X3C98S	8630	1.984 -		SFAM01.1	PPB	10/30/2022 14:53	SMPL
CCV099	CCV099	21851	5 -		SFAM01.1	PPB	10/30/2022 14:58	SMPL
CCB099	CCB099	-17	0.0115 -		SFAM01.1	PPB	10/30/2022 15:00	SMPL