

LB122683


SFAM01.1

INSTRUMENT ID: CVL

Linear

 μ Abs.:

45269



A= 0.0000e+000

B= 2.2150e-004 slope

C= -6.3494e-002 intercept

Rho= 0.9998265

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	W/D
0.00	0.000	0.018	0.018	368	0.000	368					1
0.05	0.050					185					1
0.20	0.200	0.169	-0.031	1049	0.0 %	1049	0				16
2.50	2.500	2.434	-0.066	11276	0.0 %	11276					13
5.00	5.000	5.144	0.144	23508	0.0 %	23508					23
7.50	7.500	7.472	-0.028	34020	0.0 %	34020					0 23
10.0	10.000	9.964	-0.036	45269	0.0 %	45269					0 11/21/22

LB122683

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	368	-		0 SFAM01.1	PPB	10/30/2022 15:55	Std
	0.2 S01	1049	-		0.2 SFAM01.1	PPB	10/30/2022 15:57	Std
	2.5 S02	11276	-		2.5 SFAM01.1	PPB	10/30/2022 15:59	Std
	5 S03	23508	-		5 SFAM01.1	PPB	10/30/2022 16:02	Std
	7.5 S04	34020	-		7.5 SFAM01.1	PPB	10/30/2022 16:04	Std
	10 S05	45269	-		10 SFAM01.1	PPB	10/30/2022 16:06	Std
ICV097	ICV097	17795	3.8781 -		SFAM01.1	PPB	10/30/2022 16:09	SMPL
ICB097	ICB097	471	0.0408 -		SFAM01.1	PPB	10/30/2022 16:11	SMPL
CCV001	CCV001	23395	5.1185 -		SFAM01.1	PPB	10/30/2022 16:13	SMPL
CCB001	CCB001	319	0.0072 -		SFAM01.1	PPB	10/30/2022 16:16	SMPL
PB148679BL	PBS679	259	-0.0061 -		SFAM01.1	PPB	10/30/2022 16:18	SMPL
N5138-19	X3D02	30639	6.723 -		SFAM01.1	PPB	10/30/2022 16:20	SMPL
N5138-20	X3D02D	30693	6.735 -		SFAM01.1	PPB	10/30/2022 16:25	SMPL
N5138-21	X3D02S	40823	8.9788 -		SFAM01.1	PPB	10/30/2022 16:31	SMPL
CCV002	CCV002	23510	5.144 -		SFAM01.1	PPB	10/30/2022 16:34	SMPL
CCB002	CCB002	395	0.024 -		SFAM01.1	PPB	10/30/2022 16:36	SMPL