

LB12268T


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

49936



A= 0.0000e+000

B= 1.9854e-004

C= -5.8478e-003

Rho= 0.9998634

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.006	-0.006	0	0.000	0					
0.2	0.200	0.189	-0.011	979	0.0 %	979					6
2.5	2.500	2.467	-0.033	12457	0.0 %	12457					1
5.0	5.000	5.034	0.034	25383	0.0 %	25383					1
7.5	7.500	7.608	0.108	38347	0.0 %	38347					1
10.0	10.000	9.909	-0.091	49936	0.0 %	49936					1

LB122685 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 31 Oct 2022 08:44:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	0	-	0.0000	7471B	PPB	31 Oct 2022 09:00:56	S	Std
0.2 - 1	50.2	979	-	0.2000	7471B	PPB	31 Oct 2022 09:03:13	S	Std
2.5 - 1	52.5	12457	-	2.5000	7471B	PPB	31 Oct 2022 09:05:30	S	Std
5.0 - 1	55	25383	-	5.0000	7471B	PPB	31 Oct 2022 09:07:47	S	Std
7.5 - 1	57.5	38347	-	7.5000	7471B	PPB	31 Oct 2022 09:10:03	S	Std
10.0 - 1	510	49936	-	10.0000	7471B	PPB	31 Oct 2022 09:12:19	S	Std
ICV98 - 1	ICV98	20884	4.1405	-	7471B	PPB	31 Oct 2022 09:15:03	U	SMPL
ICB98 - 1	ICB98	17	-0.0025	-	7471B	PPB	31 Oct 2022 09:17:18	U	SMPL
CCV03 - 1	CCV03	25478	5.0526	-	7471B	PPB	31 Oct 2022 09:19:33	U	SMPL
CCB03 - 1	CCB03	-89	-0.0235	-	7471B	PPB	31 Oct 2022 09:21:49	U	SMPL
CRA - 1	CRA	941	0.1810	-	7471B	PPB	31 Oct 2022 09:24:03	U	SMPL
HighStd - 1	HighStd	49865	9.8944	-	7471B	PPB	31 Oct 2022 09:26:19	U	SMPL
ChkStd - 1	ChkStd	35695	7.0811	-	7471B	PPB	31 Oct 2022 09:28:34	U	SMPL
PB148634BL - 1	PBS	-122	-0.0301	-	7471B	PPB	31 Oct 2022 09:30:51	U	SMPL
PB148634BS - 1	LCSS	19894	3.9439	-	7471B	PPB	31 Oct 2022 09:33:07	U	SMPL
N5307-01 - 1	CL-02-102622	1532	0.2983	-	7471B	PPB	31 Oct 2022 09:35:23	U	SMPL
N5310-17 - 1	BIN 1-2-(2840)-COMP-1	499	0.0932	-	7471B	PPB	31 Oct 2022 09:37:39	U	SMPL
N5310-18 - 1	BIN 1-2-(2840)-COMP-2	352	0.0640	-	7471B	PPB	31 Oct 2022 09:39:56	U	SMPL
N5310-19 - 1	BIN 1-2-(2840)-COMP-3	565	0.1063	-	7471B	PPB	31 Oct 2022 09:42:14	U	SMPL
N5310-20 - 1	BIN 1-2-(2840)-COMP-4	387	0.0710	-	7471B	PPB	31 Oct 2022 09:44:29	U	SMPL
CCV04 - 1	CCV04	25373	5.0318	-	7471B	PPB	31 Oct 2022 09:46:44	U	SMPL
CCB04 - 1	CCB04	-120	-0.0297	-	7471B	PPB	31 Oct 2022 09:49:00	U	SMPL
N5310-21 - 1	BIN 1-2-(2840)-COMP-5	576	0.1085	-	7471B	PPB	31 Oct 2022 09:51:15	U	SMPL
N5313-01 - 1	COMPOSITE-SW-QUARTER	3951	0.7786	-	7471B	PPB	31 Oct 2022 09:53:30	U	SMPL
N5313-08 - 1	COMPOSITE-NW-QUARTER	2382	0.4671	-	7471B	PPB	31 Oct 2022 09:55:45	U	SMPL
N5313-15 - 1	COMPOSITE-SE-QUARTER	4675	0.9223	-	7471B	PPB	31 Oct 2022 09:58:01	U	SMPL
N5316-01 - 1	COMPOSITE-NE-QUARTER	2931	0.5761	-	7471B	PPB	31 Oct 2022 10:00:17	U	SMPL
N5316-08 - 1	COMPOSITE-PILE-2	1729	0.3374	-	7471B	PPB	31 Oct 2022 10:02:33	U	SMPL
N5327-01 - 1	LACY-05	35400	7.0225	-	7471B	PPB	31 Oct 2022 10:04:49	U	SMPL
N5327-04 - 1	LACY-03	30675	6.0844	-	7471B	PPB	31 Oct 2022 10:07:06	U	SMPL
N5330-01 - 1	TR-5-102722	2127	0.4165	-	7471B	PPB	31 Oct 2022 10:09:24	U	SMPL
N5331-01 - 1	ROLLOFF-RB21502	21439	4.2507	-	7471B	PPB	31 Oct 2022 10:11:39	U	SMPL
CCV05 - 1	CCV05	25693	5.0953	-	7471B	PPB	31 Oct 2022 10:13:55	U	SMPL
CCB05 - 1	CCB05	-37	-0.0132	-	7471B	PPB	31 Oct 2022 10:16:10	U	SMPL
N5332-01 - 1	MH-4-WC	4314	0.8507	-	7471B	PPB	31 Oct 2022 10:18:26	U	SMPL
N5335-01 - 1	SB-1	584	0.1101	-	7471B	PPB	31 Oct 2022 10:20:41	U	SMPL
N5335-02 - 1	SB-2	390	0.0716	-	7471B	PPB	31 Oct 2022 10:22:57	U	SMPL
N5335-03 - 1	SB-3	4636	0.9146	-	7471B	PPB	31 Oct 2022 10:25:12	U	SMPL
N5335-04 - 1	SB-4	1055	0.2036	-	7471B	PPB	31 Oct 2022 10:27:28	U	SMPL
N5335-04DUP - 1	SB-4DUP	955	0.1838	-	7471B	PPB	31 Oct 2022 10:29:44	U	SMPL
N5335-04MS - 1	SB-4MS	19044	3.7752	-	7471B	PPB	31 Oct 2022 10:32:00	U	SMPL
N5335-04MSD - 1	SB-4MSD	19881	3.9414	-	7471B	PPB	31 Oct 2022 10:34:17	U	SMPL
PB148640BL - 1	PBS	485	0.0904	-	7471B	PPB	31 Oct 2022 10:36:35	U	SMPL
PB148640BS - 1	LCSS	20021	3.9692	-	7471B	PPB	31 Oct 2022 10:38:51	U	SMPL
CCV06 - 1	CCV06	25258	5.0089	-	7471B	PPB	31 Oct 2022 10:43:44	U	SMPL
CCB06 - 1	CCB06	-46	-0.0150	-	7471B	PPB	31 Oct 2022 10:48:40	U	SMPL
N5234-01 - 1	LSP-SS-03-01-03	3029	0.5955	-	7471B	PPB	31 Oct 2022 10:50:56	U	SMPL
N5234-02 - 1	LSP-SS-03-03-05	1211	0.2346	-	7471B	PPB	31 Oct 2022 10:53:11	U	SMPL
N5234-03 - 1	LSP-SS-03-01-03-D	729	0.1389	-	7471B	PPB	31 Oct 2022 10:55:27	U	SMPL
N5234-04 - 1	LSP-SS-04-05-15	2947	0.5793	-	7471B	PPB	31 Oct 2022 10:57:42	U	SMPL
N5234-05 - 1	LSP-SS-04-15-3	3664	0.7216	-	7471B	PPB	31 Oct 2022 10:59:58	U	SMPL
N5234-06 - 1	LSP-SS-06-00-20	1883	0.3680	-	7471B	PPB	31 Oct 2022 11:02:14	U	SMPL
N5234-07 - 1	LSP-SS-06-20-40	1796	0.3507	-	7471B	PPB	31 Oct 2022 11:04:31	U	SMPL
N5234-08 - 1	LSP-SS-05-08-13	2711	0.5324	-	7471B	PPB	31 Oct 2022 11:06:47	U	SMPL
N5234-09 - 1	LSP-SS-07-00-20	293	0.0523	-	7471B	PPB	31 Oct 2022 11:09:04	U	SMPL
N5234-10 - 1	LSP-SS-07-20-40	2281	0.4470	-	7471B	PPB	31 Oct 2022 11:11:20	U	SMPL
CCV07 - 1	CCV07	25187	4.9948	-	7471B	PPB	31 Oct 2022 11:18:38	U	SMPL
CCB07 - 1	CCB07	-81	-0.0219	-	7471B	PPB	31 Oct 2022 11:20:53	U	SMPL
N5235-01 - 1	LSP-SS-01-00-20	1684	0.3285	-	7471B	PPB	31 Oct 2022 11:28:45	U	SMPL
N5235-01DUP - 1	LSP-SS-01-00-20DUP	1629	0.3176	-	7471B	PPB	31 Oct 2022 11:31:00	U	SMPL
N5235-02 - 1	N5235-01MS	21739	4.3103	-	7471B	PPB	31 Oct 2022 11:33:16	U	SMPL
N5235-03 - 1	N5235-01MSD	22662	4.4935	-	7471B	PPB	31 Oct 2022 11:35:31	U	SMPL
N5235-04 - 1	LSP-DS-01-09-11	664	0.1260	-	7471B	PPB	31 Oct 2022 11:37:47	U	SMPL
N5235-05 - 1	LSP-SS-02-01-03	949	0.1826	-	7471B	PPB	31 Oct 2022 11:40:04	U	SMPL
N5235-06 - 1	LSP-SS-02-03-05	2111	0.4133	-	7471B	PPB	31 Oct 2022 11:42:20	U	SMPL
N5339-01 - 1	NB-07-102722	1701	0.3319	-	7471B	PPB	31 Oct 2022 11:44:37	U	SMPL
N5341-01 - 1	WC-1	2868	0.5636	-	7471B	PPB	31 Oct 2022 11:46:54	U	SMPL
N5341-04 - 1	WC-2	5034	0.9936	-	7471B	PPB	31 Oct 2022 11:49:10	U	SMPL
CCV08 - 1	CCV08	26911	5.3371	-	7471B	PPB	31 Oct 2022 11:51:26	U	SMPL
CCB08 - 1	CCB08	-136	-0.0328	-	7471B	PPB	31 Oct 2022 11:53:42	U	SMPL
N5359-01 - 1	HD-01-102822	7800	1.5428	-	7471B	PPB	31 Oct 2022 11:55:58	U	SMPL
PB148644BL - 1	PBS	1636	0.3190	-	7471B	PPB	31 Oct 2022 11:58:15	U	SMPL
PB148644BS - 1	LCSS	22289	4.4195	-	7471B	PPB	31 Oct 2022 12:00:31	U	SMPL
N5219-01 - 1	BACKYARD-SURFACE-A	8316	1.6452	-	7471B	PPB	31 Oct 2022 12:02:47	U	SMPL
N5219-02 - 1	BACKYARD-BOTTOM-A	3076	0.6049	-	7471B	PPB	31 Oct 2022 12:05:03	U	SMPL
N5219-03 - 1	BACKYARD-SURFACE-B	3827	0.7540	-	7471B	PPB	31 Oct 2022 12:07:19	U	SMPL
N5219-04 - 1	BACKYARD-BOTTOM-B	3856	0.7597	-	7471B	PPB	31 Oct 2022 12:09:35	U	SMPL
N5272-01 - 1	20221102-COMP-01	10049	1.9893	-	7471B	PPB	31 Oct 2022 12:11:52	U	SMPL
N5272-02 - 1	20221103-COMP-02	6223	1.2297	-	7471B	PPB	31 Oct 2022 12:14:09	U	SMPL
N5272-03 - 1	20221104-COMP-03	8784	1.7381	-	7471B	PPB	31 Oct 2022 12:16:25	U	SMPL
CCV09 - 1	CCV09	26075	5.1711	-	7471B	PPB	31 Oct 2022 12:18:42	U	SMPL

LB122685 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 31 Oct 2022 08:44:42

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
CCB09 - 1	CCB09	-52	-0.0162		-7471B	PPB	31 Oct 2022 12:23:39	U	SMPL
N5272-04 - 1	20221109-COMP-08	5601	1.1062		-7471B	PPB	31 Oct 2022 12:25:55	U	SMPL
N5272-05 - 1	20221110-COMP-09	10211	2.0215		-7471B	PPB	31 Oct 2022 12:28:11	U	SMPL
N5273-01 - 1	LIDEN-GAS-PIPELINE	7951	1.5728		-7471B	PPB	31 Oct 2022 12:30:27	U	SMPL
N5336-01 - 1	SB-01-0-2.0	1930	0.3773		-7471B	PPB	31 Oct 2022 12:40:41	U	SMPL
N5336-01DUP - 1	SB-01-0-2.0DUP	1879	0.3672		-7471B	PPB	31 Oct 2022 12:42:57	U	SMPL
N5336-01MS - 1	SB-01-0-2.0MS	24259	4.8106		-7471B	PPB	31 Oct 2022 12:45:13	U	SMPL
N5336-01MSD - 1	SB-01-0-2.0MSD	21895	4.3412		-7471B	PPB	31 Oct 2022 12:47:29	U	SMPL
N5336-02 - 1	SB-01-16.5-17.0	2093	0.4097		-7471B	PPB	31 Oct 2022 12:49:46	U	SMPL
N5336-03 - 1	SB-01-16.5-17.0-DUP	1630	0.3178		-7471B	PPB	31 Oct 2022 12:52:03	U	SMPL
N5336-05 - 1	SB-02-0-2.0	2124	0.4159		-7471B	PPB	31 Oct 2022 12:54:19	U	SMPL
CCV10 - 1	CCV10	24999	4.9575		-7471B	PPB	31 Oct 2022 12:59:13	U	SMPL
CCB10 - 1	CCB10	-115	-0.0287		-7471B	PPB	31 Oct 2022 13:01:29	U	SMPL
N5336-06 - 1	SB-02-16.5-17.0	1791	0.3497		-7471B	PPB	31 Oct 2022 13:03:45	U	SMPL
N5337-02 - 1	SB-01-COMP	8944	1.7699		-7471B	PPB	31 Oct 2022 13:06:01	U	SMPL
N5337-04 - 1	SB-02-COMP	9811	1.9420		-7471B	PPB	31 Oct 2022 13:08:17	U	SMPL
N5337-06 - 1	SB-03-COMP	10424	2.0638		-7471B	PPB	31 Oct 2022 13:10:34	U	SMPL
N5335-04LX5 - 1		68	0.0077		-7471B	PPB	31 Oct 2022 13:12:51	U	SMPL
N5335-04A - 1		20639	4.0919		-7471B	PPB	31 Oct 2022 13:15:07	U	SMPL
N5235-01LX5 - 1		-33	-0.0124		-7471B	PPB	31 Oct 2022 13:17:24	U	SMPL
N5235-01A - 1		21423	4.2475		-7471B	PPB	31 Oct 2022 13:19:41	U	SMPL
N5336-01LX5 - 1		292	0.0521		-7471B	PPB	31 Oct 2022 13:21:58	U	SMPL
N5336-01A - 1		23693	4.6982		-7471B	PPB	31 Oct 2022 13:24:14	U	SMPL
CCV11 - 1	CCV11	24940	4.9458		-7471B	PPB	31 Oct 2022 13:28:12	U	SMPL
CCB11 - 1	CCB11	-109	-0.0275		-7471B	PPB	31 Oct 2022 13:30:27	U	SMPL