



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122691

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	82.4	10/31/2022	11:04
2	N5352-01		1	14	100.00	>212.0	>212.0	10/31/2022	11:35
3	N5352-01DUP		1	13	100.00	>212.0	>212.0	10/31/2022	12:05
4	N5377-02		1	12	100.00	>212.0	>212.0	10/31/2022	12:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 122691

WORKLIST(Hardcopy Internal Chain)

WorkList Name : flash-1031 WorkList ID : 164567 Department : Wet-Chemistry Date : 10-31-2022 10:54:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/04/2022	Water	N5352-01	Flash Point	Cool 4 deg C	PSEG03	F11	CMS-WATERPIT-1028	10/28/2022	1010B
11/02/2022	Water	N5377-02	Flash Point	Cool 4 deg C	PARS02	F11	LIQUID-WC-20221028	10/28/2022	1010B

Date/Time 10/31/2022 10:55
Raw Sample Received by: SI (WC)
Raw Sample Relinquished by: CP (SM)

Date/Time 10/31/2022 13:45
Raw Sample Received by: CP (SM)
Raw Sample Relinquished by: SI (WC)