


LB122698

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.2295e-004

C= -3.6024e-002

Rho= 0.9995597

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/21
0.0	0.000	0.030	0.030	294	0.000	294					
0.2	0.200	0.215	0.015	1124	0.0 %	1124					
2.5	2.500	2.493	-0.007	11345	0.0 %	11345					
5.0	5.000	4.838	-0.162	21860	0.0 %	21860					
7.5	7.500	7.698	0.198	34688	0.0 %	34688					
10.0	10.000	9.927	-0.073	44686	0.0 %	44686					

136071

LB122698 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 31 Oct 2022 15:17:56

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	294	-	0.0000	7470A	PPB	31 Oct 2022 15:18:47	S	Std
0.2 - 1	50.2	1124	-	0.2000	7470A	PPB	31 Oct 2022 15:21:04	S	Std
2.5 - 1	52.5	11345	-	2.5000	7470A	PPB	31 Oct 2022 15:23:21	S	Std
5.0 - 1	55	21860	-	5.0000	7470A	PPB	31 Oct 2022 15:25:38	S	Std
7.5 - 1	57.5	34688	-	7.5000	7470A	PPB	31 Oct 2022 15:27:54	S	Std
10.0 - 1	59.1	44686	-	10.0000	7470A	PPB	31 Oct 2022 15:32:49	S	Std
ICV99 - 1	ICV99	18075	3.9939	-	7470A	PPB	31 Oct 2022 15:35:39	U	SMPL
ICB99 - 1	ICB99	358	0.0438	-	7470A	PPB	31 Oct 2022 15:37:54	U	SMPL
CCV12 - 1	CCV12	24531	5.4333	-	7470A	PPB	31 Oct 2022 15:40:09	U	SMPL
CCB12 - 1	CCB12	195	0.0075	-	7470A	PPB	31 Oct 2022 15:42:24	U	SMPL
CRA - 1	CRA	1029	0.1934	-	7470A	PPB	31 Oct 2022 15:44:39	U	SMPL
HighStd - 1	HighStd	44251	9.8299	-	7470A	PPB	31 Oct 2022 15:46:54	U	SMPL
ChkStd - 1	ChkStd	31889	7.0738	-	7470A	PPB	31 Oct 2022 15:49:09	U	SMPL
PB148637BL - 1	PBW	192	0.0068	-	7470A	PPB	31 Oct 2022 15:51:24	U	SMPL
PB148637BS - 1	LCSW	16460	3.6338	-	7470A	PPB	31 Oct 2022 15:53:41	U	SMPL
N5313-02 - 1	COMPOSITE-SW-QUARTER	232	0.0157	-	7470A	PPB	31 Oct 2022 15:55:57	U	SMPL
N5313-09 - 1	COMPOSITE-NW-QUARTER	241	0.0177	-	7470A	PPB	31 Oct 2022 15:58:13	U	SMPL
N5313-16 - 1	COMPOSITE-SE-QUARTER	187	0.0057	-	7470A	PPB	31 Oct 2022 16:00:30	U	SMPL
N5316-02 - 1	COMPOSITE-NE-QUARTER	668	0.1129	-	7470A	PPB	31 Oct 2022 16:02:47	U	SMPL
N5316-09 - 1	COMPOSITE-PILE-2	587	0.0949	-	7470A	PPB	31 Oct 2022 16:05:02	U	SMPL
CCV13 - 1	CCV13	23619	5.2299	-	7470A	PPB	31 Oct 2022 16:07:17	U	SMPL
CCB13 - 1	CCB13	187	0.0057	-	7470A	PPB	31 Oct 2022 16:09:32	U	SMPL
N5327-03 - 1	LACY-05	474	0.0697	-	7470A	PPB	31 Oct 2022 16:11:47	U	SMPL
N5327-06 - 1	LACY-03	438	0.0616	-	7470A	PPB	31 Oct 2022 16:14:03	U	SMPL
N5332-02 - 1	MH-4-WC	300	0.0309	-	7470A	PPB	31 Oct 2022 16:16:18	U	SMPL
N5332-02DUP - 1	MH-4-WCDUP	259	0.0217	-	7470A	PPB	31 Oct 2022 16:18:34	U	SMPL
N5332-02MS - 1	MH-4-WCMS	16465	3.6349	-	7470A	PPB	31 Oct 2022 16:20:50	U	SMPL
N5332-02MSD - 1	MH-4-WCMSD	16519	3.6470	-	7470A	PPB	31 Oct 2022 16:23:06	U	SMPL
PB148604TB - 1	PB148604TB	241	0.0177	-	7470A	PPB	31 Oct 2022 16:25:23	U	SMPL
N5332-02LX5 - 1		51	-0.0247	-	7470A	PPB	31 Oct 2022 16:27:40	U	SMPL
N5332-02A - 1		16685	3.6840	-	7470A	PPB	31 Oct 2022 16:29:57	U	SMPL
CCV14 - 1	CCV14	23402	5.1816	-	7470A	PPB	31 Oct 2022 16:32:12	U	SMPL
CCB14 - 1	CCB14	226	0.0144	-	7470A	PPB	31 Oct 2022 16:34:28	U	SMPL