



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB122702

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	10/31/2022	17:20
2	CAL2	1	Water	NA	NA	20.7	7.00	10/31/2022	17:21
3	CAL3	1	Water	NA	NA	20.7	10.03	10/31/2022	17:22
4	ICV	1	Water	NA	NA	20.7	6.99	10/31/2022	17:25
5	CCV1	1	Water	NA	NA	20.8	2.01	10/31/2022	17:30
6	N5381-01	1	Water	NA	NA	20.1	7.40	10/31/2022	17:32
7	N5381-02	1	Water	NA	NA	20.1	7.56	10/31/2022	17:35
8	N5383-01	1	Water	NA	NA	20.1	7.58	10/31/2022	17:39
9	N5383-02	1	Water	NA	NA	20.7	7.61	10/31/2022	17:40
10	N5385-01	1	Water	NA	NA	20.8	7.71	10/31/2022	17:43
11	N5385-02	1	Water	NA	NA	20.8	7.66	10/31/2022	17:45
12	N5391-05	1	Water	NA	NA	21.4	6.59	10/31/2022	17:49
13	N5391-06	1	Water	NA	NA	21.3	5.43	10/31/2022	17:50
14	N5394-01	1	Water	NA	NA	20.1	4.72	10/31/2022	17:52
15	N5394-01DUP	1	Water	NA	NA	20.2	4.73	10/31/2022	17:53
16	CCV2	1	Water	NA	NA	20.7	12.02	10/31/2022	18:00

WORKLIST(Hardcopy Internal Chain)

13122702

WorkList Name : ph n5381

WorkList ID : 164577

Department : Wet-Chemistry

Date : 10-31-2022 15:43:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/28/2022	Water	N5381-01	pH <u>E</u>	Cool 4 deg C	ENV170	F11	265393-7	10/28/2022	9040C
10/28/2022	Water	N5381-02	pH <u>E</u>	Cool 4 deg C	ENV170	F11	265393-8	10/28/2022	9040C
10/28/2022	Water	N5383-01	pH <u>E</u>	Cool 4 deg C	ENV170	F11	256317-7	10/28/2022	9040C
10/28/2022	Water	N5383-02	pH <u>E</u>	Cool 4 deg C	ENV170	F11	256317-8	10/28/2022	9040C
10/28/2022	Water	N5385-01	pH <u>E</u>	Cool 4 deg C	ENV170	F11	251563-5	10/28/2022	9040C
10/28/2022	Water	N5385-02	pH <u>E</u>	Cool 4 deg C	ENV170	F11	251563-6	10/28/2022	9040C
10/31/2022	Water	N5391-05	pH <u>C</u>	Cool 4 deg C	PSEG03	F11	31845	10/31/2022	9040C
10/31/2022	Water	N5391-06	pH <u>C</u>	Cool 4 deg C	PSEG03	F11	31846	10/31/2022	9040C
10/28/2022	Water	N5394-01	pH <u>A</u>	Cool 4 deg C	STER14	F11	MINI-CELL	10/28/2022	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

10/31/22 18:10

20 (SM)

20 (SM)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

10/31/22

20 (SM)

20 (SM)