



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB122703

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	10/31/2022	17:20
2	CAL2	1	Water	NA	NA	20.7	7.00	10/31/2022	17:21
3	CAL3	1	Water	NA	NA	20.7	10.03	10/31/2022	17:22
4	ICV	1	Water	NA	NA	20.7	6.99	10/31/2022	17:25
5	CCV1	1	Water	NA	NA	20.8	2.01	10/31/2022	17:30
6	N5391-01	1	Solid	20.01	20	20.7	5.86	10/31/2022	18:02
7	N5391-01DUP	1	Solid	20.02	20	20.8	5.87	10/31/2022	18:03
8	CCV2	1	Water	NA	NA	20.8	12.03	10/31/2022	18:05

UB122703

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S N5391 WorkList ID : 164578 Department : Wet-Chemistry Date : 10-31-2022 17:30:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/31/2022	Solid	N5391-01	pH		Cool 4 deg C	PSEG03 F11	31561	10/31/2022	9045D

Date/Time 10/31/22 17:35
Raw Sample Received by: JR(WOC)
Raw Sample Relinquished by: JT(SM)

Date/Time 10/31/22 18:35
Raw Sample Received by: JT(SM)
Raw Sample Relinquished by: JR(WOC)