

L3122706


7470A

INSTRUMENT ID: CV1

Linear

$\mu$  Abs.:

42209



A= 0.0000e+000

B= 2.3127e-004

C= 8.0626e-002

Rho= 0.9991196

Accept=Accepted

| Std ID | Conc.  | Calc. | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |      |
|--------|--------|-------|--------|-------|------------|-------|-------|-------|-------|-------|------|
| 0.0    | 0.000  | 0.074 | 0.074  | -28   | 0.000      | -28   |       |       |       |       | 7/11 |
| 0.2    | 0.200  | 0.234 | 0.034  | 665   | 0.0 %      | 665   |       |       |       |       | 17   |
| 2.5    | 2.500  | 2.285 | -0.215 | 9531  | 0.0 %      | 9531  |       |       |       |       | -9   |
| 5.0    | 5.000  | 5.007 | 0.007  | 21302 | 0.0 %      | 21302 |       |       |       |       | 0    |
| 7.5    | 7.500  | 7.757 | 0.257  | 33191 | 0.0 %      | 33191 |       |       |       |       | 3    |
| 10.0   | 10.000 | 9.842 | -0.158 | 42209 | 0.0 %      | 42209 |       |       |       |       | -2   |

## LB122706 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Nov 2022 10:50:57

| Sample ID       | Extended ID                                   | μ Abs | Conc   | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|-----------------------------------------------|-------|--------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 30                                            | -28   | -      | 0.000    | 7470A  | PPB   | 01 Nov 2022 10:55:00 | S    | Std  |
| 0.2 - 1         | 50.2                                          | 665   | -      | 0.200    | 7470A  | PPB   | 01 Nov 2022 10:57:18 | S    | Std  |
| 2.5 - 1         | 32.5                                          | 9531  | -      | 2.500    | 7470A  | PPB   | 01 Nov 2022 10:59:35 | S    | Std  |
| 5.0 - 1         | 55                                            | 21302 | -      | 5.000    | 7470A  | PPB   | 01 Nov 2022 11:01:52 | S    | Std  |
| 7.5 - 1         | 57.5                                          | 33191 | -      | 7.500    | 7470A  | PPB   | 01 Nov 2022 11:04:08 | S    | Std  |
| 10.0 - 1        | 57.5                                          | 42209 | -      | 10.000   | 7470A  | PPB   | 01 Nov 2022 11:06:24 | S    | Std  |
| ICV01 - 1       | ICV01                                         | 17112 | 4.0382 | -        | 7470A  | PPB   | 01 Nov 2022 11:11:57 | U    | SMPL |
| ICB01 - 1       | ICB01                                         | 11    | 0.0832 | -        | 7470A  | PPB   | 01 Nov 2022 11:14:12 | U    | SMPL |
| CCV15 - 1       | CCV15                                         | 20823 | 4.8978 | -        | 7470A  | PPB   | 01 Nov 2022 11:19:01 | U    | SMPL |
| CCB15 - 1       | CCB15                                         | -130  | 0.0506 | -        | 7470A  | PPB   | 01 Nov 2022 11:21:16 | U    | SMPL |
| CRA - 1         | CRA                                           | 621   | 0.2242 | -        | 7470A  | PPB   | 01 Nov 2022 11:23:31 | U    | SMPL |
| HighStd - 1     | HighStd                                       | 42757 | 9.9692 | -        | 7470A  | PPB   | 01 Nov 2022 11:29:21 | U    | SMPL |
| ChkStd - 1      | ChkStd                                        | 28500 | 6.6719 | -        | 7470A  | PPB   | 01 Nov 2022 11:41:10 | U    | SMPL |
| PB148702BL - 1  | PBW                                           | -140  | 0.0482 | -        | 7470A  | PPB   | 01 Nov 2022 11:43:28 | U    | SMPL |
| PB148702BS - 1  | LCSW                                          | 14628 | 3.4637 | -        | 7470A  | PPB   | 01 Nov 2022 11:48:13 | U    | SMPL |
| N5336-04 - 1    | SB-01-COMP                                    | -116  | 0.0538 | -        | 7470A  | PPB   | 01 Nov 2022 11:50:28 | U    | SMPL |
| N5336-07 - 1    | SB-02-COMP                                    | -49   | 0.0693 | -        | 7470A  | PPB   | 01 Nov 2022 11:52:45 | U    | SMPL |
| N5337-07 - 1    | COMP-01                                       | -61   | 0.0665 | -        | 7470A  | PPB   | 01 Nov 2022 11:55:02 | U    | SMPL |
| N5341-03 - 1    | WC-1                                          | -82   | 0.0617 | -        | 7470A  | PPB   | 01 Nov 2022 11:57:19 | U    | SMPL |
| N5341-06 - 1    | WC-2                                          | -77   | 0.0628 | -        | 7470A  | PPB   | 01 Nov 2022 11:59:34 | U    | SMPL |
| CCV16 - 1       | CCV16                                         | 19317 | 4.5481 | -        | 7470A  | PPB   | 01 Nov 2022 12:01:49 | U    | SMPL |
| CCB16 - 1       | CCB16                                         | -56   | 0.0677 | -        | 7470A  | PPB   | 01 Nov 2022 12:04:04 | U    | SMPL |
| N5371-02 - 1    | MH-6-WC                                       | -28   | 0.0741 | -        | 7470A  | PPB   | 01 Nov 2022 12:06:19 | U    | SMPL |
| N5371-05 - 1    | MH-7-WC                                       | -4    | 0.0797 | -        | 7470A  | PPB   | 01 Nov 2022 12:08:35 | U    | SMPL |
| N5371-08 - 1    | MH-2-WC                                       | -37   | 0.0721 | -        | 7470A  | PPB   | 01 Nov 2022 12:10:51 | U    | SMPL |
| N5377-01 - 1    | SOIL-WC-20221028                              | 192   | 0.1250 | -        | 7470A  | PPB   | 01 Nov 2022 12:13:07 | U    | SMPL |
| N5377-01DUP - 1 | SOIL-WC-20221028DUP                           | 146   | 0.1144 | -        | 7470A  | PPB   | 01 Nov 2022 12:15:22 | U    | SMPL |
| N5377-01MS - 1  | SOIL-WC-20221028MS                            | 14009 | 3.3205 | -        | 7470A  | PPB   | 01 Nov 2022 12:17:37 | U    | SMPL |
| N5377-01MSD - 1 | SOIL-WC-20221028MSD                           | 14605 | 3.4584 | -        | 7470A  | PPB   | 01 Nov 2022 12:22:29 | U    | SMPL |
| PB148703BL - 1  | PBW                                           | -22   | 0.0755 | -        | 7470A  | PPB   | 01 Nov 2022 12:24:46 | U    | SMPL |
| PB148703BS - 1  | LCSW                                          | 13586 | 3.2227 | -        | 7470A  | PPB   | 01 Nov 2022 12:27:03 | U    | SMPL |
| N5333-01 - 1    | GLEASON-WASH-TANK-DRUM                        | 337   | 0.1586 | -        | 7470A  | PPB   | 01 Nov 2022 12:29:18 | U    | SMPL |
| CCV17 - 1       | CCV17                                         | 19413 | 4.5703 | -        | 7470A  | PPB   | 01 Nov 2022 12:31:34 | U    | SMPL |
| CCB17 - 1       | CCB17                                         | -31   | 0.0735 | -        | 7470A  | PPB   | 01 Nov 2022 12:33:49 | U    | SMPL |
| N5377-02 - 1    | LIQUID-WC-20221028                            | 38    | 0.0894 | -        | 7470A  | PPB   | 01 Nov 2022 12:36:05 | U    | SMPL |
| N5377-02DUP - 1 | LIQUID-WC-20221028DUP                         | 39    | 0.0896 | -        | 7470A  | PPB   | 01 Nov 2022 12:38:20 | U    | SMPL |
| N5377-02MS - 1  | LIQUID-WC-20221028MS                          | 14953 | 3.5388 | -        | 7470A  | PPB   | 01 Nov 2022 12:45:50 | U    | SMPL |
| N5377-02MSD - 1 | LIQUID-WC-20221028MSD                         | 15052 | 3.5618 | -        | 7470A  | PPB   | 01 Nov 2022 12:48:06 | U    | SMPL |
| PB148719BL - 1  | PBW                                           | -70   | 0.0644 | -        | 7470A  | PPB   | 01 Nov 2022 12:50:22 | U    | SMPL |
| PB148719BS - 1  | LCSW                                          | 14749 | 3.4917 | -        | 7470A  | PPB   | 01 Nov 2022 12:52:38 | U    | SMPL |
| N5306-06 - 1    | DISSOLVED-20221166-COMP-01-MODIFIED-ELUTRIATE | -22   | 0.0755 | -        | 7470A  | PPB   | 01 Nov 2022 12:54:55 | U    | SMPL |
| N5306-07 - 1    | DISSOLVED-20221167-COMP-02-MODIFIED-ELUTRIATE | 50    | 0.0922 | -        | 7470A  | PPB   | 01 Nov 2022 12:57:11 | U    | SMPL |
| N5306-08 - 1    | DISSOLVED-20221168-COMP-03-MODIFIED-ELUTRIATE | -3    | 0.0799 | -        | 7470A  | PPB   | 01 Nov 2022 12:59:28 | U    | SMPL |
| N5306-09 - 1    | DISSOLVED-20221173-COMP-08-MODIFIED-ELUTRIATE | -26   | 0.0746 | -        | 7470A  | PPB   | 01 Nov 2022 13:01:44 | U    | SMPL |
| CCV18 - 1       | CCV18                                         | 19552 | 4.6025 | -        | 7470A  | PPB   | 01 Nov 2022 13:04:00 | U    | SMPL |
| CCB18 - 1       | CCB18                                         | -168  | 0.0418 | -        | 7470A  | PPB   | 01 Nov 2022 13:06:15 | U    | SMPL |
| N5306-10 - 1    | DISSOLVED-20221174-COMP-09-MODIFIED-ELUTRIATE | 28    | 0.0866 | -        | 7470A  | PPB   | 01 Nov 2022 13:08:31 | U    | SMPL |
| N5352-01 - 1    | CMS-WATERPIT-1028                             | -56   | 0.0677 | -        | 7470A  | PPB   | 01 Nov 2022 13:10:46 | U    | SMPL |
| N5352-01DUP - 1 | CMS-WATERPIT-1028DUP                          | -81   | 0.0619 | -        | 7470A  | PPB   | 01 Nov 2022 13:13:02 | U    | SMPL |
| N5352-01MS - 1  | CMS-WATERPIT-1028MS                           | 14522 | 3.4392 | -        | 7470A  | PPB   | 01 Nov 2022 13:17:57 | U    | SMPL |
| N5352-01MSD - 1 | CMS-WATERPIT-1028MSD                          | 17333 | 4.0893 | -        | 7470A  | PPB   | 01 Nov 2022 13:23:28 | U    | SMPL |
| N5355-01 - 1    | 5518-1                                        | -51   | 0.0688 | -        | 7470A  | PPB   | 01 Nov 2022 13:25:44 | U    | SMPL |
| N5355-02 - 1    | 5518-2                                        | 6     | 0.0820 | -        | 7470A  | PPB   | 01 Nov 2022 13:28:00 | U    | SMPL |
| N5357-01 - 1    | 5051-1                                        | -34   | 0.0728 | -        | 7470A  | PPB   | 01 Nov 2022 13:30:17 | U    | SMPL |
| N5357-02 - 1    | 5051-2                                        | -41   | 0.0711 | -        | 7470A  | PPB   | 01 Nov 2022 13:32:34 | U    | SMPL |
| N5360-15 - 1    | DISSOLVED-20221169-COMP-04-MODIFIED-ELUTRIATE | -15   | 0.0772 | -        | 7470A  | PPB   | 01 Nov 2022 13:34:49 | U    | SMPL |
| CCV19 - 1       | CCV19                                         | 19303 | 4.5449 | -        | 7470A  | PPB   | 01 Nov 2022 13:37:05 | U    | SMPL |
| CCB19 - 1       | CCB19                                         | -125  | 0.0517 | -        | 7470A  | PPB   | 01 Nov 2022 13:39:21 | U    | SMPL |
| N5360-16 - 1    | DISSOLVED-20221171-COMP-06-MODIFIED-ELUTRIATE | 35    | 0.0887 | -        | 7470A  | PPB   | 01 Nov 2022 13:41:36 | U    | SMPL |
| N5360-17 - 1    | DISSOLVED-20221172-COMP-07-MODIFIED-ELUTRIATE | 22    | 0.0857 | -        | 7470A  | PPB   | 01 Nov 2022 13:43:52 | U    | SMPL |
| N5360-18 - 1    | DISSOLVED-20221175-COMP-10-MODIFIED-ELUTRIATE | 10    | 0.0829 | -        | 7470A  | PPB   | 01 Nov 2022 13:46:07 | U    | SMPL |
| N5360-19 - 1    | DISSOLVED-20221179-COMP-14-MODIFIED-ELUTRIATE | -20   | 0.0760 | -        | 7470A  | PPB   | 01 Nov 2022 13:48:23 | U    | SMPL |
| N5362-03 - 1    | ORA-1509                                      | 133   | 0.1114 | -        | 7470A  | PPB   | 01 Nov 2022 13:50:39 | U    | SMPL |
| PB148641TB - 1  | PB148641TB                                    | -44   | 0.0704 | -        | 7470A  | PPB   | 01 Nov 2022 13:52:56 | U    | SMPL |
| PB148672TB - 1  | PB148672TB                                    | -67   | 0.0651 | -        | 7470A  | PPB   | 01 Nov 2022 13:55:12 | U    | SMPL |
| CCV20 - 1       | CCV20                                         | 20621 | 4.8497 | -        | 7470A  | PPB   | 01 Nov 2022 14:00:07 | U    | SMPL |
| CCB20 - 1       | CCB20                                         | -74   | 0.0635 | -        | 7470A  | PPB   | 01 Nov 2022 14:02:24 | U    | SMPL |
| N5377-01LX5 - 1 |                                               | 38    | 0.0894 | -        | 7470A  | PPB   | 01 Nov 2022 14:04:39 | U    | SMPL |
| N5377-01A - 1   |                                               | 15283 | 3.6152 | -        | 7470A  | PPB   | 01 Nov 2022 14:06:55 | U    | SMPL |
| N5377-02LX5 - 1 |                                               | -84   | 0.0612 | -        | 7470A  | PPB   | 01 Nov 2022 14:09:11 | U    | SMPL |
| N5377-02A - 1   |                                               | 15830 | 3.7417 | -        | 7470A  | PPB   | 01 Nov 2022 14:11:26 | U    | SMPL |
| N5352-01LX5 - 1 |                                               | -111  | 0.0550 | -        | 7470A  | PPB   | 01 Nov 2022 14:13:44 | U    | SMPL |
| N5352-01A - 1   |                                               | 16760 | 3.9568 | -        | 7470A  | PPB   | 01 Nov 2022 14:16:00 | U    | SMPL |
| CCV21 - 1       | CCV21                                         | 22629 | 5.3141 | -        | 7470A  | PPB   | 01 Nov 2022 14:18:16 | U    | SMPL |
| CCB21 - 1       | CCB21                                         | -90   | 0.0598 | -        | 7470A  | PPB   | 01 Nov 2022 14:20:32 | U    | SMPL |