



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Sohil

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122716

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	11/01/2022	14:10
2	N5387-01		1	14	100.00	>212.0	>212.0	11/01/2022	16:10
3	N5391-02		1	13	100.00	>212.0	>212.0	11/01/2022	16:40
4	N5391-02DUP		1	13	100.00	>212.0	>212.0	11/01/2022	17:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 122716

WorkList Name : FP-OIL-1101 WorkList ID : 164596 Department : Wet-Chemistry Date : 11-01-2022 14:09:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/07/2022	Solid	N5387-01	B Flash Point	Cool 4 deg C	PSEG03	F11	101722	10/31/2022	1010B
11/07/2022	Solid	N5391-02	A Flash Point	Cool 4 deg C	PSEG03	F11	15496	10/31/2022	1010B

Date/Time 11/1/22 14:10
 Raw Sample Received by: 5J (wc)
 Raw Sample Relinquished by: CP (SM)

Date/Time 11/1/22 14:30
 Raw Sample Received by: CP (SM)
 Raw Sample Relinquished by: 5J (wc)