



Analytical Summary Report

Analysis Method: 9041A

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB122720

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940
3.0-5.5 pH paper refill	W1934
Test Paper, PH, Short Range 6.0/8.0	W2350
Test Paper,PH Short Range 9.0/10.0	W1939
0.0-3.0 pH paper refill	W1931
10.0-12.0 pH paper refill	W1940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.0	11/01/2022	17:25
2	CAL2	1	Water	NA	NA	20.7	7.0	11/01/2022	17:26
3	CAL3	1	Water	NA	NA	20.7	10.0	11/01/2022	17:29
4	ICV	1	Water	NA	NA	20.8	7.0	11/01/2022	17:30
5	CCV1	1	Water	NA	NA	20.7	2.0	11/01/2022	17:32
6	N5317-03	1	Solid	20.02	20	20.3	9.7	11/01/2022	17:38
7	CCV2	1	Water	NA	NA	20.7	12.0	11/01/2022	18:20

WORKLIST(Hardcopy Internal Chain)

122720

WorkList Name : ph n5317

WorkList ID : 164621

Department : Wet-Chemistry

Date : 11-01-2022 13:39:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/24/2022	Solid	N5317-03	pH	Cool 4 deg C	CHEM02	QA Of	PT-CORR-SOIL	10/24/2022	9041A

17:10

Date/Time 11/01/22 13:40 1P
 Raw Sample Received by: 70(WC)
 Raw Sample Relinquished by: CP SR

Date/Time 11/01/22 18:30
 Raw Sample Received by: CP SR
 Raw Sample Relinquished by: 70(WC)