



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB122721

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	11/02/2022	09:35
2	CAL2	1	Water	NA	NA	20.7	7.00	11/02/2022	09:36
3	CAL3	1	Water	NA	NA	20.7	10.03	11/02/2022	09:37
4	ICV	1	Water	NA	NA	20.8	7.00	11/02/2022	09:39
5	CCV1	1	Water	NA	NA	20.7	2.01	11/02/2022	09:40
6	N5412-03	1	Water	NA	NA	20.7	7.95	11/02/2022	09:42
7	N5412-04	1	Water	NA	NA	20.2	6.93	11/02/2022	09:43
8	N5412-04DUP	1	Water	NA	NA	20.3	6.94	11/02/2022	09:44
9	CCV2	1	Water	NA	NA	20.8	12.03	11/02/2022	09:45

11/02/22

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N5412 w WorkList ID : 164623 Department : Wet-Chemistry Date : 11-02-2022 09:27:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/01/2022	Water	N5412-03	Corrosivity	Cool 4 deg C	PSEG03	F11	D-2748	11/01/2022	9040C
11/01/2022	Water	N5412-04	Corrosivity	Cool 4 deg C	PSEG03	F11	D-2763	11/01/2022	9040C

Date/Time 11/02/22 09:30
Raw Sample Received by: 70 (WC)
Raw Sample Relinquished by: CP SR

Date/Time 11/02/22 11:00
Raw Sample Received by: CP SR
Raw Sample Relinquished by: 70 (WC)