



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB122727

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	11/02/2022	13:20
2	CAL2	1	Water	NA	NA	20.7	7.00	11/02/2022	13:21
3	CAL3	1	Water	NA	NA	20.7	10.03	11/02/2022	13:22
4	ICV	1	Water	NA	NA	20.8	7.02	11/02/2022	13:25
5	CCV1	1	Water	NA	NA	20.7	2.01	11/02/2022	13:30
6	N5404-01	1	Water	NA	NA	20.1	7.55	11/02/2022	13:31
7	N5404-02	1	Water	NA	NA	20.2	7.66	11/02/2022	13:33
8	N5420-02	1	Water	NA	NA	20.7	4.84	11/02/2022	13:35
9	N5420-03	1	Water	NA	NA	20.3	5.52	11/02/2022	13:36
10	N5420-04	1	Water	NA	NA	20.7	5.31	11/02/2022	13:40
11	N5420-05	1	Water	NA	NA	20.6	5.50	11/02/2022	13:42
12	N5420-05DUP	1	Water	NA	NA	20.7	5.51	11/02/2022	13:43
13	CCV2	1	Water	NA	NA	20.7	12.03	11/02/2022	13:45

122727

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N5420 WorkList ID : 184644 Department : Wet-Chemistry Date : 11-02-2022 13:10:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/31/2022	Water	N5404-01	pH <u>E</u>	Cool 4 deg C	ENV170	F11	551B-3	10/31/2022	9040C
10/31/2022	Water	N5404-02	pH <u>E</u>	Cool 4 deg C	ENV170	F11	551B-4	10/31/2022	9040C
11/01/2022	Water	N5420-02	pH <u>D</u>	Cool 4 deg C	TETR06	F11	RW8-PT-IN01-11012022	11/01/2022	9040C
11/01/2022	Water	N5420-03	pH <u>D</u>	Cool 4 deg C	TETR06	F11	RW8-PT-OUT01-11012022	11/01/2022	9040C
11/01/2022	Water	N5420-04	pH <u>D</u>	Cool 4 deg C	TETR06	F11	RW8-PT-IN02-11012022	11/01/2022	9040C
11/01/2022	Water	N5420-05	pH <u>D</u>	Cool 4 deg C	TETR06	F11	RW8-PT-OUT-02-11012022	11/01/2022	9040C

Date/Time 11/02/22 13:15
 Raw Sample Received by: JP (COC)
 Raw Sample Relinquished by: JP (COC)

Date/Time 11/02/22 14:30
 Raw Sample Received by: JP (COC)
 Raw Sample Relinquished by: JP (COC)