

L3122728


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

47258



A= 0.0000e+000

B= 2.0899e-004

C= -1.6834e-002

Rho= 0.9996831

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.017	-0.017	1	0.000	1					7
0.2	0.200	0.191	-0.009	995	0.0 %	995					5
2.5	2.500	2.441	-0.059	11760	0.0 %	11760					2
5.0	5.000	5.074	0.074	24361	0.0 %	24361					1
7.5	7.500	7.651	0.151	36688	0.0 %	36688					2
10.0	10.000	9.860	-0.140	47258	0.0 %	47258					1

LB122728 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Nov 2022 13:34:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	51	1	-	0.0000	7471B	PPB	02 Nov 2022 13:59:29	S	Std
0.2 - 1	51.2	995	-	0.2000	7471B	PPB	02 Nov 2022 14:01:46	S	Std
2.5 - 1	52.5	11760	-	2.5000	7471B	PPB	02 Nov 2022 14:04:03	S	Std
5.0 - 1	55	24361	-	5.0000	7471B	PPB	02 Nov 2022 14:06:20	S	Std
7.5 - 1	57.5	36688	-	7.5000	7471B	PPB	02 Nov 2022 14:08:36	S	Std
10.0 - 1	51	47258	-	10.0000	7471B	PPB	02 Nov 2022 14:16:34	S	Std
ICV03 - 1	ICV03	19469	4.0520	-	7471B	PPB	02 Nov 2022 14:19:24	U	SMPL
ICB03 - 1	ICB03	-1	-0.0170	-	7471B	PPB	02 Nov 2022 14:21:39	U	SMPL
CCV26 - 1	CCV26	25558	5.3245	-	7471B	PPB	02 Nov 2022 14:23:54	U	SMPL
CCB26 - 1	CCB26	-120	-0.0419	-	7471B	PPB	02 Nov 2022 14:26:09	U	SMPL
CRA - 1	CRA	965	0.1848	-	7471B	PPB	02 Nov 2022 14:28:25	U	SMPL
HighStd - 1	HighStd	47944	10.0030	-	7471B	PPB	02 Nov 2022 14:30:40	U	SMPL
ChkStd - 1	ChkStd	34422	7.1770	-	7471B	PPB	02 Nov 2022 14:32:55	U	SMPL
PB148761BL - 1	PBS	-155	-0.0492	-	7471B	PPB	02 Nov 2022 14:35:11	U	SMPL
PB148761BS - 1	LCSS	17839	3.7113	-	7471B	PPB	02 Nov 2022 14:37:29	U	SMPL
N5415-01 - 1	R-14	537	0.0954	-	7471B	PPB	02 Nov 2022 14:48:47	U	SMPL
N5415-01DUP - 1	R-14DUP	483	0.0841	-	7471B	PPB	02 Nov 2022 14:51:03	U	SMPL
N5415-01MS - 1	R-14MS	19566	4.0723	-	7471B	PPB	02 Nov 2022 14:53:19	U	SMPL
N5415-01MSD - 1	R-14MSD	19106	3.9761	-	7471B	PPB	02 Nov 2022 14:55:36	U	SMPL
N5415-02 - 1	R-15	3093	0.6296	-	7471B	PPB	02 Nov 2022 14:57:51	U	SMPL
CCV27 - 1	CCV27	26003	5.4175	-	7471B	PPB	02 Nov 2022 15:00:07	U	SMPL
CCB27 - 1	CCB27	-73	-0.0321	-	7471B	PPB	02 Nov 2022 15:02:22	U	SMPL
N5415-03 - 1	R-16	1851	0.3700	-	7471B	PPB	02 Nov 2022 15:04:37	U	SMPL
N5415-04 - 1	R-17	1281	0.2509	-	7471B	PPB	02 Nov 2022 15:06:52	U	SMPL
N5415-05 - 1	R-18	697	0.1288	-	7471B	PPB	02 Nov 2022 15:09:08	U	SMPL
N5415-06 - 1	R-19	1759	0.3508	-	7471B	PPB	02 Nov 2022 15:11:23	U	SMPL
N5415-07 - 1	R-20	1169	0.2275	-	7471B	PPB	02 Nov 2022 15:13:39	U	SMPL
N5415-08 - 1	R-21	1994	0.3999	-	7471B	PPB	02 Nov 2022 15:15:55	U	SMPL
N5415-09 - 1	R-22	1701	0.3387	-	7471B	PPB	02 Nov 2022 15:18:12	U	SMPL
N5415-10 - 1	R-23	416	0.0701	-	7471B	PPB	02 Nov 2022 15:20:29	U	SMPL
N5415-11 - 1	R-24	2210	0.4450	-	7471B	PPB	02 Nov 2022 15:22:46	U	SMPL
N5415-12 - 1	R-25	2044	0.4103	-	7471B	PPB	02 Nov 2022 15:25:01	U	SMPL
CCV28 - 1	CCV28	26140	5.4462	-	7471B	PPB	02 Nov 2022 15:27:17	U	SMPL
CCB28 - 1	CCB28	-92	-0.0361	-	7471B	PPB	02 Nov 2022 15:29:32	U	SMPL
N5415-13 - 1	R-26	1118	0.2168	-	7471B	PPB	02 Nov 2022 15:31:48	U	SMPL
N5415-14 - 1	R-27	1443	0.2847	-	7471B	PPB	02 Nov 2022 15:34:03	U	SMPL
N5415-15 - 1	R-28	304	0.0467	-	7471B	PPB	02 Nov 2022 15:36:19	U	SMPL
N5415-16 - 1	R-29	489	0.0854	-	7471B	PPB	02 Nov 2022 15:38:35	U	SMPL
N5416-01 - 1	CC-1	4569	0.9380	-	7471B	PPB	02 Nov 2022 15:40:51	U	SMPL
PB148762BL - 1	PBS	-60	-0.0294	-	7471B	PPB	02 Nov 2022 15:43:07	U	SMPL
PB148762BS - 1	LCSS	18457	3.8405	-	7471B	PPB	02 Nov 2022 15:45:24	U	SMPL
N5395-01 - 1	S-1-103122	1189	0.2317	-	7471B	PPB	02 Nov 2022 15:47:40	U	SMPL
N5395-02 - 1	S-12-103122	5343	1.0998	-	7471B	PPB	02 Nov 2022 15:49:57	U	SMPL
N5395-03 - 1	S-11-103122	10134	2.1011	-	7471B	PPB	02 Nov 2022 15:52:13	U	SMPL
CCV29 - 1	CCV29	26322	5.4842	-	7471B	PPB	02 Nov 2022 15:54:28	U	SMPL
CCB29 - 1	CCB29	-73	-0.0321	-	7471B	PPB	02 Nov 2022 15:56:44	U	SMPL
N5395-04 - 1	S-10-103122	11224	2.3289	-	7471B	PPB	02 Nov 2022 15:58:59	U	SMPL
N5395-05 - 1	S-4-103122	12179	2.5285	-	7471B	PPB	02 Nov 2022 16:01:15	U	SMPL
N5395-06 - 1	S-5-103122	9200	1.9059	-	7471B	PPB	02 Nov 2022 16:03:30	U	SMPL
N5395-07 - 1	S-3-103122	11333	2.3517	-	7471B	PPB	02 Nov 2022 16:05:46	U	SMPL
N5395-08 - 1	S-2-103122	7389	1.5274	-	7471B	PPB	02 Nov 2022 16:08:02	U	SMPL
N5395-09 - 1	S-6-103122	9125	1.8902	-	7471B	PPB	02 Nov 2022 16:10:18	U	SMPL
N5395-10 - 1	S-7-103122	9414	1.9506	-	7471B	PPB	02 Nov 2022 16:12:34	U	SMPL
N5395-11 - 1	S-8-103122	11781	2.4453	-	7471B	PPB	02 Nov 2022 16:14:51	U	SMPL
N5395-12 - 1	S-9-103122	10378	2.1521	-	7471B	PPB	02 Nov 2022 16:17:08	U	SMPL
N5395-13 - 1	S-13-103122	11578	2.4029	-	7471B	PPB	02 Nov 2022 16:19:24	U	SMPL
CCV30 - 1	CCV30	26287	5.4769	-	7471B	PPB	02 Nov 2022 16:21:39	U	SMPL
CCB30 - 1	CCB30	-89	-0.0354	-	7471B	PPB	02 Nov 2022 16:23:55	U	SMPL
N5396-01 - 1	TP-A	7118	1.4708	-	7471B	PPB	02 Nov 2022 16:26:10	U	SMPL
N5396-04 - 1	TP-B	478	0.0831	-	7471B	PPB	02 Nov 2022 16:28:26	U	SMPL
N5396-07 - 1	TP-F	2831	0.5748	-	7471B	PPB	02 Nov 2022 16:30:42	U	SMPL
N5396-10 - 1	TP-G	181	0.0210	-	7471B	PPB	02 Nov 2022 16:32:58	U	SMPL
N5396-13 - 1	TP-I	16805	3.4952	-	7471B	PPB	02 Nov 2022 16:35:14	U	SMPL
N5396-16 - 1	TP-H	5306	1.0921	-	7471B	PPB	02 Nov 2022 16:42:24	U	SMPL
N5396-16DUP - 1	TP-HDUP	5372	1.1059	-	7471B	PPB	02 Nov 2022 16:44:40	U	SMPL
N5396-16MS - 1	TP-HMS	25725	5.3594	-	7471B	PPB	02 Nov 2022 16:46:57	U	SMPL
N5396-16MSD - 1	TP-HMSD	26905	5.6061	-	7471B	PPB	02 Nov 2022 16:49:13	U	SMPL
N5397-01 - 1	ALBANY-STREET	7850	1.6237	-	7471B	PPB	02 Nov 2022 16:51:30	U	SMPL
CCV31 - 1	CCV31	23526	4.8999	-	7471B	PPB	02 Nov 2022 16:56:22	U	SMPL
CCB31 - 1	CCB31	-56	-0.0285	-	7471B	PPB	02 Nov 2022 16:58:38	U	SMPL
N5415-01LX5 - 1		66	-0.0030	-	7471B	PPB	02 Nov 2022 17:00:54	U	SMPL
N5415-01A - 1		21527	4.4821	-	7471B	PPB	02 Nov 2022 17:03:10	U	SMPL
N5396-16LX5 - 1		960	0.1838	-	7471B	PPB	02 Nov 2022 17:05:26	U	SMPL
N5396-16A - 1		26332	5.4863	-	7471B	PPB	02 Nov 2022 17:07:42	U	SMPL
CCV32 - 1	CCV32	25756	5.3659	-	7471B	PPB	02 Nov 2022 17:12:38	U	SMPL
CCB32 - 1	CCB32	-85	-0.0346	-	7471B	PPB	02 Nov 2022 17:14:54	U	SMPL