



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 11/3/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:40
In Date: 11/02/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:29
Out Date: 11/03/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122736

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N5402-01	GBM72	1	0.92	8.92	9.84	9.66	98.0	
N5402-02	GBM72D	2	0.92	8.92	9.84	9.66	98.0	
N5402-03	GBM72S	3	0.92	8.92	9.84	9.66	98.0	
N5402-04	GBM73	4	0.96	8.70	9.66	9.11	93.7	
N5402-05	GBM74	5	0.95	8.71	9.66	9.23	95.1	
N5402-06	GBM75	6	0.96	9.01	9.97	9.87	98.9	
N5402-07	GBM76	7	0.92	8.58	9.5	9.26	97.2	
N5402-08	GBM77	8	0.93	8.73	9.66	9.22	95.0	
N5402-09	GBM78	9	0.92	8.98	9.9	9.63	97.0	
N5402-10	GBM79	10	0.93	8.66	9.59	9.47	98.6	
N5402-11	GBM80	11	0.91	8.71	9.62	9.47	98.3	
N5402-12	GBM81	12	0.95	8.76	9.71	9.56	98.3	
N5402-13	GBM82	13	0.94	8.73	9.67	9.51	98.2	
N5402-14	GBM83	14	0.92	8.78	9.7	9.08	92.9	
N5402-15	GBM84	15	0.92	8.89	9.81	9.26	93.8	
N5402-16	GBM85	16	0.91	8.88	9.79	9.36	95.2	
N5402-17	GBM86	17	0.94	9.02	9.96	9.57	95.7	
N5402-18	GBM87	18	0.94	8.77	9.71	9.63	99.1	
N5402-19	GBM88	19	0.96	8.62	9.58	9.41	98.0	
N5402-20	GBM89	20	0.96	8.59	9.55	9.41	98.4	
N5402-21	GBM90	21	0.93	8.72	9.65	9.41	97.2	
N5402-22	GBM91	22	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$