



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Sohil

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122739

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	83.0	11/03/2022	09:27
2	N5417-07		1	13	100.00	>212.0	>212.0	11/03/2022	10:00
3	N5417-07DUP		1	14	100.00	>212.0	>212.0	11/03/2022	10:30
4	N5417-08		1	12	100.00	>212.0	>212.0	11/03/2022	11:00
5	N5417-09		1	13	100.00	>212.0	>212.0	11/03/2022	11:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB122739

WORKLIST(Hardcopy Internal Chain)

WorkList Name: FP-110322

WorkList ID: 164665

Department: Wet-Chemistry

Date: 11-03-2022 09:25:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/14/2022	Water	N5417-07	Flash Point	Cool 4 deg C	PSEG03	F11	AUD-1904	11/01/2022	1010B
11/14/2022	Water	N5417-08	Flash Point	Cool 4 deg C	PSEG03	F11	AUD-1905	11/01/2022	1010B
11/14/2022	Water	N5417-09	Flash Point	Cool 4 deg C	PSEG03	F11	AUD-1906	11/01/2022	1010B

Date/Time

11/3/22 9:25

Raw Sample Received by:

SJ(WC)

Raw Sample Relinquished by:

CP(SM)

Date/Time

11/3/22 12:00

Raw Sample Received by:

CP(SM)

Raw Sample Relinquished by:

SJ(WC)