

LB122740


245.1

INSTRUMENT ID: CVL

Linear

 $\mu$  Abs.:

41783



A= 0.0000e+000

B= 2.3870e-004

C= 7.1840e-004

Rho= 0.9999878

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.0    | 0.000  | -0.001 | -0.001 | -8    | 0.000      | -8    |       |       |       |       |    |
| 0.05   | 0.050  | 0.052  | 0.002  | 213   | 0.0 %      | 213   | 0     |       |       |       | 4  |
| 0.2    | 0.200  | 0.197  | -0.003 | 823   | 0.0 %      | 823   |       |       |       |       | 2  |
| 2.5    | 2.500  | 2.495  | -0.005 | 10448 | 0.0 %      | 10448 |       |       |       |       | 0  |
| 5.0    | 5.000  | 4.993  | -0.007 | 20915 | 0.0 %      | 20915 |       |       |       |       | 0  |
| 7.5    | 7.500  | 7.540  | 0.040  | 31585 | 0.0 %      | 31585 |       |       |       |       | 1  |
| 10.0   | 10.000 | 9.974  | -0.026 | 41783 | 0.0 %      | 41783 |       |       |       |       | 0  |

## LB122740 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 03 Nov 2022 09:34:07

| Sample ID       | Extended ID | $\mu$ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|-------------|------------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 58.1        | -8         | -       | 0.0000   | 245.1  | PPB   | 03 Nov 2022 09:37:26 | S    | Std  |
| 0.05 - 1        | 58.2        | 213        | -       | 0.0500   | 245.1  | PPB   | 03 Nov 2022 09:39:43 | S    | Std  |
| 0.2 - 1         | 58.5        | 823        | -       | 0.2000   | 245.1  | PPB   | 03 Nov 2022 09:42:00 | S    | Std  |
| 2.5 - 1         | 58.5        | 10448      | -       | 2.5000   | 245.1  | PPB   | 03 Nov 2022 09:44:17 | S    | Std  |
| 5.0 - 1         | 59.5        | 20915      | -       | 5.0000   | 245.1  | PPB   | 03 Nov 2022 09:46:34 | S    | Std  |
| 7.5 - 1         | 59.8        | 31585      | -       | 7.5000   | 245.1  | PPB   | 03 Nov 2022 09:48:50 | S    | Std  |
| 10.0 - 1        |             | 41783      | -       | 10.0000  | 245.1  | PPB   | 03 Nov 2022 09:56:18 | S    | Std  |
| ICV04 - 1       | ICV04       | 16098      | 3.8434  | -245.1   |        | PPB   | 03 Nov 2022 09:58:58 | U    | SMPL |
| ICB04 - 1       | ICB04       | 93         | 0.0229  | -245.1   |        | PPB   | 03 Nov 2022 10:01:13 | U    | SMPL |
| CCV33 - 1       | CCV33       | 20752      | 4.9543  | -245.1   |        | PPB   | 03 Nov 2022 10:03:28 | U    | SMPL |
| CCB33 - 1       | CCB33       | -49        | -0.0110 | -245.1   |        | PPB   | 03 Nov 2022 10:05:43 | U    | SMPL |
| CRA - 1         | CRA         | 145        | 0.0353  | -245.1   |        | PPB   | 03 Nov 2022 10:10:42 | U    | SMPL |
| HighStd - 1     | HighStd     | 42059      | 10.0403 | -245.1   |        | PPB   | 03 Nov 2022 10:12:57 | U    | SMPL |
| ChkStd - 1      | ChkStd      | 29324      | 7.0004  | -245.1   |        | PPB   | 03 Nov 2022 10:15:13 | U    | SMPL |
| PB148783BL - 1  | PBW         | 17         | 0.0048  | -245.1   |        | PPB   | 03 Nov 2022 10:20:04 | U    | SMPL |
| PB148783BS - 1  | LCSW        | 15855      | 3.7854  | -245.1   |        | PPB   | 03 Nov 2022 10:22:19 | U    | SMPL |
| N5329-01 - 1    | BNM-1       | 66         | 0.0165  | -245.1   |        | PPB   | 03 Nov 2022 10:24:35 | U    | SMPL |
| N5329-01DUP - 1 | BNM-1DUP    | 55         | 0.0138  | -245.1   |        | PPB   | 03 Nov 2022 10:26:52 | U    | SMPL |
| N5329-01MS - 1  | BNM-1MS     | 15206      | 3.6304  | -245.1   |        | PPB   | 03 Nov 2022 10:29:08 | U    | SMPL |
| N5329-01MSD - 1 | BNM-1MSD    | 16846      | 4.0219  | -245.1   |        | PPB   | 03 Nov 2022 10:31:25 | U    | SMPL |
| N5329-02 - 1    | BNM-1       | -37        | -0.0081 | -245.1   |        | PPB   | 03 Nov 2022 10:33:41 | U    | SMPL |
| CCV34 - 1       | CCV34       | 20427      | 4.8767  | -245.1   |        | PPB   | 03 Nov 2022 10:35:56 | U    | SMPL |
| CCB34 - 1       | CCB34       | 19         | 0.0053  | -245.1   |        | PPB   | 03 Nov 2022 10:38:11 | U    | SMPL |
| N5329-03 - 1    | CBH-594     | 462        | 0.1110  | -245.1   |        | PPB   | 03 Nov 2022 10:40:27 | U    | SMPL |
| N5329-04 - 1    | CBH-594     | 146        | 0.0356  | -245.1   |        | PPB   | 03 Nov 2022 10:42:42 | U    | SMPL |
| N5329-01LX5 - 1 |             | -57        | -0.0129 | -245.1   |        | PPB   | 03 Nov 2022 10:44:58 | U    | SMPL |
| N5329-01A - 1   |             | 16580      | 3.9584  | -245.1   |        | PPB   | 03 Nov 2022 10:47:13 | U    | SMPL |
| CCV35 - 1       | CCV35       | 20401      | 4.8705  | -245.1   |        | PPB   | 03 Nov 2022 10:49:29 | U    | SMPL |
| CCB35 - 1       | CCB35       | -6         | -0.0007 | -245.1   |        | PPB   | 03 Nov 2022 10:51:45 | U    | SMPL |