

LB122744


7470A

INSTRUMENT ID: CY2

Linear

μ Abs.:

41122



A= 0.0000e+000

B= 2.4279e-004

C= -1.9242e-002

Rho= 0.9999613

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.021	-0.021	-9	0.000	-9					
0.2	0.200	0.195	-0.005	882	0.0 %	882					-3
2.5	2.500	2.536	0.036	10525	0.0 %	10525					1
5.0	5.000	4.974	-0.026	20567	0.0 %	20567					-1
7.5	7.500	7.551	0.051	31181	0.0 %	31181					1
10.0	10.000	9.965	-0.035	41122	0.0 %	41122					0

LB122744 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Nov 2022 12:18:57

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-9	-	0.0000	7470A	PPB	03 Nov 2022 12:22:52	S	Std
0.2 - 1	50.2	882	-	0.2000	7470A	PPB	03 Nov 2022 12:25:09	S	Std
2.5 - 1	52.5	10525	-	2.5000	7470A	PPB	03 Nov 2022 12:27:26	S	Std
5.0 - 1	55	20567	-	5.0000	7470A	PPB	03 Nov 2022 12:29:43	S	Std
7.5 - 1	57.5	31181	-	7.5000	7470A	PPB	03 Nov 2022 12:31:59	S	Std
10.0 - 1	51	41122	-	10.0000	7470A	PPB	03 Nov 2022 12:36:51	S	Std
ICV05 - 1	ICV05	17216	4.1607	-	7470A	PPB	03 Nov 2022 12:39:34	U	SMPL
ICB05 - 1	ICB05	49	-0.0073	-	7470A	PPB	03 Nov 2022 12:41:49	U	SMPL
CCV36 - 1	CCV36	20637	4.9913	-	7470A	PPB	03 Nov 2022 12:44:04	U	SMPL
CCB36 - 1	CCB36	-95	-0.0423	-	7470A	PPB	03 Nov 2022 12:46:20	U	SMPL
CRA - 1	CRA	754	0.1638	-	7470A	PPB	03 Nov 2022 12:48:35	U	SMPL
HighStd - 1	HighStd	41551	10.0690	-	7470A	PPB	03 Nov 2022 12:50:50	U	SMPL
ChkStd - 1	ChkStd	30134	7.2971	-	7470A	PPB	03 Nov 2022 12:53:06	U	SMPL
PB148767BL - 1	PBW	-57	-0.0331	-	7470A	PPB	03 Nov 2022 12:55:22	U	SMPL
PB148767BS - 1	LCSW	16989	4.1056	-	7470A	PPB	03 Nov 2022 12:57:38	U	SMPL
N5380-02 - 1	TP-E	-51	-0.0316	-	7470A	PPB	03 Nov 2022 12:59:55	U	SMPL
N5380-05 - 1	TP-D	-47	-0.0307	-	7470A	PPB	03 Nov 2022 13:02:11	U	SMPL
N5380-08 - 1	TP-C	-56	-0.0328	-	7470A	PPB	03 Nov 2022 13:04:29	U	SMPL
N5396-02 - 1	TP-A	-61	-0.0341	-	7470A	PPB	03 Nov 2022 13:06:45	U	SMPL
N5396-05 - 1	TP-B	-62	-0.0343	-	7470A	PPB	03 Nov 2022 13:09:01	U	SMPL
CCV37 - 1	CCV37	21320	5.1571	-	7470A	PPB	03 Nov 2022 13:11:16	U	SMPL
CCB37 - 1	CCB37	-89	-0.0409	-	7470A	PPB	03 Nov 2022 13:13:31	U	SMPL
N5396-08 - 1	TP-F	0	-0.0192	-	7470A	PPB	03 Nov 2022 13:15:46	U	SMPL
N5396-11 - 1	TP-G	0	-0.0192	-	7470A	PPB	03 Nov 2022 13:18:02	U	SMPL
N5396-14 - 1	TP-I	-8	-0.0212	-	7470A	PPB	03 Nov 2022 13:20:18	U	SMPL
N5396-17 - 1	TP-H	-25	-0.0253	-	7470A	PPB	03 Nov 2022 13:22:34	U	SMPL
N5397-03 - 1	ALBANY-STREET	-26	-0.0256	-	7470A	PPB	03 Nov 2022 13:24:49	U	SMPL
N5406-02 - 1	SB-01-COMP	-14	-0.0226	-	7470A	PPB	03 Nov 2022 13:27:06	U	SMPL
N5406-04 - 1	SB-02-COMP	-4	-0.0202	-	7470A	PPB	03 Nov 2022 13:29:22	U	SMPL
N5406-06 - 1	SB-03-COMP	-4	-0.0202	-	7470A	PPB	03 Nov 2022 13:31:39	U	SMPL
N5412-01 - 1	D-2757	6	-0.0178	-	7470A	PPB	03 Nov 2022 13:33:57	U	SMPL
N5412-02 - 1	D-2770	8	-0.0173	-	7470A	PPB	03 Nov 2022 13:36:13	U	SMPL
CCV38 - 1	CCV38	20832	5.0386	-	7470A	PPB	03 Nov 2022 13:38:29	U	SMPL
CCB38 - 1	CCB38	-82	-0.0392	-	7470A	PPB	03 Nov 2022 13:40:44	U	SMPL
N5412-02DUP - 1	D-2770DUP	31	-0.0117	-	7470A	PPB	03 Nov 2022 13:43:00	U	SMPL
N5412-02MS - 1	D-2770MS	16897	4.0832	-	7470A	PPB	03 Nov 2022 13:45:15	U	SMPL
N5412-02MSD - 1	D-2770MSD	16208	3.9159	-	7470A	PPB	03 Nov 2022 13:47:31	U	SMPL
PB148768BL - 1	PBW	-52	-0.0319	-	7470A	PPB	03 Nov 2022 13:49:46	U	SMPL
PB148768BS - 1	LCSW	17093	4.1308	-	7470A	PPB	03 Nov 2022 13:52:02	U	SMPL
N5412-03 - 1	D-2748	64	-0.0037	-	7470A	PPB	03 Nov 2022 13:54:18	U	SMPL
N5412-04 - 1	D-2763	-53	-0.0321	-	7470A	PPB	03 Nov 2022 13:56:34	U	SMPL
N5412-04DUP - 1	D-2763DUP	-39	-0.0287	-	7470A	PPB	03 Nov 2022 13:58:51	U	SMPL
N5412-04MS - 1	D-2763MS	16715	4.0390	-	7470A	PPB	03 Nov 2022 14:01:08	U	SMPL
N5412-04MSD - 1	D-2763MSD	16184	3.9101	-	7470A	PPB	03 Nov 2022 14:03:24	U	SMPL
CCV39 - 1	CCV39	19035	4.6023	-	7470A	PPB	03 Nov 2022 14:05:39	U	SMPL
CCB39 - 1	CCB39	-38	-0.0285	-	7470A	PPB	03 Nov 2022 14:07:55	U	SMPL
PB148784BL - 1	PBW	60	-0.0047	-	7470A	PPB	03 Nov 2022 14:10:10	U	SMPL
PB148784BS - 1	LCSW	15880	3.8363	-	7470A	PPB	03 Nov 2022 14:12:25	U	SMPL
N5304-01 - 1	DEEP-WELL	692	0.1488	-	7470A	PPB	03 Nov 2022 14:14:41	U	SMPL
N5304-01DUP - 1	DEEP-WELLDUP	386	0.0745	-	7470A	PPB	03 Nov 2022 14:16:56	U	SMPL
N5304-01MS - 1	DEEP-WELLS	17152	4.1451	-	7470A	PPB	03 Nov 2022 14:23:46	U	SMPL
N5304-01MSD - 1	DEEP-WELLSMSD	17413	4.2085	-	7470A	PPB	03 Nov 2022 14:26:02	U	SMPL
N5306-01 - 1	20221166-COMP-01-MODIFIED-ELUTRIATE	34	-0.0110	-	7470A	PPB	03 Nov 2022 14:28:19	U	SMPL
N5306-02 - 1	20221167-COMP-02-MODIFIED-ELUTRIATE	76	-0.0008	-	7470A	PPB	03 Nov 2022 14:30:35	U	SMPL
N5306-03 - 1	20221168-COMP-03-MODIFIED-ELUTRIATE	-16	-0.0231	-	7470A	PPB	03 Nov 2022 14:32:52	U	SMPL
N5306-04 - 1	20221173-COMP-08-MODIFIED-ELUTRIATE	5	-0.0180	-	7470A	PPB	03 Nov 2022 14:35:08	U	SMPL
CCV40 - 1	CCV40	20812	5.0337	-	7470A	PPB	03 Nov 2022 14:37:23	U	SMPL
CCB40 - 1	CCB40	-126	-0.0498	-	7470A	PPB	03 Nov 2022 14:39:39	U	SMPL
N5306-05 - 1	20221174-COMP-09-MODIFIED-ELUTRIATE	82	0.0007	-	7470A	PPB	03 Nov 2022 14:41:54	U	SMPL
N5338-01 - 1	COMP(67-69)TOTE	1304	0.2974	-	7470A	PPB	03 Nov 2022 14:44:10	U	SMPL
N5360-10 - 1	20221169-COMP-04-MODIFIED-ELUTRIATE	42	-0.0090	-	7470A	PPB	03 Nov 2022 14:46:25	U	SMPL
N5360-11 - 1	20221171-COMP-06-MODIFIED-ELUTRIATE	86	0.0016	-	7470A	PPB	03 Nov 2022 14:48:41	U	SMPL
N5360-12 - 1	20221172-COMP-07-MODIFIED-ELUTRIATE	-21	-0.0243	-	7470A	PPB	03 Nov 2022 14:50:56	U	SMPL
N5360-13 - 1	20221175-COMP-10-MODIFIED-ELUTRIATE	-78	-0.0382	-	7470A	PPB	03 Nov 2022 14:53:13	U	SMPL
N5360-14 - 1	20221179-COMP-14-MODIFIED-ELUTRIATE	23	-0.0137	-	7470A	PPB	03 Nov 2022 14:55:29	U	SMPL
N5381-01 - 1	265393-7	-146	-0.0547	-	7470A	PPB	03 Nov 2022 14:57:46	U	SMPL
N5381-02 - 1	265393-8	-69	-0.0360	-	7470A	PPB	03 Nov 2022 15:00:03	U	SMPL
N5383-01 - 1	256317-7	119	0.0097	-	7470A	PPB	03 Nov 2022 15:02:19	U	SMPL
CCV41 - 1	CCV41	21365	5.1680	-	7470A	PPB	03 Nov 2022 15:04:36	U	SMPL
CCB41 - 1	CCB41	-80	-0.0387	-	7470A	PPB	03 Nov 2022 15:06:52	U	SMPL
N5383-02 - 1	256317-8	175	0.0232	-	7470A	PPB	03 Nov 2022 15:09:08	U	SMPL
N5385-01 - 1	251563-5	125	0.0111	-	7470A	PPB	03 Nov 2022 15:11:24	U	SMPL
N5385-02 - 1	251563-6	93	0.0033	-	7470A	PPB	03 Nov 2022 15:13:40	U	SMPL
N5404-01 - 1	551B-3	39	-0.0098	-	7470A	PPB	03 Nov 2022 15:15:56	U	SMPL
N5404-02 - 1	551B-4	72	-0.0018	-	7470A	PPB	03 Nov 2022 15:18:12	U	SMPL
PB148712TB - 1	PB148712TB	-9	-0.0214	-	7470A	PPB	03 Nov 2022 15:20:28	U	SMPL
PB148746TB - 1	PB148746TB	0	-0.0192	-	7470A	PPB	03 Nov 2022 15:22:44	U	SMPL
CCV42 - 1	CCV42	21536	5.2095	-	7470A	PPB	03 Nov 2022 15:25:01	U	SMPL
CCB42 - 1	CCB42	-74	-0.0372	-	7470A	PPB	03 Nov 2022 15:27:18	U	SMPL
N5412-02LX5 - 1		-17	-0.0234	-	7470A	PPB	03 Nov 2022 15:29:34	U	SMPL
N5412-02A - 1		17059	4.1225	-	7470A	PPB	03 Nov 2022 15:31:50	U	SMPL
N5412-04LX5 - 1		-85	-0.0409	-	7470A	PPB	03 Nov 2022 15:34:07	U	SMPL
N5412-04A - 1		17091	4.1303	-	7470A	PPB	03 Nov 2022 15:36:22	U	SMPL
N5304-01LX5 - 1		33	-0.0112	-	7470A	PPB	03 Nov 2022 15:38:39	U	SMPL
N5304-01A - 1		17274	4.1747	-	7470A	PPB	03 Nov 2022 15:40:55	U	SMPL
CCV43 - 1	CCV43	21387	5.1734	-	7470A	PPB	03 Nov 2022 15:43:11	U	SMPL
CCB43 - 1	CCB43	-49	-0.0311	-	7470A	PPB	03 Nov 2022 15:45:27	U	SMPL