



PERCENT SOLID

Supervisor: Iwona

Analyst: JIGNESH

Date: 11/4/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 17:10

In Date: 11/03/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:25

Out Date: 11/04/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB122746

| Lab ID   | Client SampleID     | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments        |
|----------|---------------------|--------|----------------|--------------|-------------------------|---------------------------|---------|-----------------|
| N5435-01 | PTS-TP-COMP-1103    | 29     | 0.91           | 9.04         | 9.95                    | 8.53                      | 84.3    |                 |
| N5435-03 | PTS-SOIL-PILE-1103  | 30     | 0.96           | 9.01         | 9.97                    | 9.27                      | 92.2    |                 |
| N5435-05 | PTS-CON-1103        | 31     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CONCRETE SAMPLE |
| N5436-01 | 250FERRYBLVD-TFS-01 | 1      | 0.91           | 8.95         | 9.86                    | 8.18                      | 81.2    |                 |
| N5436-02 | 250FERRYBLVD-TFS-02 | 2      | 0.92           | 8.94         | 9.86                    | 8.14                      | 80.8    |                 |
| N5436-03 | 250FERRYBLVD-TFS-03 | 3      | 0.97           | 8.70         | 9.67                    | 8.04                      | 81.3    |                 |
| N5436-04 | 250FERRYBLVD-TFS-04 | 4      | 0.91           | 9.03         | 9.94                    | 7.98                      | 78.3    |                 |
| N5436-05 | 250FERRYBLVD-TFS-05 | 5      | 0.92           | 8.72         | 9.64                    | 7.94                      | 80.5    |                 |
| N5436-06 | 250FERRYBLVD-TFS-06 | 6      | 0.91           | 8.85         | 9.76                    | 7.92                      | 79.2    |                 |
| N5436-07 | 250FERRYBLVD-TFS-07 | 7      | 0.91           | 8.64         | 9.55                    | 7.91                      | 81.0    |                 |
| N5436-08 | 250FERRYBLVD-TFS-08 | 8      | 0.9            | 9.06         | 9.96                    | 8.24                      | 81.0    |                 |
| N5436-09 | 250FERRYBLVD-TFS-09 | 9      | 0.91           | 8.69         | 9.6                     | 7.88                      | 80.2    |                 |
| N5436-10 | 250FERRYBLVD-TFS-10 | 10     | 0.92           | 8.73         | 9.65                    | 8.00                      | 81.1    |                 |
| N5436-11 | 250FERRYBLVD-TFS-11 | 11     | 0.9            | 8.74         | 9.64                    | 7.98                      | 81.0    |                 |
| N5436-12 | 250FERRYBLVD-TFS-12 | 12     | 0.91           | 8.63         | 9.54                    | 7.94                      | 81.5    |                 |
| N5436-13 | 250FERRYBLVD-TFS-13 | 13     | 0.91           | 8.66         | 9.57                    | 7.98                      | 81.6    |                 |
| N5449-01 | COMP-1              | 14     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | debris          |
| N5449-02 | 17203               | 15     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | debris          |
| N5450-01 | BC-279313-FRONT     | 16     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N5450-02 | BC-279313-END       | 17     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N5450-03 | SOM-2022-92A        | 18     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N5450-04 | SOM-2022-92B        | 19     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N5450-05 | SOM-2022-91A        | 20     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC REEL       |
| N5451-01 | LFR-SOIL-1103       | 32     | 0.91           | 8.81         | 9.72                    | 6.11                      | 59.0    |                 |
| N5452-01 | SOIL-PILE           | 21     | 0.93           | 8.85         | 9.78                    | 8.63                      | 87.0    |                 |
| N5453-01 | TR-6-110322         | 33     | 0.91           | 8.86         | 9.77                    | 8.85                      | 89.6    |                 |
| N5453-02 | TR-6-110322-EPH-2   | 34     | 0.91           | 9.07         | 9.98                    | 9.00                      | 89.2    |                 |
| N5459-01 | EO-03-110322        | 22     | 0.91           | 8.64         | 9.55                    | 8.83                      | 91.7    |                 |



PERCENT SOLID

Supervisor: Iwona  
Analyst: JIGNESH  
Date: 11/4/2022

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:10  
In Date: 11/03/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:25  
Out Date: 11/04/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB122746

| Lab ID   | Client SampleID    | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|--------------------|--------|----------------|--------------|-------------------------|---------------------------|---------|----------|
| N5459-02 | EO-03-110322-EPH-2 | 23     | 0.91           | 8.64         | 9.55                    | 8.83                      | 91.7    |          |
| N5459-03 | EO-02-110322       | 24     | 0.95           | 8.57         | 9.52                    | 8.72                      | 90.7    |          |
| N5459-04 | EO-02-110322-EPH-2 | 25     | 0.95           | 8.57         | 9.52                    | 8.72                      | 90.7    |          |
| N5460-01 | SH-1               | 26     | 0.91           | 8.73         | 9.64                    | 8.93                      | 91.9    |          |
| N5461-01 | TJRC-110222-G1     | 27     | 0.92           | 8.59         | 9.51                    | 8.76                      | 91.3    |          |
| N5461-02 | TJRC-110222-G2     | 28     | 0.98           | 8.71         | 9.69                    | 8.91                      | 91.0    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

# WORKLIST(Hardcopy Internal Chain)

13122746

WorkList Name : %1-111322

WorkList ID : 164656

Department : Wet-Chemistry

Date : 11-03-2022 08:23:40

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Raw Sample Storage Location | Customer Sample     | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|-----------------------------|---------------------|--------------|--------------|
| 11/08/2022 | Solid  | N5460-01 | Percent Solids | Cool 4 deg C | EARTH03  | F11                         | SH-1                | 11/03/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5436-01 | Percent Solids | Cool 4 deg C | NOBI03   | F11                         | 250FERRYBLVD-TFS-01 | 11/02/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5436-02 | Percent Solids | Cool 4 deg C | NOBI03   | F11                         | 250FERRYBLVD-TFS-02 | 11/02/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5436-03 | Percent Solids | Cool 4 deg C | NOBI03   | F11                         | 250FERRYBLVD-TFS-03 | 11/02/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5436-04 | Percent Solids | Cool 4 deg C | NOBI03   | F11                         | 250FERRYBLVD-TFS-04 | 11/02/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5436-05 | Percent Solids | Cool 4 deg C | NOBI03   | F11                         | 250FERRYBLVD-TFS-05 | 11/02/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5436-06 | Percent Solids | Cool 4 deg C | NOBI03   | F11                         | 250FERRYBLVD-TFS-06 | 11/02/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5436-13 | Percent Solids | Cool 4 deg C | NOBI03   | F11                         | 250FERRYBLVD-TFS-13 | 11/02/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5436-07 | Percent Solids | Cool 4 deg C | NOBI03   | F11                         | 250FERRYBLVD-TFS-07 | 11/02/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5436-08 | Percent Solids | Cool 4 deg C | NOBI03   | F11                         | 250FERRYBLVD-TFS-08 | 11/02/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5436-09 | Percent Solids | Cool 4 deg C | NOBI03   | F11                         | 250FERRYBLVD-TFS-09 | 11/02/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5436-10 | Percent Solids | Cool 4 deg C | NOBI03   | F11                         | 250FERRYBLVD-TFS-10 | 11/02/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5436-11 | Percent Solids | Cool 4 deg C | NOBI03   | F11                         | 250FERRYBLVD-TFS-11 | 11/02/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5436-12 | Percent Solids | Cool 4 deg C | NOBI03   | F11                         | 250FERRYBLVD-TFS-12 | 11/02/2022   | Chemtech -SO |
| 11/08/2022 | Solid  | N5435-01 | Percent Solids | Cool 4 deg C | PSEG03   | F11                         | PTS-TP-COMP-1103    | 11/03/2022   | Chemtech -SO |
| 11/08/2022 | Solid  | N5435-03 | Percent Solids | Cool 4 deg C | PSEG03   | F11                         | PTS-SOIL-PILE-1103  | 11/03/2022   | Chemtech -SO |
| 11/08/2022 | Solid  | N5435-05 | Percent Solids | Cool 4 deg C | PSEG03   | F11                         | PTS-CON-1103        | 11/03/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5449-01 | Percent Solids | Cool 4 deg C | PSEG03   | F11                         | COMP-1              | 11/03/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5449-02 | Percent Solids | Cool 4 deg C | PSEG03   | F11                         | 17203               | 11/03/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5450-01 | Percent Solids | Cool 4 deg C | PSEG03   | F11                         | BC-279313-FRONT     | 11/03/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5450-02 | Percent Solids | Cool 4 deg C | PSEG03   | F11                         | BC-279313-END       | 11/03/2022   | Chemtech -SO |

Date/Time 11/03/22 16:10 Date/Time 11/03/22 Date/Time 17:20  
 Raw Sample Received by: JP/arc Raw Sample Received by: CL Raw Sample Received by: CL  
 Raw Sample Relinquished by: CL Raw Sample Relinquished by: JP/arc Raw Sample Relinquished by: JP/arc

# WORKLIST(Hardcopy Internal Chain)

13122746

WorkList Name : %1-111322

WorkList ID : 164656

Department : Wet-Chemistry

Date : 11-03-2022 08:23:40

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Raw Sample Storage Location | Customer Sample    | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|-----------------------------|--------------------|--------------|--------------|
| 11/10/2022 | Solid  | N5450-03 | Percent Solids | Cool 4 deg C | PSEG03   | F11                         | SOM-2022-92A       | 11/03/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5450-04 | Percent Solids | Cool 4 deg C | PSEG03   | F11                         | SOM-2022-92B       | 11/03/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5450-05 | Percent Solids | Cool 4 deg C | PSEG03   | F11                         | SOM-2022-91A       | 11/03/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5451-01 | Percent Solids | Cool 4 deg C | PSEG03   | F11                         | LFR-SOIL-1103      | 11/03/2022   | Chemtech -SO |
| 11/10/2022 | Solid  | N5452-01 | Percent Solids | Cool 4 deg C | PSEG03   | F11                         | SOIL-PILE          | 11/03/2022   | Chemtech -SO |
| 11/08/2022 | Solid  | N5453-01 | Percent Solids | Cool 4 deg C | PSEG05   | F11                         | TR-6-110322        | 11/03/2022   | Chemtech -SO |
| 11/08/2022 | Solid  | N5453-02 | Percent Solids | Cool 4 deg C | PSEG05   | F11                         | TR-6-110322-EPH-2  | 11/03/2022   | Chemtech -SO |
| 11/08/2022 | Solid  | N5459-01 | Percent Solids | Cool 4 deg C | PSEG05   | F11                         | EO-03-110322       | 11/03/2022   | Chemtech -SO |
| 11/08/2022 | Solid  | N5459-02 | Percent Solids | Cool 4 deg C | PSEG05   | F11                         | EO-03-110322-EPH-2 | 11/03/2022   | Chemtech -SO |
| 11/08/2022 | Solid  | N5459-03 | Percent Solids | Cool 4 deg C | PSEG05   | F11                         | EO-02-110322       | 11/03/2022   | Chemtech -SO |
| 11/08/2022 | Solid  | N5459-04 | Percent Solids | Cool 4 deg C | PSEG05   | F11                         | EO-02-110322-EPH-2 | 11/03/2022   | Chemtech -SO |
| 11/12/2022 | Solid  | N5461-01 | Percent Solids | Cool 4 deg C | ENTE01   | F11                         | TJRC-110222-G1     | 11/02/2022   | Chemtech -SO |
| 11/12/2022 | Solid  | N5461-02 | Percent Solids | Cool 4 deg C | ENTE01   | F11                         | TJRC-110222-G2     | 11/02/2022   | Chemtech -SO |

Date/Time 11/03/22 16:10  
 Raw Sample Received by: JP (COC)  
 Raw Sample Relinquished by: CP SR

Date/Time 11/03/22 17:20  
 Raw Sample Received by: CP SR  
 Raw Sample Relinquished by: JP (COC)