



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: Sohil

Run Number: LB122747

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP99684
calibration std. phosphate 0.5 ppm	WP99683
calibration std. phosphate 0.3 ppm	WP99682
calibration std. phosphate 0.1 ppm	WP99681
calibration std. phosphate 0.05 ppm	WP99680
calibration std. 0 ppm	WP99679
phosphate CCV std.	WP99685
Combined reagent	WP99759
Phenolphthalein indicator	WP96861
Sodium hydroxide, 1N	WP98644
Phosphate ICV-LCS Std	WP99686

Intercept: -0.0018

Slope: 0.6533

Regression: 0.999959

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.003		11/03/2022	13:00
2	CAL2	0.05	1	50	50	0.032	0.052	-4	11/03/2022	13:00
3	CAL3	0.10	1	50	50	0.064	0.101	-1	11/03/2022	13:01
4	CAL4	0.30	1	50	50	0.190	0.294	2	11/03/2022	13:01
5	CAL5	0.50	1	50	50	0.324	0.499	0.2	11/03/2022	13:02
6	CAL6	1.00	1	50	50	0.653	1.002	-0.2	11/03/2022	13:03



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: Sohil

Run Number: LB122747

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.321	0.494	11/03/2022	13:05
2	ICB		1	50	50	0.001	0.004	11/03/2022	13:05
3	CCV1	0.50	1	50	50	0.326	0.502	11/03/2022	13:06
4	CCB1		1	50	50	0.002	0.006	11/03/2022	13:06
5	PB148788BL		1	50	50	0.003	0.007	11/03/2022	13:07
6	PB148788BS	0.50	1	50	50	0.334	0.514	11/03/2022	13:07
7	N5295-01		1	50	50	1.067	1.636	11/03/2022	13:08
8	N5297-01		1	50	50	0.638	0.979	11/03/2022	13:08
9	N5297-01DUP		1	50	50	0.637	0.978	11/03/2022	13:09
10	N5297-01MS	0.50	1	50	50	0.967	1.483	11/03/2022	13:09
11	N5297-01MSD	0.50	1	50	50	0.969	1.486	11/03/2022	13:10
12	N5363-01		1	50	50	1.464	2.244	11/03/2022	13:10
13	N5295-01		10	50	50	0.261	0.402	11/03/2022	13:11
14	N5297-01MS	0.50	2	50	50	0.463	0.711	11/03/2022	13:11
15	CCV2	0.50	1	50	50	0.318	0.490	11/03/2022	13:12
16	CCB2		1	50	50	0.002	0.006	11/03/2022	13:12
17	N5297-01MSD	0.50	2	50	50	0.460	0.707	11/03/2022	13:13
18	N5363-01		10	50	50	0.420	0.646	11/03/2022	13:13
19	CCV3	0.50	1	50	50	0.323	0.497	11/03/2022	13:14
20	CCB3		1	50	50	0.003	0.007	11/03/2022	13:14