

Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: Sohil

Run Number: LB122748

BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP99684
calibration std. phosphate 0.5 ppm	WP99683
calibration std. phosphate 0.3 ppm	WP99682
calibration std. phosphate 0.1 ppm	WP99681
calibration std. phosphate 0.05 ppm	WP99680
calibration std. 0 ppm	WP99679
phosphate CCV std.	WP99685
Combined reagent	WP99759
Phenolphthalein indicator	WP96861
Sodium hydroxide, 1N	WP98644
Phosphate ICV-LCS Std	WP99686

Intercept: -0.0018

Slope: 0.6533

Regression: 0.999959

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.003		11/03/2022	13:00
2	CAL2	0.05	1	50	50	0.032	0.052	-4	11/03/2022	13:00
3	CAL3	0.10	1	50	50	0.064	0.101	-1	11/03/2022	13:01
4	CAL4	0.30	1	50	50	0.190	0.294	2	11/03/2022	13:01
5	CAL5	0.50	1	50	50	0.324	0.499	0.2	11/03/2022	13:02
6	CAL6	1.00	1	50	50	0.653	1.002	-0.2	11/03/2022	13:03



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: Sohil

Run Number: LB122748

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.321	0.494	11/03/2022	13:05
2	ICB		1	50	50	0.001	0.004	11/03/2022	13:05
3	CCV1	0.50	1	50	50	0.328	0.505	11/03/2022	13:25
4	CCB1		1	50	50	0.003	0.007	11/03/2022	13:25
5	PB148796BL		1	1.00	50	0.002	0.006	11/03/2022	13:26
6	PB148796BS	0.50	1	1.00	50	0.336	0.517	11/03/2022	13:27
7	N5317-08		1	1.00	50	1.961	3.004	11/03/2022	13:27
8	N5317-08		50	1.00	50	0.606	0.930	11/03/2022	13:28
9	CCV2	0.50	1	50	50	0.324	0.499	11/03/2022	13:28
10	CCB2		1	50	50	0.001	0.004	11/03/2022	13:29