



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB122752

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	11/04/2022	08:00
2	CAL2	1	Water	NA	NA	20.7	7.00	11/04/2022	08:02
3	CAL3	1	Water	NA	NA	20.7	10.03	11/04/2022	08:03
4	ICV	1	Water	NA	NA	20.7	6.98	11/04/2022	08:05
5	CCV1	1	Water	NA	NA	20.7	2.01	11/04/2022	08:10
6	N5437-01	1	Water	NA	NA	20.1	7.43	11/04/2022	08:11
7	N5437-01DUP	1	Water	NA	NA	20.1	7.44	11/04/2022	08:12
8	N5437-02	1	Water	NA	NA	20.1	7.52	11/04/2022	08:15
9	N5439-01	1	Water	NA	NA	20.3	7.79	11/04/2022	08:19
10	N5439-02	1	Water	NA	NA	20.3	7.80	11/04/2022	08:20
11	N5441-01	1	Water	NA	NA	20.1	7.83	11/04/2022	08:22
12	N5441-02	1	Water	NA	NA	20.1	7.85	11/04/2022	08:24
13	N5454-01	1	Water	NA	NA	24.1	9.84	11/04/2022	08:25
14	N5454-02	1	Water	NA	NA	24.1	6.93	11/04/2022	08:30
15	CCV2	1	Water	NA	NA	20.7	12.03	11/04/2022	08:33

UB 12152

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n5457 WorkList ID : 164700 Department : Wet-Chemistry Date : 11-04-2022 07:37:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/02/2022	Water	N5437-01	pH <u>E</u>	Cool 4 deg C	ENV170	F11	596C-3	11/02/2022	9040C
11/02/2022	Water	N5437-02	pH <u>E</u>	Cool 4 deg C	ENV170	F11	596C-4	11/02/2022	9040C
11/02/2022	Water	N5439-01	pH <u>E</u>	Cool 4 deg C	ENV170	F11	506D-3	11/02/2022	9040C
11/02/2022	Water	N5439-02	pH <u>E</u>	Cool 4 deg C	ENV170	F11	506D-4	11/02/2022	9040C
11/02/2022	Water	N5441-01	pH <u>E</u>	Cool 4 deg C	ENV170	F11	5486-5	11/02/2022	9040C
11/02/2022	Water	N5441-02	pH <u>E</u>	Cool 4 deg C	ENV170	F11	5486-6	11/02/2022	9040C
11/03/2022	Water	N5454-01	pH <u>C</u>	Cool 4 deg C	PSEG03	F11	TRE-1160	11/03/2022	9040C
11/03/2022	Water	N5454-02	pH <u>C</u>	Cool 4 deg C	PSEG03	F11	TRE-1674	11/03/2022	9040C

Date/Time
Raw Sample Received by:
Raw Sample Relinquished by:

11/04/22 07:40
JD (DC)
CD SR

Date/Time
Raw Sample Received by:
Raw Sample Relinquished by:

11/04/22 09:145
CD SR
JD (DC)