

LB122757


7470A

INSTRUMENT ID: CVL

Linear

μ Abs.:

47520



A= 0.0000e+000

B= 2.1322e-004

C= -7.3381e-003

Rho= 0.9994752

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	U/eD
0.0	0.000	-0.008	-0.008	-1	0.000	-1					
0.2	0.200	0.170	-0.030	834	0.0 %	834					-15
2.5	2.500	2.672	0.172	12564	0.0 %	12564					7
5.0	5.000	4.925	-0.075	23132	0.0 %	23132					-2
7.5	7.500	7.315	-0.185	34343	0.0 %	34343					-2
10.0	10.000	10.125	0.125	47520	0.0 %	47520					1

LB122757 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Nov 2022 11:28:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-1	-	0.0000	7470A	PPB	04 Nov 2022 11:36:01	S	Std
0.2 - 1	50.2	834	-	0.2000	7470A	PPB	04 Nov 2022 11:38:18	S	Std
2.5 - 1	52.5	12564	-	2.5000	7470A	PPB	04 Nov 2022 11:40:35	S	Std
5.0 - 1	55	23132	-	5.0000	7470A	PPB	04 Nov 2022 11:42:52	S	Std
7.5 - 1	57.5	34343	-	7.5000	7470A	PPB	04 Nov 2022 11:45:09	S	Std
10.0 - 1	59	47520	-	10.0000	7470A	PPB	04 Nov 2022 11:47:24	S	Std
ICV06 - 1	ICV06	18921	4.0271	-7470A	PPB	04 Nov 2022 11:52:53	U	SMPL	
ICB06 - 1	ICB06	25	-0.0020	-7470A	PPB	04 Nov 2022 11:55:07	U	SMPL	
CCV44 - 1	CCV44	23141	4.9269	-7470A	PPB	04 Nov 2022 11:57:22	U	SMPL	
CCB44 - 1	CCB44	-49	-0.0178	-7470A	PPB	04 Nov 2022 11:59:37	U	SMPL	
CRA - 1	CRA	812	0.1658	-7470A	PPB	04 Nov 2022 12:01:52	U	SMPL	
HighStd - 1	HighStd	47222	10.0615	-7470A	PPB	04 Nov 2022 12:04:07	U	SMPL	
ChkStd - 1	ChkStd	32384	6.8977	-7470A	PPB	04 Nov 2022 12:06:22	U	SMPL	
PB148798BL - 1	PBW	-38	-0.0154	-7470A	PPB	04 Nov 2022 12:08:38	U	SMPL	
PB148798BS - 1	LCSW	18061	3.8437	-7470A	PPB	04 Nov 2022 12:10:54	U	SMPL	
N5418-02 - 1	TP-J	-4	-0.0082	-7470A	PPB	04 Nov 2022 12:13:10	U	SMPL	
N5418-02DUP - 1	TP-JDUP	14	-0.0044	-7470A	PPB	04 Nov 2022 12:15:26	U	SMPL	
N5418-02MS - 1	TP-JMS	17222	3.6648	-7470A	PPB	04 Nov 2022 12:17:43	U	SMPL	
N5418-02MSD - 1	TP-JMSD	16578	3.5275	-7470A	PPB	04 Nov 2022 12:20:01	U	SMPL	
N5418-05 - 1	TP-K	-12	-0.0099	-7470A	PPB	04 Nov 2022 12:22:16	U	SMPL	
CCV45 - 1	CCV45	22822	4.8589	-7470A	PPB	04 Nov 2022 12:24:31	U	SMPL	
CCB45 - 1	CCB45	-8	-0.0090	-7470A	PPB	04 Nov 2022 12:26:46	U	SMPL	
N5418-08 - 1	TP-L	32	-0.0005	-7470A	PPB	04 Nov 2022 12:29:01	U	SMPL	
N5418-11 - 1	TP-M	28	-0.0014	-7470A	PPB	04 Nov 2022 12:31:17	U	SMPL	
N5418-14 - 1	TP-N	17	-0.0037	-7470A	PPB	04 Nov 2022 12:33:32	U	SMPL	
N5418-17 - 1	TP-O	-25	-0.0127	-7470A	PPB	04 Nov 2022 12:35:48	U	SMPL	
N5425-02 - 1	371	-49	-0.0178	-7470A	PPB	04 Nov 2022 12:38:03	U	SMPL	
N5418-02LX5 - 1		-35	-0.0148	-7470A	PPB	04 Nov 2022 12:40:20	U	SMPL	
N5418-02A - 1		18829	4.0075	-7470A	PPB	04 Nov 2022 12:42:36	U	SMPL	
PB148770TB - 1	PB148770TB	-42	-0.0163	-7470A	PPB	04 Nov 2022 12:44:53	U	SMPL	
CV46 - 1	CCV46	23134	4.9254	-7470A	PPB	04 Nov 2022 12:47:10	U	SMPL	
CB46 - 1	CCB46	-27	-0.0131	-7470A	PPB	04 Nov 2022 12:49:25	U	SMPL	