

LB122759


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

52374



A= 0.0000e+000

B= 1.8894e-004

C= 3.0835e-002

Rho= 0.9998494

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.027	0.027	-19	0.000	-19					
0.2	0.200	0.188	-0.012	832	0.0 %	832					-6
2.5	2.500	2.457	-0.043	12842	0.0 %	12842					-2
5.0	5.000	4.975	-0.025	26168	0.0 %	26168					-1
7.5	7.500	7.626	0.126	40198	0.0 %	40198					2
10.0	10.000	9.926	-0.074	52374	0.0 %	52374					-1

LB122759 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Nov 2022 14:16:25

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-19	-	0.0000	7471B	PPB	04 Nov 2022 14:22:24	S	Std
0.2 - 1	50.2	832	-	0.2000	7471B	PPB	04 Nov 2022 14:24:41	S	Std
2.5 - 1	52.5	12842	-	2.5000	7471B	PPB	04 Nov 2022 14:26:58	S	Std
5.0 - 1	55	26168	-	5.0000	7471B	PPB	04 Nov 2022 14:29:15	S	Std
7.5 - 1	57.5	40198	-	7.5000	7471B	PPB	04 Nov 2022 14:31:31	S	Std
10.0 - 1	510	52374	-	10.0000	7471B	PPB	04 Nov 2022 14:33:47	S	Std
ICV07 - 1	ICV07	21355	4.0657	-7471B	PPB	04 Nov 2022 14:36:32	U	SMPL	
ICB07 - 1	ICB07	7	0.0322	-7471B	PPB	04 Nov 2022 14:38:47	U	SMPL	
CCV47 - 1	CCV47	27005	5.1332	-7471B	PPB	04 Nov 2022 14:41:02	U	SMPL	
CCB47 - 1	CCB47	-70	0.0176	-7471B	PPB	04 Nov 2022 14:43:17	U	SMPL	
CRA - 1	CRA	850	0.1914	-7471B	PPB	04 Nov 2022 14:45:32	U	SMPL	
HighStd - 1	HighStd	52824	10.0115	-7471B	PPB	04 Nov 2022 14:47:47	U	SMPL	
ChkStd - 1	ChkStd	38283	7.2641	-7471B	PPB	04 Nov 2022 14:50:03	U	SMPL	
PB148807BL - 1	PBS	-125	0.0072	-7471B	PPB	04 Nov 2022 14:52:20	U	SMPL	
PB148807BS - 1	LCSS	19809	3.7736	-7471B	PPB	04 Nov 2022 14:54:36	U	SMPL	
N5318-01 - 1	20221105-COMP-04	8208	1.5817	-7471B	PPB	04 Nov 2022 14:56:52	U	SMPL	
N5318-02 - 1	20221106-COMP-05	13690	2.6175	-7471B	PPB	04 Nov 2022 14:59:08	U	SMPL	
N5318-03 - 1	20221107-COMP-06	10747	2.0614	-7471B	PPB	04 Nov 2022 15:01:25	U	SMPL	
N5318-04 - 1	20221108-COMP-07	6040	1.1720	-7471B	PPB	04 Nov 2022 15:03:42	U	SMPL	
N5318-05 - 1	20221111-COMP-10	8994	1.7302	-7471B	PPB	04 Nov 2022 15:05:57	U	SMPL	
CCV48 - 1	CCV48	26574	5.0518	-7471B	PPB	04 Nov 2022 15:08:12	U	SMPL	
CCB48 - 1	CCB48	-105	0.0110	-7471B	PPB	04 Nov 2022 15:10:27	U	SMPL	
N5318-06 - 1	20221115-COMP-14	11720	2.2452	-7471B	PPB	04 Nov 2022 15:12:43	U	SMPL	
N5418-01 - 1	TP-J	5704	1.1086	-7471B	PPB	04 Nov 2022 15:14:58	U	SMPL	
N5418-04 - 1	TP-K	1017	0.2230	-7471B	PPB	04 Nov 2022 15:17:13	U	SMPL	
N5418-07 - 1	TP-L	2056	0.4193	-7471B	PPB	04 Nov 2022 15:19:29	U	SMPL	
N5418-10 - 1	TP-M	20779	3.9569	-7471B	PPB	04 Nov 2022 15:21:44	U	SMPL	
N5418-13 - 1	TP-N	265	0.0809	-7471B	PPB	04 Nov 2022 15:24:01	U	SMPL	
N5418-16 - 1	TP-O	10100	1.9392	-7471B	PPB	04 Nov 2022 15:26:17	U	SMPL	
N5418-16DUP - 1	TP-ODUP	10139	1.9465	-7471B	PPB	04 Nov 2022 15:28:34	U	SMPL	
N5418-16MS - 1	TP-OMS	30500	5.7936	-7471B	PPB	04 Nov 2022 15:30:51	U	SMPL	
N5418-16MSD - 1	TP-OMSD	29490	5.6027	-7471B	PPB	04 Nov 2022 15:33:06	U	SMPL	
CCV49 - 1	CCV49	27032	5.1383	-7471B	PPB	04 Nov 2022 15:35:22	U	SMPL	
CCB49 - 1	CCB49	-46	0.0221	-7471B	PPB	04 Nov 2022 15:37:37	U	SMPL	
N5452-01 - 1	SOIL-PILE	8091	1.5596	-7471B	PPB	04 Nov 2022 15:39:53	U	SMPL	
PB148808BL - 1	PBS	-20	0.0271	-7471B	PPB	04 Nov 2022 15:42:08	U	SMPL	
PB148808BS - 1	LCSS	19612	3.7364	-7471B	PPB	04 Nov 2022 15:44:24	U	SMPL	
N5417-01 - 1	AUD-1731	283	0.0843	-7471B	PPB	04 Nov 2022 15:46:39	U	SMPL	
N5417-01DUP - 1	AUD-1731DUP	289	0.0854	-7471B	PPB	04 Nov 2022 15:48:55	U	SMPL	
N5417-01MS - 1	AUD-1731MS	21481	4.0895	-7471B	PPB	04 Nov 2022 15:51:11	U	SMPL	
N5417-01MSD - 1	AUD-1731MSD	21756	4.1415	-7471B	PPB	04 Nov 2022 15:53:28	U	SMPL	
N5417-02 - 1	AUD-1732	370	0.1007	-7471B	PPB	04 Nov 2022 15:55:45	U	SMPL	
N5425-01 - 1	371	903	0.2014	-7471B	PPB	04 Nov 2022 16:01:37	U	SMPL	
N5430-03 - 1	SOIL-COMP-1102	3344	0.6627	-7471B	PPB	04 Nov 2022 16:03:52	U	SMPL	
CCV50 - 1	CCV50	26817	5.0977	-7471B	PPB	04 Nov 2022 16:06:07	U	SMPL	
CCB50 - 1	CCB50	-97	0.0125	-7471B	PPB	04 Nov 2022 16:08:23	U	SMPL	
N5433-01 - 1	CL-01-110222	2637	0.5291	-7471B	PPB	04 Nov 2022 16:10:38	U	SMPL	
N5436-01 - 1	250FERRYBLVD-TFS-01	6580	1.2741	-7471B	PPB	04 Nov 2022 16:12:54	U	SMPL	
N5436-02 - 1	250FERRYBLVD-TFS-02	7203	1.3918	-7471B	PPB	04 Nov 2022 16:15:09	U	SMPL	
N5436-03 - 1	250FERRYBLVD-TFS-03	7008	1.3549	-7471B	PPB	04 Nov 2022 16:17:25	U	SMPL	
N5436-04 - 1	250FERRYBLVD-TFS-04	5563	1.0819	-7471B	PPB	04 Nov 2022 16:19:45	U	SMPL	
N5436-05 - 1	250FERRYBLVD-TFS-05	5696	1.1070	-7471B	PPB	04 Nov 2022 16:22:02	U	SMPL	
N5436-06 - 1	250FERRYBLVD-TFS-06	6981	1.3498	-7471B	PPB	04 Nov 2022 16:24:18	U	SMPL	
N5436-07 - 1	250FERRYBLVD-TFS-07	6147	1.1923	-7471B	PPB	04 Nov 2022 16:26:35	U	SMPL	
N5436-08 - 1	250FERRYBLVD-TFS-08	7044	1.3617	-7471B	PPB	04 Nov 2022 16:28:53	U	SMPL	
N5436-09 - 1	250FERRYBLVD-TFS-09	5962	1.1573	-7471B	PPB	04 Nov 2022 16:31:08	U	SMPL	
CCV51 - 1	CCV51	26773	5.0894	-7471B	PPB	04 Nov 2022 16:33:23	U	SMPL	
CCB51 - 1	CCB51	-66	0.0184	-7471B	PPB	04 Nov 2022 16:35:39	U	SMPL	
N5436-10 - 1	250FERRYBLVD-TFS-10	9141	1.7580	-7471B	PPB	04 Nov 2022 16:37:55	U	SMPL	
N5436-11 - 1	250FERRYBLVD-TFS-11	7006	1.3546	-7471B	PPB	04 Nov 2022 16:40:11	U	SMPL	
N5436-12 - 1	250FERRYBLVD-TFS-12	6369	1.2342	-7471B	PPB	04 Nov 2022 16:42:26	U	SMPL	
N5436-13 - 1	250FERRYBLVD-TFS-13	8482	1.6334	-7471B	PPB	04 Nov 2022 16:44:42	U	SMPL	
N5418-16LX5 - 1		1957	0.4006	-7471B	PPB	04 Nov 2022 16:46:57	U	SMPL	
N5418-16A - 1		34274	6.5066	-7471B	PPB	04 Nov 2022 16:49:14	U	SMPL	
N5417-01LX5 - 1		-79	0.0159	-7471B	PPB	04 Nov 2022 16:51:30	U	SMPL	
N5417-01A - 1		21891	4.1670	-7471B	PPB	04 Nov 2022 16:53:46	U	SMPL	
CCV52 - 1	CCV52	26959	5.1245	-7471B	PPB	04 Nov 2022 16:56:02	U	SMPL	
CCB52 - 1	CCB52	-69	0.0178	-7471B	PPB	04 Nov 2022 16:58:18	U	SMPL	
				-7471B	PPB	04 Nov 2022 17:00:44	U	SMPL	