



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Sohil

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122764

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	12	28.00	82.4	82.1	11/04/2022	11:45
2	N5454-01		1	14	100.00	>212.0	>212.0	11/04/2022	12:15
3	N5454-02		1	13	100.00	>212.0	>212.0	11/04/2022	13:00
4	N5454-02DUP		1	14	100.00	>212.0	>212.0	11/04/2022	13:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : flashpoint-110422 WorkList ID : 164714 Department : Wet-Chemistry Date : 11-04-2022 11:42:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/10/2022	Water	N5454-01	Flash Point	Cool 4 deg C	PSEG03	F11	TRE-1160	11/03/2022	1010B
11/10/2022	Water	N5454-02	Flash Point	Cool 4 deg C	PSEG03	F11	TRE-1674	11/03/2022	1010B

Date/Time 11/4/22 11:45
 Raw Sample Received by: SJ(WC)
 Raw Sample Relinquished by: CP(SM)

Date/Time 11/4/22 15:00
 Raw Sample Received by: CP(SM)
 Raw Sample Relinquished by: SJ(WC)