

Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: Sohil

Run Number: LB122765

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP98723
COD calibration std. 100 ppm	WP98721
COD calibration std. 10 ppm	WP98719
COD calibration std. 150 ppm	WP98722
COD calibration std. 50 ppm	WP98720
COD calibration std. 0 ppm	WP98718
COD Digestion Vials Low Level 0-150Mg/L	W2945
COD ICV-LCS std, 50ppm	WP99774
COD CCV std, 50ppm	WP99773
COD LOD std, 5ppm	WP99775

Temp In (C): 149	Date In: 11/04/2022	Time In: 10:20
Temp Out (C): 151	Date Out: 11/04/2022	Time Out: 12:20

Intercept: -0.5982

Slope: 1.0064

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	0.594		08/29/2022	13:40
2	CAL2	10	1	Water	9.000	9.537	4.6	08/29/2022	13:40
3	CAL3	50	1	Water	50.000	50.276	-0.6	08/29/2022	13:41
4	CAL4	100	1	Water	99.000	98.965	1	08/29/2022	13:41
5	CAL5	150	1	Water	151.000	150.634	-0.4	08/29/2022	13:42



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: Sohil

Run Number: LB122765

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	49.000	49.283	08/29/2022	13:42
2	ICB		NA	NA	1	Water	0.000	0.594	08/29/2022	13:43
3	CCV1	50	NA	NA	1	Water	50.000	50.276	11/04/2022	15:40
4	CCB1		NA	NA	1	Water	0.000	0.594	11/04/2022	15:40
5	LB122765BL		1	50	1	Solid	1.000	1.588	11/04/2022	15:41
6	LB122765BS	50	1	50	1	Solid	49.000	49.283	11/04/2022	15:41
7	N4955-03		1	50	1	Solid	5.000	5.563	11/04/2022	15:42
8	CCV2	50	NA	NA	1	Water	50.000	50.276	11/04/2022	15:42
9	CCB2		NA	NA	1	Water	1.000	1.588	11/04/2022	15:43