

LB 122769


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46361



A= 0.0000e+000

B= 2.1682e-004 *slope*C= -2.9818e-002 *intercept*

Rho= 0.9999091

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>o/d</i>
0.00	0.000	0.073	0.073	476	0.000	476					
0.05	0.050					185	0				
0.20	0.200	0.195	-0.005	1036	0.0 %	1036					
2.50	2.500	2.447	-0.053	11423	0.0 %	11423					
5.00	5.000	4.929	-0.071	22872	0.0 %	22872					
7.50	7.500	7.534	0.034	34884	0.0 %	34884					
10.0	10.000	10.022	0.022	46361	0.0 %	46361					

o/d
 1123
 0012

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	476	-		0 SFAM01.1	PPB	11/5/2022 16:07	Std
	0.2 S01	1036	-		0.2 SFAM01.1	PPB	11/5/2022 16:10	Std
	2.5 S02	11423	-		2.5 SFAM01.1	PPB	11/5/2022 16:12	Std
	5 S03	22872	-		5 SFAM01.1	PPB	11/5/2022 16:14	Std
	7.5 S04	34884	-		7.5 SFAM01.1	PPB	11/5/2022 16:16	Std
	10 S05	46361	-		10 SFAM01.1	PPB	11/5/2022 16:19	Std
ICV008	ICV008	18508	3.983 -		SFAM01.1	PPB	11/5/2022 16:24	SMPL
ICB008	ICB008	-35	-0.0374 -		SFAM01.1	PPB	11/5/2022 16:29	SMPL
CCV053	CCV053	22776	4.9084 -		SFAM01.1	PPB	11/5/2022 16:31	SMPL
CCB053	CCB053	4	-0.029 -		SFAM01.1	PPB	11/5/2022 16:33	SMPL
PB148841BL	PBS841	132	-0.0012 -		SFAM01.1	PPB	11/5/2022 16:38	SMPL
PB148842BL	PBS842	68	-0.0151 -		SFAM01.1	PPB	11/5/2022 16:43	SMPL
N5205-01	MH0AE5	500	0.0786 -		SFAM01.1	PPB	11/5/2022 16:45	SMPL
N5205-02	MH0AE7	7427	1.5805 -		SFAM01.1	PPB	11/5/2022 16:55	SMPL
N5205-03	MH0AE7D	7181	1.5271 -		SFAM01.1	PPB	11/5/2022 16:57	SMPL
N5205-04	MH0AE7S	18634	4.0104 -		SFAM01.1	PPB	11/5/2022 16:59	SMPL
N5205-05	MH0AE8	8745	1.8663 -		SFAM01.1	PPB	11/5/2022 17:01	SMPL
N5205-06	MH0AE9	1841	0.3693 -		SFAM01.1	PPB	11/5/2022 17:04	SMPL
N5205-07	MH0AF0	1719	0.3429 -		SFAM01.1	PPB	11/5/2022 17:06	SMPL
N5205-08	MH0AF1	16399	3.5258 -		SFAM01.1	PPB	11/5/2022 17:08	SMPL
N5205-09	MH0AF2	3317	0.6894 -		SFAM01.1	PPB	11/5/2022 17:10	SMPL
N5205-10	MH0AF3	16944	3.6439 -		SFAM01.1	PPB	11/5/2022 17:13	SMPL
CCV054	CCV054	22772	4.9075 -		SFAM01.1	PPB	11/5/2022 17:15	SMPL
CCB054	CCB054	75	-0.0136 -		SFAM01.1	PPB	11/5/2022 17:17	SMPL
N5205-11	MH0AF5	2716	0.5591 -		SFAM01.1	PPB	11/5/2022 17:19	SMPL
N5205-12	MH0AF6	3985	0.8342 -		SFAM01.1	PPB	11/5/2022 17:22	SMPL
N5205-13	MH0AF7	1340	0.2607 -		SFAM01.1	PPB	11/5/2022 17:24	SMPL
N5205-14	MH0AF8	1305	0.2531 -		SFAM01.1	PPB	11/5/2022 17:26	SMPL
N5205-15	MH0AF9	3572	0.7447 -		SFAM01.1	PPB	11/5/2022 17:29	SMPL
N5205-16	MH0AG0	9425	2.0137 -		SFAM01.1	PPB	11/5/2022 17:31	SMPL
N5205-17	MH0AG1	7311	1.5553 -		SFAM01.1	PPB	11/5/2022 17:33	SMPL
N5205-18	MH0AG2	18229	3.9225 -		SFAM01.1	PPB	11/5/2022 17:35	SMPL
N5205-19	MH0AG3	2549	0.5228 -		SFAM01.1	PPB	11/5/2022 17:38	SMPL
N5205-20	MH0AG4	5290	1.1171 -		SFAM01.1	PPB	11/5/2022 17:40	SMPL
N5205-21	MH0AG5	1066064	HIGH -		SFAM01.1	PPB	11/5/2022 17:42	
CCV055	CCV055	21824	4.702 -		SFAM01.1	PPB	11/5/2022 17:44	SMPL
CCB055	CCB055	27	-0.024 -		SFAM01.1	PPB	11/5/2022 17:51	SMPL
N5206-01	MH0AF4	3697	0.7718 -		SFAM01.1	PPB	11/5/2022 18:01	SMPL
N5206-02	MH0AF4D	3780	0.7898 -		SFAM01.1	PPB	11/5/2022 18:03	SMPL
N5206-03	MH0AF4S	15641	3.3614 -		SFAM01.1	PPB	11/5/2022 18:05	SMPL
N5206-04	MH0AG6	5950	1.2602 -		SFAM01.1	PPB	11/5/2022 18:08	SMPL
N5206-05	MH0AG7	37	-0.0218 -		SFAM01.1	PPB	11/5/2022 18:10	SMPL
N5206-06	MH0AG8	8377	1.7865 -		SFAM01.1	PPB	11/5/2022 18:12	SMPL
N5206-07	MH0AG9	12399	2.6585 -		SFAM01.1	PPB	11/5/2022 18:15	SMPL
N5206-08	MH0AH0	8671	1.8502 -		SFAM01.1	PPB	11/5/2022 18:17	SMPL
N5206-09	MH0AH1	1220	0.2347 -		SFAM01.1	PPB	11/5/2022 18:19	SMPL

LB122769

INSTRUMENT ID : CV1

N5206-10	MH0AH2	12172	2.6093 -	SFAM01.1	PPB	11/5/2022 18:21	SMPL
CCV056	CCV056	22809	4.9156 -	SFAM01.1	PPB	11/5/2022 18:24	SMPL
CCB056	CCB056	31	-0.0231 -	SFAM01.1	PPB	11/5/2022 18:36	SMPL
N5206-11	MH0AH3	17039	3.6645 -	SFAM01.1	PPB	11/5/2022 18:39	SMPL
N5206-12	MH0AH4	436	0.0647 -	SFAM01.1	PPB	11/5/2022 18:41	SMPL
N5206-13	MH0AH5	12120	2.598 -	SFAM01.1	PPB	11/5/2022 18:43	SMPL
N5206-14	MH0AH6	998	0.1866 -	SFAM01.1	PPB	11/5/2022 18:45	SMPL
N5206-15	MH0AH7	550	0.0894 -	SFAM01.1	PPB	11/5/2022 18:48	SMPL
PB148843BL	PBS843	-5	-0.0309 -	SFAM01.1	PPB	11/5/2022 18:50	SMPL
N4519-04	MDL-SOIL	351	0.0463 -	SFAM01.1	PPB	11/5/2022 18:52	SMPL
N5205-21DLX10	MH0AG5	214118	46.3947 -	SFAM01.1	PPB	11/5/2022 18:54	SMPL
N5205-21DLX100	MH0AG5	21016	4.5268 -	SFAM01.1	PPB	11/5/2022 19:05	SMPL
CCV057	CCV057	22669	4.8852 -	SFAM01.1	PPB	11/5/2022 19:07	SMPL
CCB057	CCB057	98	-0.0086 -	SFAM01.1	PPB	11/5/2022 19:12	SMPL