



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB122773

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	11/06/2022	12:30
2	CAL2	1	Water	NA	NA	20.7	7.00	11/06/2022	12:31
3	CAL3	1	Water	NA	NA	20.8	10.03	11/06/2022	12:33
4	ICV	1	Water	NA	NA	20.8	6.99	11/06/2022	12:35
5	CCV1	1	Water	NA	NA	20.8	2.01	11/06/2022	12:39
6	N5479-04	1	Solid	20.02	20	23.2	9.12	11/06/2022	12:40
7	N5491-02	1	Solid	20.03	20	22.6	7.90	11/06/2022	12:44
8	N5491-05	1	Solid	20.04	20	22.7	7.55	11/06/2022	12:46
9	N5491-08	1	Solid	20.02	20	23.1	7.80	11/06/2022	12:50
10	N5491-11	1	Solid	20.03	20	23.2	7.69	11/06/2022	13:00
11	N5491-11DUP	1	Solid	20.04	20	23.3	7.70	11/06/2022	13:02
12	CCV2	1	Water	NA	NA	20.8	12.03	11/06/2022	13:05

WJ 122773

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N5491 WorkList ID : 164755 Department: Wet-Chemistry Date : 11-06-2022 07:38:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/04/2022	Solid	N5479-04	Corrosivity	Cool 4 deg C	TULL02	F11	TUL-ROS-1104	11/04/2022	9045D
11/04/2022	Solid	N5491-02	Corrosivity	Cool 4 deg C	PSEG03	F11	TP-1	11/04/2022	9045D
11/04/2022	Solid	N5491-05	Corrosivity	Cool 4 deg C	PSEG03	F11	TP-2	11/04/2022	9045D
11/04/2022	Solid	N5491-08	Corrosivity	Cool 4 deg C	PSEG03	F11	TP-3	11/04/2022	9045D
11/04/2022	Solid	N5491-11	Corrosivity	Cool 4 deg C	PSEG03	F11	TP-4	11/04/2022	9045D

Date/Time 11/06/22 10:00
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: CP SR

Date/Time 11/06/22 15:00
 Raw Sample Received by: CP SR
 Raw Sample Relinquished by: JDC