

LB122779


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

56180



A= 0.0000e+000

B= 1.7863e-004

C= 9.1809e-002

Rho= 0.9993821

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.0	0.000	0.097	0.097	31	0.000	31					
0.2	0.200	0.195	-0.005	578	0.0 %	578					
2.5	2.500	2.388	-0.112	12852	0.0 %	12852					
5.0	5.000	5.115	0.115	28120	0.0 %	28120					
7.5	7.500	7.278	-0.222	40228	0.0 %	40228					
10.0	10.000	10.127	0.127	56180	0.0 %	56180					

0/01
 1
 2
 3
 4
 5

LB112779 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 07 Nov 2022 12:55:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	31	-	0.0000	7471B	PPB	07 Nov 2022 12:58:23	S	Std
0.2 - 1	50.2	578	-	0.2000	7471B	PPB	07 Nov 2022 13:00:40	S	Std
2.5 - 1	52.5	12852	-	2.5000	7471B	PPB	07 Nov 2022 13:02:58	S	Std
5.0 - 1	55	28120	-	5.0000	7471B	PPB	07 Nov 2022 13:05:15	S	Std
7.5 - 1	57.5	40228	-	7.5000	7471B	PPB	07 Nov 2022 13:10:05	S	Std
10.0 - 1	510	56180	-	10.0000	7471B	PPB	07 Nov 2022 13:14:59	S	Std
ICV10 - 1	ICV10	21329	3.9018	-	7471B	PPB	07 Nov 2022 13:17:51	U	SMPL
ICB10 - 1	ICB10	-30	0.0865	-	7471B	PPB	07 Nov 2022 13:20:06	U	SMPL
CCV66 - 1	CCV66	27822	5.0617	-	7471B	PPB	07 Nov 2022 13:22:21	U	SMPL
CCB66 - 1	CCB66	-163	0.0627	-	7471B	PPB	07 Nov 2022 13:24:36	U	SMPL
CRA - 1	CRA	534	0.1872	-	7471B	PPB	07 Nov 2022 13:26:51	U	SMPL
HighStd - 1	HighStd	56811	10.2400	-	7471B	PPB	07 Nov 2022 13:29:07	U	SMPL
ChkStd - 1	ChkStd	38870	7.0352	-	7471B	PPB	07 Nov 2022 13:31:22	U	SMPL
PB148859BL - 1	PBS	-196	0.0568	-	7471B	PPB	07 Nov 2022 13:33:39	U	SMPL
PB148859BS - 1	LCSS	19752	3.6201	-	7471B	PPB	07 Nov 2022 13:35:55	U	SMPL
N5435-01 - 1	PTS-TP-COMP-1103	17720	3.2572	-	7471B	PPB	07 Nov 2022 13:38:11	U	SMPL
N5435-03 - 1	PTS-SOIL-PILE-1103	45302	8.1842	-	7471B	PPB	07 Nov 2022 13:40:28	U	SMPL
N5453-01 - 1	TR-6-110322	6711	1.2906	-	7471B	PPB	07 Nov 2022 13:42:46	U	SMPL
N5459-01 - 1	EO-03-110322	3662	0.7460	-	7471B	PPB	07 Nov 2022 13:45:03	U	SMPL
N5459-03 - 1	EO-02-110322	34148	6.1917	-	7471B	PPB	07 Nov 2022 13:47:19	U	SMPL
CCV67 - 1	CCV67	27794	5.0567	-	7471B	PPB	07 Nov 2022 13:49:35	U	SMPL
CCB67 - 1	CCB67	-150	0.0650	-	7471B	PPB	07 Nov 2022 13:51:50	U	SMPL
N5460-01 - 1	SH-1	807	0.2360	-	7471B	PPB	07 Nov 2022 13:54:06	U	SMPL
N5478-01 - 1	MH-NYE-AVE	28358	5.1574	-	7471B	PPB	07 Nov 2022 13:56:21	U	SMPL
N5479-03 - 1	TUL-ROS-1104	10432	1.9553	-	7471B	PPB	07 Nov 2022 13:58:37	U	SMPL
N5481-01 - 1	VNJ-231	3103	0.6461	-	7471B	PPB	07 Nov 2022 14:00:53	U	SMPL
N5485-01 - 1	OR-03-110422	2554	0.5480	-	7471B	PPB	07 Nov 2022 14:03:08	U	SMPL
N5491-01 - 1	TP-1	386	0.1608	-	7471B	PPB	07 Nov 2022 14:05:25	U	SMPL
N5491-01DUP - 1	TP-1DUP	504	0.1818	-	7471B	PPB	07 Nov 2022 14:07:41	U	SMPL
N5491-01MS - 1	TP-1MS	20313	3.7203	-	7471B	PPB	07 Nov 2022 14:09:58	U	SMPL
N5491-01MSD - 1	TP-1MSD	20751	3.7986	-	7471B	PPB	07 Nov 2022 14:12:15	U	SMPL
N5491-04 - 1	TP-2	975	0.2660	-	7471B	PPB	07 Nov 2022 14:14:30	U	SMPL
CCV68 - 1	CCV68	28334	5.1531	-	7471B	PPB	07 Nov 2022 14:16:46	U	SMPL
CCB68 - 1	CCB68	-70	0.0793	-	7471B	PPB	07 Nov 2022 14:19:01	U	SMPL
N5491-07 - 1	TP-3	816	0.2376	-	7471B	PPB	07 Nov 2022 14:21:17	U	SMPL
N5491-10 - 1	TP-4	3307	0.6825	-	7471B	PPB	07 Nov 2022 14:23:33	U	SMPL
N5491-01LX5 - 1		-27	0.0870	-	7471B	PPB	07 Nov 2022 14:25:48	U	SMPL
N5491-01A - 1		21149	3.8697	-	7471B	PPB	07 Nov 2022 14:28:05	U	SMPL
PB148644BL - 1	PBS	-156	0.0639	-	7471B	PPB	07 Nov 2022 14:30:21	U	SMPL
PB148644BS - 1	LCSS	21243	3.8865	-	7471B	PPB	07 Nov 2022 14:32:37	U	SMPL
CCV69 - 1	CCV69	28271	5.1419	-	7471B	PPB	07 Nov 2022 14:34:53	U	SMPL
CCB69 - 1	CCB69	-163	0.0627	-	7471B	PPB	07 Nov 2022 14:37:10	U	SMPL