



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122781

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	12	28.00	82.4	82.4	11/07/2022	13:30
2	N5482-02		1	15	100.00	>212.0	>212.0	11/07/2022	14:00
3	N5482-02DUP		1	14	100.00	>212.0	>212.0	11/07/2022	14:30
4	N5487-02		1	12	100.00	>212.0	>212.0	11/07/2022	15:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

6122781

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp11-7 WorkList ID : 164783 Department : Wet-Chemistry Date : 11-07-2022 13:22:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/11/2022	Water	N5482-02	Flash Point	Cool 4 deg C	PSEG03	F11	FLOOR-DRAIN-WATER	11/04/2022	1010B
11/11/2022	Water	N5487-02	Flash Point	Cool 4 deg C	PSEG03	F11	FMC-0666	11/04/2022	1010B

Date/Time 11/7/22 13:25
Raw Sample Received by: SJ(WC)
Raw Sample Relinquished by: CP(SN)

Date/Time 11/7/22 15:30
Raw Sample Received by: CP(SN)
Raw Sample Relinquished by: SJ(WC)