

LB122786


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42882



A= 0.0000e+000

B= 2.3456e-004

C= -9.3134e-003

Rho= 0.9998088

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Q/D
0.0	0.000	-0.025	-0.025	-67	0.000	-67					
0.2	0.200	0.177	-0.023	793	0.0 %	793					-12
2.5	2.500	2.569	0.069	10993	0.0 %	10993					3
5.0	5.000	5.066	0.066	21639	0.0 %	21639					1
7.5	7.500	7.364	-0.136	31435	0.0 %	31435					-2
10.0	10.000	10.049	0.049	42882	0.0 %	42882					0

LB122786 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Nov 2022 15:23:59

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-67	-	0.0000	7470A	PPB	07 Nov 2022 15:28:28	S	Std
0.2 - 1	50.2	793	-	0.2000	7470A	PPB	07 Nov 2022 15:30:45	S	Std
2.5 - 1	52.5	10993	-	2.5000	7470A	PPB	07 Nov 2022 15:33:02	S	Std
5.0 - 1	55	21639	-	5.0000	7470A	PPB	07 Nov 2022 15:35:19	S	Std
7.5 - 1	57.5	31435	-	7.5000	7470A	PPB	07 Nov 2022 15:37:35	S	Std
10.0 - 1	59	42882	-	10.0000	7470A	PPB	07 Nov 2022 15:39:51	S	Std
ICV11 - 1	ICV11	16276	3.8083	-	7470A	PPB	07 Nov 2022 15:42:34	U	SMPL
ICB11 - 1	ICB11	15	-0.0058	-	7470A	PPB	07 Nov 2022 15:44:48	U	SMPL
CCV70 - 1	CCV70	21964	5.1425	-	7470A	PPB	07 Nov 2022 15:47:04	U	SMPL
CCB70 - 1	CCB70	-109	-0.0349	-	7470A	PPB	07 Nov 2022 15:49:18	U	SMPL
CRA - 1	CRA	803	0.1790	-	7470A	PPB	07 Nov 2022 15:51:34	U	SMPL
HighStd - 1	HighStd	43333	10.1547	-	7470A	PPB	07 Nov 2022 15:53:48	U	SMPL
ChkStd - 1	ChkStd	29522	6.9153	-	7470A	PPB	07 Nov 2022 15:56:04	U	SMPL
PB148862BL - 1	PBW	-102	-0.0332	-	7470A	PPB	07 Nov 2022 15:58:20	U	SMPL
PB148862BS - 1	LCSW	16824	3.9369	-	7470A	PPB	07 Nov 2022 16:00:36	U	SMPL
N5479-04 - 1	TUL-ROS-1104	-68	-0.0253	-	7470A	PPB	07 Nov 2022 16:02:52	U	SMPL
N5491-02 - 1	TP-1	-62	-0.0238	-	7470A	PPB	07 Nov 2022 16:05:08	U	SMPL
N5491-05 - 1	TP-2	-82	-0.0285	-	7470A	PPB	07 Nov 2022 16:07:25	U	SMPL
N5491-08 - 1	TP-3	-88	-0.0300	-	7470A	PPB	07 Nov 2022 16:09:42	U	SMPL
N5491-11 - 1	TP-4	-93	-0.0311	-	7470A	PPB	07 Nov 2022 16:11:57	U	SMPL
CCV71 - 1	CCV71	21791	5.1019	-	7470A	PPB	07 Nov 2022 16:14:13	U	SMPL
CCB71 - 1	CCB71	-126	-0.0389	-	7470A	PPB	07 Nov 2022 16:16:28	U	SMPL
N5491-11DUP - 1	TP-4DUP	-18	-0.0135	-	7470A	PPB	07 Nov 2022 16:18:44	U	SMPL
N5491-11MS - 1	TP-4MS	16169	3.7832	-	7470A	PPB	07 Nov 2022 16:21:03	U	SMPL
N5491-11MSD - 1	TP-4MSD	16055	3.7565	-	7470A	PPB	07 Nov 2022 16:23:20	U	SMPL
PB148874BL - 1	PBW	-22	-0.0145	-	7470A	PPB	07 Nov 2022 16:25:35	U	SMPL
PB148874BS - 1	LCSW	15439	3.6120	-	7470A	PPB	07 Nov 2022 16:27:51	U	SMPL
N5272-11 - 1	SPLP-LEACHATE-AMENDED-COMP-01	-21	-0.0142	-	7470A	PPB	07 Nov 2022 16:30:08	U	SMPL
N5272-11DUP - 1	SPLP-LEACHATE-AMENDED-COMP-01DUP	14	-0.0060	-	7470A	PPB	07 Nov 2022 16:32:24	U	SMPL
N5272-11MS - 1	SPLP-LEACHATE-AMENDED-COMP-01MS	14899	3.4853	-	7470A	PPB	07 Nov 2022 16:34:41	U	SMPL
N5272-11MSD - 1	SPLP-LEACHATE-AMENDED-COMP-01MSD	14584	3.4115	-	7470A	PPB	07 Nov 2022 16:36:58	U	SMPL
N5272-12 - 1	SPLP-LEACHATE-AMENDED-COMP-02	-77	-0.0274	-	7470A	PPB	07 Nov 2022 16:39:14	U	SMPL
CCV72 - 1	CCV72	21898	5.1270	-	7470A	PPB	07 Nov 2022 16:41:29	U	SMPL
CCB72 - 1	CCB72	-94	-0.0314	-	7470A	PPB	07 Nov 2022 16:43:45	U	SMPL
N5272-13 - 1	SPLP-LEACHATE-AMENDED-COMP-03	66	0.0062	-	7470A	PPB	07 Nov 2022 16:46:01	U	SMPL
N5272-14 - 1	SPLP-LEACHATE-AMENDED-COMP-08	-20	-0.0140	-	7470A	PPB	07 Nov 2022 16:48:16	U	SMPL
N5272-15 - 1	SPLP-LEACHATE-AMENDED-COMP-09	-43	-0.0194	-	7470A	PPB	07 Nov 2022 16:50:32	U	SMPL
PB148847TB - 1	PB148847TB	-60	-0.0234	-	7470A	PPB	07 Nov 2022 16:52:49	U	SMPL
PB148862TB - 1	PB148862TB	-25	-0.0152	-	7470A	PPB	07 Nov 2022 16:55:08	U	SMPL
N5491-11LX5 - 1		-108	-0.0346	-	7470A	PPB	07 Nov 2022 16:57:26	U	SMPL
N5491-11A - 1		16361	3.8283	-	7470A	PPB	07 Nov 2022 16:59:43	U	SMPL
N5272-11LX5 - 1		-110	-0.0351	-	7470A	PPB	07 Nov 2022 17:01:59	U	SMPL
N5272-11A - 1		16290	3.8116	-	7470A	PPB	07 Nov 2022 17:04:17	U	SMPL
CCV73 - 1	CCV73	22002	5.1514	-	7470A	PPB	07 Nov 2022 17:06:34	U	SMPL
CCB73 - 1	CCB73	-70	-0.0257	-	7470A	PPB	07 Nov 2022 17:08:49	U	SMPL

PB148847TB

11/8/22