



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB122791

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	11/08/2022	08:40
2	CAL2	1	Water	NA	NA	20.9	7.00	11/08/2022	08:42
3	CAL3	1	Water	NA	NA	20.8	10.03	11/08/2022	08:45
4	ICV	1	Water	NA	NA	20.8	6.98	11/08/2022	08:46
5	CCV1	1	Water	NA	NA	20.7	2.01	11/08/2022	08:50
6	N5504-02	1	Water	NA	NA	20.1	12.59	11/08/2022	10:00
7	N5504-02DUP	1	Water	NA	NA	20.2	12.60	11/08/2022	10:01
8	CCV2	1	Water	NA	NA	20.7	12.03	11/08/2022	10:02

VB 122791

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N5002 w WorkList ID : 164801 Department : Wet-Chemistry Date : 11-08-2022 08:26:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/07/2022	Water	N5504-02	Corrosivity	Cool 4 deg C	TULL02	F11	SV34211L	11/07/2022	9040C

Date/Time 11/08/12 08:40
Raw Sample Received by: JDC
Raw Sample Relinquished by: JDC

Date/Time 11/08/12 10:50
Raw Sample Received by: JDC
Raw Sample Relinquished by: JDC