

LB122792


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4255E



A= 0.0000e+000

B= 2.3480e-004

C= -2.5747e-003

Rho= 0.9999583

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.001	0.001	15	0.000	15					
0.2	0.200	0.183	-0.017	792	0.0 %	792					-9
2.5	2.500	2.483	-0.017	10586	0.0 %	10586					-1
5.0	5.000	5.072	0.072	21614	0.0 %	21614					1
7.5	7.500	7.470	-0.030	31825	0.0 %	31825					0
10.0	10.000	9.990	-0.010	42558	0.0 %	42558					0

LB122792 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Nov 2022 09:11:05

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	15	-	0.00007470A	PPB		08 Nov 2022 09:26:54	S	Std
0.2 - 1	50.2	792	-	0.20007470A	PPB		08 Nov 2022 09:29:11	S	Std
2.5 - 1	52.5	10586	-	2.50007470A	PPB		08 Nov 2022 09:31:28	S	Std
5.0 - 1	55	21614	-	5.00007470A	PPB		08 Nov 2022 09:33:45	S	Std
7.5 - 1	57.5	31825	-	7.50007470A	PPB		08 Nov 2022 09:36:01	S	Std
10.0 - 1	58	42558	-	10.00007470A	PPB		08 Nov 2022 09:38:18	S	Std
ICV12 - 1	ICV12	17311	4.0621	-7470A	PPB		08 Nov 2022 09:41:04	U	SMPL
ICB12 - 1	ICB12	52	0.0096	-7470A	PPB		08 Nov 2022 09:43:19	U	SMPL
CCV74 - 1	CCV74	21611	5.0717	-7470A	PPB		08 Nov 2022 09:45:34	U	SMPL
CCB74 - 1	CCB74	-79	-0.0211	-7470A	PPB		08 Nov 2022 09:47:49	U	SMPL
CRA - 1	CRA	747	0.1728	-7470A	PPB		08 Nov 2022 09:50:04	U	SMPL
HighStd - 1	HighStd	42590	9.9977	-7470A	PPB		08 Nov 2022 09:52:19	U	SMPL
ChkStd - 1	ChkStd	30077	7.0596	-7470A	PPB		08 Nov 2022 09:54:34	U	SMPL
PB148553BL - 1	PBW	-80	-0.0214	-7470A	PPB		08 Nov 2022 09:56:50	U	SMPL
PB148553BS - 1	LCSW	15632	3.6679	-7470A	PPB		08 Nov 2022 09:59:06	U	SMPL
N5226-01 - 1	20221093-198-SITE-WATER	-16	-0.0063	-7470A	PPB		08 Nov 2022 10:01:22	U	SMPL
N5226-02 - 1	20221094-199-SITE-WATER	-30	-0.0096	-7470A	PPB		08 Nov 2022 10:03:38	U	SMPL
N5226-03 - 1	20221096-201-SITE-WATER	48	0.0087	-7470A	PPB		08 Nov 2022 10:05:55	U	SMPL
N5226-04 - 1	20221097-202-SITE-WATER	-43	-0.0127	-7470A	PPB		08 Nov 2022 10:08:12	U	SMPL
N5226-05 - 1	20221098-203-SITE-WATER	259	0.0582	-7470A	PPB		08 Nov 2022 10:10:27	U	SMPL
CCV75 - 1	CCV75	21439	5.0314	-7470A	PPB		08 Nov 2022 10:12:42	U	SMPL
CCB75 - 1	CCB75	-103	-0.0268	-7470A	PPB		08 Nov 2022 10:14:57	U	SMPL
N5226-06 - 1	20221101-PC-SITE-WATER	19	0.0019	-7470A	PPB		08 Nov 2022 10:17:13	U	SMPL
N5226-07 - 1	20221091-FIELD-BLANK-1	-17	-0.0066	-7470A	PPB		08 Nov 2022 10:19:28	U	SMPL
N5235-07 - 1	LSP-GW-03-DUP	154	0.0336	-7470A	PPB		08 Nov 2022 10:21:44	U	SMPL
N5236-02 - 1	LSP-GW-02	424	0.0970	-7470A	PPB		08 Nov 2022 10:24:00	U	SMPL
N5236-02DUP - 1	LSP-GW-02DUP	364	0.0829	-7470A	PPB		08 Nov 2022 10:26:15	U	SMPL
N5236-02MS - 1	LSP-GW-02MS	17020	3.9938	-7470A	PPB		08 Nov 2022 10:31:02	U	SMPL
N5236-02MSD - 1	LSP-GW-02MSD	17243	4.0461	-7470A	PPB		08 Nov 2022 10:33:18	U	SMPL
N5236-03 - 1	LSP-GW-03	165	0.0362	-7470A	PPB		08 Nov 2022 10:35:35	U	SMPL
N5237-01 - 1	LSP-GW-01	76	0.0153	-7470A	PPB		08 Nov 2022 10:37:52	U	SMPL
N5237-04 - 1	LSP-RB	92	0.0190	-7470A	PPB		08 Nov 2022 10:40:07	U	SMPL
CCV76 - 1	CCV76	21369	5.0149	-7470A	PPB		08 Nov 2022 10:42:23	U	SMPL
CCB76 - 1	CCB76	-100	-0.0261	-7470A	PPB		08 Nov 2022 10:44:39	U	SMPL
N5260-01 - 1	MW-3	529	0.1216	-7470A	PPB		08 Nov 2022 10:46:54	U	SMPL
N5260-02 - 1	MW-3	230	0.0514	-7470A	PPB		08 Nov 2022 10:49:09	U	SMPL
N5266-01 - 1	20221095-200-SITE-WATER	-13	-0.0056	-7470A	PPB		08 Nov 2022 10:51:25	U	SMPL
N5266-02 - 1	20221099-204-SITE-WATER	-15	-0.0061	-7470A	PPB		08 Nov 2022 10:53:41	U	SMPL
N5266-03 - 1	20221100-204.1-SITE-WATER	-17	-0.0066	-7470A	PPB		08 Nov 2022 10:55:56	U	SMPL
N5266-04 - 1	20221092-FIELD-BLANK-2	-38	-0.0115	-7470A	PPB		08 Nov 2022 10:58:12	U	SMPL
N5236-02LX5 - 1		-30	-0.0096	-7470A	PPB		08 Nov 2022 11:00:29	U	SMPL
N5236-02A - 1		17166	4.0280	-7470A	PPB		08 Nov 2022 11:02:46	U	SMPL
CCV77 - 1	CCV77	21298	4.9983	-7470A	PPB		08 Nov 2022 11:05:02	U	SMPL
CCB77 - 1	CCB77	-58	-0.0162	-7470A	PPB		08 Nov 2022 11:07:18	U	SMPL