



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB122793

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	11/08/2022	08:35
2	CAL2	1	Water	NA	NA	20.3	7.00	11/08/2022	08:36
3	CAL3	1	Water	NA	NA	20.8	10.03	11/08/2022	08:37
4	ICV	1	Water	NA	NA	20.7	6.98	11/08/2022	08:40
5	CCV1	1	Water	NA	NA	20.7	2.01	11/08/2022	08:42
6	N5468-01	1	Water	NA	NA	20.1	4.75	11/08/2022	08:45
7	N5474-01	1	Water	NA	NA	20.4	7.66	11/08/2022	08:46
8	N5474-02	1	Water	NA	NA	20.2	7.71	11/08/2022	08:47
9	N5476-01	1	Water	NA	NA	20.1	7.68	11/08/2022	08:50
10	N5476-02	1	Water	NA	NA	20.1	7.71	11/08/2022	08:53
11	N5499-01	1	Water	NA	NA	22.1	6.74	11/08/2022	09:00
12	N5500-01	1	Water	NA	NA	20.2	6.89	11/08/2022	09:02
13	N5512-07	1	Water	NA	NA	22.7	6.89	11/08/2022	09:05
14	N5512-09	1	Water	NA	NA	22.3	8.17	11/08/2022	09:07
15	N5512-09DUP	1	Water	NA	NA	22.4	8.18	11/08/2022	09:09
16	CCV2	1	Water	NA	NA	20.8	12.03	11/08/2022	09:10

WORKLIST(Hardcopy Internal Chain)

183 122793

WorkList Name : ph n5500 w

WorkList ID : 164803

Department : Wet-Chemistry

Date : 11-08-2022 08:28:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/02/2022	Water	N5468-01	pH	Cool 4 deg C	STER04	N11	POOL-WATER	11/02/2022	9040C
11/03/2022	Water	N5474-01	pH	Cool 4 deg C	ENV170	F11	5592-3	11/03/2022	9040C
11/03/2022	Water	N5474-02	pH	Cool 4 deg C	ENV170	F11	5592-4	11/03/2022	9040C
11/03/2022	Water	N5476-01	pH	Cool 4 deg C	ENV170	F11	RED-006-5	11/03/2022	9040C
11/03/2022	Water	N5476-02	pH	Cool 4 deg C	ENV170	F11	RED-006-6	11/03/2022	9040C
11/07/2022	Water	N5499-01	pH	Cool 4 deg C	PSEG03	F11	SCP122869-TOTE	11/07/2022	9040C
11/07/2022	Water	N5500-01	pH	Cool 4 deg C	PSEG03	F11	NWB-1851	11/07/2022	9040C
11/07/2022	Water	N5512-07	pH	Cool 4 deg C	PSEG03	G11	AUD-1911-TOTE	11/07/2022	9040C
11/07/2022	Water	N5512-09	pH	Cool 4 deg C	PSEG03	G11	AUD-1896-TOTE	11/07/2022	9040C

11/08/22 08:30

Date/Time
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

11/08/22

Date/Time
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]