



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB122794

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	11/08/2022	08:35
2	CAL2	1	Water	NA	NA	20.3	7.00	11/08/2022	08:36
3	CAL3	1	Water	NA	NA	20.8	10.03	11/08/2022	08:37
4	ICV	1	Water	NA	NA	20.7	6.98	11/08/2022	08:40
5	CCV1	1	Water	NA	NA	20.7	2.01	11/08/2022	08:42
6	N5508-01	1	Solid	20.02	20	20.3	8.17	11/08/2022	09:00
7	N5508-01DUP	1	Solid	20.03	20	20.4	8.18	11/08/2022	09:01
8	CCV2	1	Water	NA	NA	20.7	12.03	11/08/2022	09:05

122794
13 122794

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n5508 s WorkList ID : 164802 Department : Wet-Chemistry Date : 11-08-2022 08:27:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/07/2022	Solid	N5508-01	pH		Cool 4 deg C	PSEG05	G11	AU-06-110722	11/07/2022 9045D

Date/Time 11/08/22 08:30
Raw Sample Received by: 70 (GOC)
Raw Sample Relinquished by: [Signature]

Date/Time 11/08/22 11:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: 70 (GOC)