

LB 122799


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51600



A= 0.0000e+000

B= 1.9128e-004

C= -3.0032e-002

Rho= 0.9995880

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/d
0.0	0.000	0.007	0.007	192	0.000	192					
0.2	0.200	0.165	-0.035	1021	0.0 %	1021					-18
2.5	2.500	2.444	-0.056	12933	0.0 %	12933					-2
5.0	5.000	5.061	0.061	26614	0.0 %	26614					1
7.5	7.500	7.683	0.183	40325	0.0 %	40325					2
10.0	10.000	9.840	-0.160	51600	0.0 %	51600					-2

LB122799 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 08 Nov 2022 15:35:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	192	-	0.000	7471B	PPB	08 Nov 2022 15:37:47	S	Std
0.2 - 1	50-2	1021	-	0.200	7471B	PPB	08 Nov 2022 15:40:04	S	Std
2.5 - 1	52-5	12933	-	2.500	7471B	PPB	08 Nov 2022 15:42:21	S	Std
5.0 - 1	55	26614	-	5.000	7471B	PPB	08 Nov 2022 15:44:38	S	Std
7.5 - 1	57-5	40325	-	7.500	7471B	PPB	08 Nov 2022 15:46:54	S	Std
10.0 - 1	310	51600	-	10.000	7471B	PPB	08 Nov 2022 15:49:10	S	Std
ICV13 - 1	ICV13	21592	4.1001	-	7471B	PPB	08 Nov 2022 15:52:01	U	SMPL
ICB13 - 1	ICB13	107	-0.0096	-	7471B	PPB	08 Nov 2022 15:54:16	U	SMPL
CCV78 - 1	CCV78	27010	5.1365	-	7471B	PPB	08 Nov 2022 15:56:31	U	SMPL
CCB78 - 1	CCB78	17	-0.0268	-	7471B	PPB	08 Nov 2022 15:58:46	U	SMPL
CRA - 1	CRA	1365	0.2311	-	7471B	PPB	08 Nov 2022 16:03:31	U	SMPL
HighStd - 1	HighStd	51729	9.8648	-	7471B	PPB	08 Nov 2022 16:05:45	U	SMPL
ChkStd - 1	ChkStd	37882	7.2161	-	7471B	PPB	08 Nov 2022 16:08:00	U	SMPL
PB148904BL - 1	PBS	-107	-0.0505	-	7471B	PPB	08 Nov 2022 16:10:16	U	SMPL
PB148904BS - 1	LCSS	20102	3.8151	-	7471B	PPB	08 Nov 2022 16:12:33	U	SMPL
N5451-01 - 1	LFR-SOIL-1103	3895	0.7150	-	7471B	PPB	08 Nov 2022 16:14:49	U	SMPL
N5494-01 - 1	SB-19	1107	0.1817	-	7471B	PPB	08 Nov 2022 16:17:05	U	SMPL
N5494-01DUP - 1	SB-19DUP	1028	0.1666	-	7471B	PPB	08 Nov 2022 16:19:22	U	SMPL
N5494-01MS - 1	SB-19MS	21916	4.1621	-	7471B	PPB	08 Nov 2022 16:21:39	U	SMPL
N5494-01MSD - 1	SB-19MSD	22413	4.2572	-	7471B	PPB	08 Nov 2022 16:23:54	U	SMPL
CCV79 - 1	CCV79	27034	5.1411	-	7471B	PPB	08 Nov 2022 16:26:09	U	SMPL
CCB79 - 1	CCB79	108	-0.0094	-	7471B	PPB	08 Nov 2022 16:28:24	U	SMPL
N5494-03 - 1	SB-18	860	0.1345	-	7471B	PPB	08 Nov 2022 16:30:40	U	SMPL
N5494-05 - 1	SB-17	737	0.1109	-	7471B	PPB	08 Nov 2022 16:32:55	U	SMPL
N5494-07 - 1	SB-16	1536	0.2638	-	7471B	PPB	08 Nov 2022 16:35:10	U	SMPL
N5494-09 - 1	SB-15	3058	0.5549	-	7471B	PPB	08 Nov 2022 16:37:26	U	SMPL
N5494-11 - 1	SB-24	1159	0.1917	-	7471B	PPB	08 Nov 2022 16:39:42	U	SMPL
N5494-13 - 1	SB-23	807	0.1243	-	7471B	PPB	08 Nov 2022 16:41:58	U	SMPL
N5494-15 - 1	SB-22	695	0.1029	-	7471B	PPB	08 Nov 2022 16:44:14	U	SMPL
N5494-17 - 1	SB-20	866	0.1356	-	7471B	PPB	08 Nov 2022 16:46:31	U	SMPL
N5494-19 - 1	SB-21	3442	0.6284	-	7471B	PPB	08 Nov 2022 16:48:48	U	SMPL
N5508-01 - 1	AU-06-110722	7129	1.3336	-	7471B	PPB	08 Nov 2022 16:51:04	U	SMPL
CCV80 - 1	CCV80	27263	5.1849	-	7471B	PPB	08 Nov 2022 16:53:19	U	SMPL
CCB80 - 1	CCB80	138	-0.0036	-	7471B	PPB	08 Nov 2022 16:55:35	U	SMPL
PB148905BL - 1	PBS	48	-0.0209	-	7471B	PPB	08 Nov 2022 16:57:50	U	SMPL
PB148905BS - 1	LCSS	20025	3.8004	-	7471B	PPB	08 Nov 2022 17:00:06	U	SMPL
N5495-01 - 1	SB-32A	7658	1.4348	-	7471B	PPB	08 Nov 2022 17:02:21	U	SMPL
N5495-02 - 1	SB-32B	1612	0.2783	-	7471B	PPB	08 Nov 2022 17:04:37	U	SMPL
N5495-04 - 1	SB-31A	5897	1.0980	-	7471B	PPB	08 Nov 2022 17:06:52	U	SMPL
N5495-05 - 1	SB-31B	898	0.1417	-	7471B	PPB	08 Nov 2022 17:09:08	U	SMPL
N5495-07 - 1	SB-29A	5825	1.0842	-	7471B	PPB	08 Nov 2022 17:11:25	U	SMPL
N5495-08 - 1	SB-29B	530	0.0713	-	7471B	PPB	08 Nov 2022 17:13:42	U	SMPL
N5495-10 - 1	DUP-01	404	0.0472	-	7471B	PPB	08 Nov 2022 17:15:58	U	SMPL
N5495-11 - 1	SB-14	351	0.0371	-	7471B	PPB	08 Nov 2022 17:18:14	U	SMPL
CCV81 - 1	CCV81	26555	5.0494	-	7471B	PPB	08 Nov 2022 17:20:30	U	SMPL
CCB81 - 1	CCB81	102	-0.0105	-	7471B	PPB	08 Nov 2022 17:22:45	U	SMPL
N5501-01 - 1	HS-3-110722	8044	1.5086	-	7471B	PPB	08 Nov 2022 17:25:00	U	SMPL
N5501-03 - 1	HS-4-110722	16785	3.1806	-	7471B	PPB	08 Nov 2022 17:27:16	U	SMPL
N5502-01 - 1	2182	29086	5.5336	-	7471B	PPB	08 Nov 2022 17:29:31	U	SMPL
N5505-01 - 1	HD-2-110722	6673	1.2464	-	7471B	PPB	08 Nov 2022 17:31:47	U	SMPL
N5506-01 - 1	TP-5	6064	1.1299	-	7471B	PPB	08 Nov 2022 17:34:03	U	SMPL
N5506-04 - 1	TP-6	418	0.0499	-	7471B	PPB	08 Nov 2022 17:36:19	U	SMPL
N5506-07 - 1	TP-8	608	0.0863	-	7471B	PPB	08 Nov 2022 17:38:35	U	SMPL
N5506-10 - 1	TP-9	5284	0.9807	-	7471B	PPB	08 Nov 2022 17:40:52	U	SMPL
N5506-13 - 1	TP-10	4433	0.8179	-	7471B	PPB	08 Nov 2022 17:43:09	U	SMPL
N5511-01 - 1	PL-03-110722	2888	0.5224	-	7471B	PPB	08 Nov 2022 17:45:24	U	SMPL
CCV82 - 1	CCV82	26436	5.0267	-	7471B	PPB	08 Nov 2022 17:47:40	U	SMPL
CCB82 - 1	CCB82	35	-0.0233	-	7471B	PPB	08 Nov 2022 17:49:56	U	SMPL
N5511-03 - 1	PL-04-110722	14716	2.7849	-	7471B	PPB	08 Nov 2022 17:52:11	U	SMPL
N5511-05 - 1	PL-05-110722	1125	0.1852	-	7471B	PPB	08 Nov 2022 17:54:27	U	SMPL
N5511-05DUP - 1	PL-05-110722DUP	1360	0.2301	-	7471B	PPB	08 Nov 2022 17:56:42	U	SMPL
N5511-05MS - 1	PL-05-110722MS	22771	4.3256	-	7471B	PPB	08 Nov 2022 17:58:58	U	SMPL
N5511-05MSD - 1	PL-05-110722MSD	22639	4.3004	-	7471B	PPB	08 Nov 2022 18:01:14	U	SMPL
N5494-01LX5 - 1		74	-0.0159	-	7471B	PPB	08 Nov 2022 18:03:30	U	SMPL
N5494-01A - 1		21948	4.1682	-	7471B	PPB	08 Nov 2022 18:05:46	U	SMPL
N5511-05LX5 - 1		126	-0.0059	-	7471B	PPB	08 Nov 2022 18:08:03	U	SMPL
N5511-05A - 1		22936	4.3572	-	7471B	PPB	08 Nov 2022 18:10:20	U	SMPL
CCV83 - 1	CCV83	26569	5.0521	-	7471B	PPB	08 Nov 2022 18:12:36	U	SMPL
CCB83 - 1	CCB83	28	-0.0247	-	7471B	PPB	08 Nov 2022 18:14:52	U	SMPL