



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 11/9/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 11/08/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 11/09/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122802

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N5508-01	AU-06-110722	1	0.92	8.67	9.59	8.79	90.8	
N5509-01	TP-7	2	0.92	9.01	9.93	8.93	88.9	
N5509-03	TP-7-EPH-2	3	0.92	8.76	9.68	8.69	88.7	
N5509-04	TP-13	4	0.91	8.83	9.74	8.74	88.7	
N5509-06	TP-13-EPH-2	5	0.92	8.83	9.75	8.91	90.5	
N5509-07	TP-12	6	0.93	8.64	9.57	8.8	91.1	
N5509-09	TP-12-EPH-2	7	0.91	8.98	9.89	9.1	91.2	
N5509-10	TP-11	8	0.9	8.60	9.5	8.77	91.5	
N5509-12	TP-11-EPH-2	9	0.9	9.01	9.91	9.18	91.9	
N5512-01	40179	10	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N5512-02	40180	11	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N5512-03	40181	12	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N5512-04	40184	13	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N5512-05	40185	14	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N5512-06	40186	15	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N5516-01	FIRST-HALF-NORTH-DRYING-BIN	16	0.88	9.10	9.98	7.59	73.7	
N5516-03	SECOND-HALF-SOUTH-DRYING-BIN	17	0.88	8.99	9.87	8.35	83.1	
N5517-01	SEDIMENT-BASIN	18	0.87	9.05	9.92	4.81	43.5	SLUDGE SAMPLE
N5518-01	PAT-02-110822	28	0.92	8.85	9.77	8.82	89.3	
N5518-02	PAT-02-110822-EPH-2	29	0.92	8.85	9.77	8.82	89.3	
N5518-03	PAT-03-110822	30	0.93	8.78	9.71	9.08	92.8	
N5518-04	PAT-03-110822-EPH-2	31	0.93	8.78	9.71	9.08	92.8	
N5521-01	WASTE	19	0.87	8.71	9.58	8.64	89.2	
N5521-02	VOC	20	0.86	8.86	9.72	8.75	89.1	
N5521-03	1	21	0.87	8.99	9.86	9.02	90.7	
N5521-04	2	22	0.86	8.97	9.83	8.9	89.6	
N5521-05	3	23	0.85	8.79	9.64	8.68	89.1	



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 11/9/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 11/08/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 11/09/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122802

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N5521-06	4	24	0.89	9.06	9.95	8.92	88.6	
N5521-07	5	25	0.86	8.99	9.85	8.99	90.4	
N5521-08	6	26	0.88	9.04	9.92	8.94	89.2	
N5523-01	BC258780-END-1-1	32	1.00	1.00	2.00	2.00	100.0	PILC REEL
N5523-02	BC258780-END-1-2	33	1.00	1.00	2.00	2.00	100.0	PILC REEL
N5523-03	BC260963-END-1-1	34	1.00	1.00	2.00	2.00	100.0	PILC REEL
N5523-04	BC260963-END-1-2	35	1.00	1.00	2.00	2.00	100.0	PILC REEL
N5523-05	BC260963-END-2-1	36	1.00	1.00	2.00	2.00	100.0	PILC REEL
N5523-06	BC260963-END-2-2	37	1.00	1.00	2.00	2.00	100.0	PILC REEL
N5524-01	72-11995	38	0.91	8.77	9.68	9.05	92.8	
N5524-02	72-11995-EPH-2	39	0.91	8.77	9.68	9.05	92.8	
N5524-03	2711	40	1.00	1.00	2.00	2.00	100.0	OILY DEBRIS
N5525-01	138A	41	1.00	1.00	2.00	2.00	100.0	PILC REEL
N5525-02	138B	42	1.00	1.00	2.00	2.00	100.0	PILC REEL
N5525-03	139A	43	1.00	1.00	2.00	2.00	100.0	PILC REEL
N5525-04	139B	44	1.00	1.00	2.00	2.00	100.0	PILC REEL
N5526-01	63A	45	1.00	1.00	2.00	2.00	100.0	PILC REEL
N5526-02	63B	46	1.00	1.00	2.00	2.00	100.0	PILC REEL
N5526-03	64A	47	1.00	1.00	2.00	2.00	100.0	PILC REEL
N5526-04	64B	48	1.00	1.00	2.00	2.00	100.0	PILC REEL
N5528-01	SAMPLE-3	27	0.88	8.69	9.57	7.74	78.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$