CHEMTECH

Analytical Summary Report

Analysis Method: 9045D
Analyst By : JIGNESH
Parameter: Corrosivity
Supervisor Review By : Iwona
Run Number: LB122809
Slope : 99.2
BalanceID: WC SC-4
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	11/08/2022	17:30
2	CAL2	1	Water	NA	NA	20.7	7.00	11/08/2022	17:31
3	CAL3	1	Water	NA	NA	20.7	10.03	11/08/2022	17:32
4	ICV	1	Water	NA	NA	20.8	6.99	11/08/2022	17:35
5	CCV1	1	Water	NA	NA	20.8	2.01	11/08/2022	17:36
6	N5509-02	1	Solid	20.02	20	20.1	8.03	11/08/2022	17:40
7	N5509-05	1	Solid	20.03	20	20.7	8.80	11/08/2022	17:42
8	N5509-08	1	Solid	20.04	20	20.7	8.32	11/08/2022	17:45
9	N5509-11	1	Solid	20.02	20	20.2	8.04	11/08/2022	17:47
10	N5524-03	1	Solid	20.03	20	21.8	5.10	11/08/2022	17:50
11	N5529-02	1	Solid	20.02	20	20.2	6.70	11/08/2022	18:00
12	N5529-02DUP	1	Solid	20.03	20	20.3	6.71	11/08/2022	18:01
13	CCV2	1	Water	NA	NA	20.8	12.03	11/08/2022	18:05

VB 122809
VB/22809

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY N5507 **WorkList ID :** 164831 **Department :** Wet-Chemistry **Date :** 11-08-2022 12:06:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/08/2022	Solid	N5509-02	Corrosivity	Cool 4 deg C	PSEG03	G11	TP-7	11/08/2022	9045D
11/08/2022	Solid	N5509-05	Corrosivity	Cool 4 deg C	PSEG03	G11	TP-13	11/08/2022	9045D
11/08/2022	Solid	N5509-08	Corrosivity	Cool 4 deg C	PSEG03	G11	TP-12	11/08/2022	9045D
11/08/2022	Solid	N5509-11	Corrosivity	Cool 4 deg C	PSEG03	G11	TP-11	11/08/2022	9045D
11/08/2022	Solid	N5524-03	Corrosivity	Cool 4 deg C	PSEG03	G11	2711	11/08/2022	9045D
11/08/2022	Solid	N5529-02	Corrosivity	Cool 4 deg C	PSEG03	G11	NMS-SOIL-1108	11/08/2022	9045D

Date/Time 11/08/22 17:15
Raw Sample Received by: JP (CDC)
Raw Sample Relinquished by: [signature]

Date/Time 11/08/22 18:30
Raw Sample Received by: [signature]
Raw Sample Relinquished by: JP (CDC)