



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122811

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	11/09/2022	08:30
2	N5499-01		1	12	100.00	>212.0	>212.0	11/09/2022	09:00
3	N5499-01DUP		1	13	100.00	>212.0	>212.0	11/09/2022	10:00
4	N5500-01		1	10	100.00	>212.0	>212.0	11/09/2022	10:30
5	N5504-02		1	13	100.00	>212.0	>212.0	11/09/2022	11:00
6	N5512-07		1	12	100.00	>212.0	>212.0	11/09/2022	11:30
7	N5512-09		1	14	100.00	>212.0	>212.0	11/09/2022	12:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

26122811

WorkList Name : fp11-8

WorkList ID : 164834

Department : Wet-Chemistry

Date : 11-08-2022 14:41:48

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/14/2022	Water	N5499-01	Flash Point	Cool 4 deg C	PSEG03	F11	SCP122869-TOTE	11/07/2022	1010B
11/14/2022	Water	N5500-01	Flash Point	Cool 4 deg C	PSEG03	F11	NWB-1851	11/07/2022	1010B
11/10/2022	Water	N5504-02	Flash Point	Cool 4 deg C	TULL02	F11	SV34211L	11/07/2022	1010B
11/14/2022	Water	N5512-07	Flash Point	Cool 4 deg C	PSEG03	G11	AUD-1911-TOTE	11/07/2022	1010B
11/14/2022	Water	N5512-09	Flash Point	Cool 4 deg C	PSEG03	G11	AUD-1896-TOTE	11/07/2022	1010B

Date/Time 11/9/22 12:30
 Raw Sample Received by: RM (wec)
 Raw Sample Relinquished by: SJ (SM)

Date/Time 11/9/22 12:30
 Raw Sample Received by: SJ (SM)
 Raw Sample Relinquished by: RM (wec)