

L03122012


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

45412



A= 0.0000e+000

B= 2.1829e-004

C= -2.7244e-002

Rho= 0.9996006

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01)
0.0	0.000	-0.027	-0.027	0	0.000	0					
0.2	0.200	0.237	0.037	1209	0.0 %	1209					19
2.5	2.500	2.488	-0.012	11520	0.0 %	11520					0
5.0	5.000	4.912	-0.088	22625	0.0 %	22625					-2
7.5	7.500	7.705	0.205	35423	0.0 %	35423					3
10.0	10.000	9.886	-0.114	45412	0.0 %	45412					-1

LB122812 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Nov 2022 10:50:27

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	0	-	0.0000	7470A	PPB	09 Nov 2022 11:01:34	S	Std
0.2 - 1	SD-2	1209	-	0.2000	7470A	PPB	09 Nov 2022 11:03:51	S	Std
2.5 - 1	SD-5	11520	-	2.5000	7470A	PPB	09 Nov 2022 11:06:08	S	Std
5.0 - 1	SD-5	22625	-	5.0000	7470A	PPB	09 Nov 2022 11:08:25	S	Std
7.5 - 1	SD-5	35423	-	7.5000	7470A	PPB	09 Nov 2022 11:10:41	S	Std
10.0 - 1	SD-5	45412	-	10.0000	7470A	PPB	09 Nov 2022 11:12:57	S	Std
ICV14 - 1	ICV14	17231	3.7342	-	7470A	PPB	09 Nov 2022 11:17:20	U	SMPL
ICB14 - 1	ICB14	41	-0.0183	-	7470A	PPB	09 Nov 2022 11:19:35	U	SMPL
CCV84 - 1	CCV84	21669	4.7030	-	7470A	PPB	09 Nov 2022 11:21:50	U	SMPL
CCB84 - 1	CCB84	-52	-0.0386	-	7470A	PPB	09 Nov 2022 11:26:04	U	SMPL
CRA - 1	CRA	1051	0.2022	-	7470A	PPB	09 Nov 2022 11:28:18	U	SMPL
HighStd - 1	HighStd	45365	9.8757	-	7470A	PPB	09 Nov 2022 11:30:33	U	SMPL
ChkStd - 1	ChkStd	30960	6.7312	-	7470A	PPB	09 Nov 2022 11:35:27	U	SMPL
PB148914BL - 1	PBW	17	-0.0235	-	7470A	PPB	09 Nov 2022 11:37:42	U	SMPL
PB148914BS - 1	LCSW	17599	3.8145	-	7470A	PPB	09 Nov 2022 11:43:11	U	SMPL
N5318-13 - 1	SPLP-LEACHATE-AMENDED-COMP-04	179	0.0118	-	7470A	PPB	09 Nov 2022 11:45:27	U	SMPL
N5318-14 - 1	SPLP-LEACHATE-AMENDED-COMP-05	158	0.0072	-	7470A	PPB	09 Nov 2022 11:47:43	U	SMPL
N5318-15 - 1	SPLP-LEACHATE-AMENDED-COMP-06	186	0.0134	-	7470A	PPB	09 Nov 2022 11:50:00	U	SMPL
N5318-16 - 1	SPLP-LEACHATE-AMENDED-COMP-07	152	0.0059	-	7470A	PPB	09 Nov 2022 11:52:17	U	SMPL
N5318-17 - 1	SPLP-LEACHATE-AMENDED-COMP-10	142	0.0038	-	7470A	PPB	09 Nov 2022 11:54:33	U	SMPL
CCV85 - 1	CCV85	21292	4.6207	-	7470A	PPB	09 Nov 2022 11:56:48	U	SMPL
CCB85 - 1	CCB85	-95	-0.0480	-	7470A	PPB	09 Nov 2022 11:59:03	U	SMPL
N5318-18 - 1	SPLP-LEACHATE-AMENDED-COMP-14	232	0.0234	-	7470A	PPB	09 Nov 2022 12:01:18	U	SMPL
N5318-18DUP - 1	SPLP-LEACHATE-AMENDED-COMP-14DUP	224	0.0217	-	7470A	PPB	09 Nov 2022 12:03:33	U	SMPL
N5318-18MS - 1	SPLP-LEACHATE-AMENDED-COMP-14MS	18129	3.9302	-	7470A	PPB	09 Nov 2022 12:12:23	U	SMPL
N5318-18MSD - 1	SPLP-LEACHATE-AMENDED-COMP-14MSD	18193	3.9442	-	7470A	PPB	09 Nov 2022 12:14:39	U	SMPL
PB148915BL - 1	PBW	-36	-0.0351	-	7470A	PPB	09 Nov 2022 12:16:55	U	SMPL
PB148915BS - 1	LCSW	17987	3.8992	-	7470A	PPB	09 Nov 2022 12:19:11	U	SMPL
N5482-03 - 1	FLOOR-DRAIN-WATER	742	0.1347	-	7470A	PPB	09 Nov 2022 12:21:27	U	SMPL
N5504-02 - 1	SV34211L	224	0.0217	-	7470A	PPB	09 Nov 2022 12:23:44	U	SMPL
N5504-02DUP - 1	SV34211LDUP	209	0.0184	-	7470A	PPB	09 Nov 2022 12:26:01	U	SMPL
N5504-02MS - 1	SV34211LMS	17371	3.7648	-	7470A	PPB	09 Nov 2022 12:28:17	U	SMPL
CCV86 - 1	CCV86	21424	4.6495	-	7470A	PPB	09 Nov 2022 12:30:32	U	SMPL
CCB86 - 1	CCB86	-65	-0.0414	-	7470A	PPB	09 Nov 2022 12:32:47	U	SMPL
N5504-02MSD - 1	SV34211LMSD	16474	3.5689	-	7470A	PPB	09 Nov 2022 12:35:03	U	SMPL
PB148925BL - 1	PBW	253	0.0280	-	7470A	PPB	09 Nov 2022 12:37:18	U	SMPL
PB148925BS - 1	LCSW	14973	3.2413	-	7470A	PPB	09 Nov 2022 12:39:34	U	SMPL
N5504-01 - 1	SV34211L	28	-0.0211	-	7470A	PPB	09 Nov 2022 12:41:50	U	SMPL
N5504-01DUP - 1	SV34211LDUP	69	-0.0122	-	7470A	PPB	09 Nov 2022 12:44:05	U	SMPL
N5504-01MS - 1	SV34211LMS	12673	2.7392	-	7470A	PPB	09 Nov 2022 12:46:21	U	SMPL
N5504-01MSD - 1	SV34211LMSD	13033	2.8178	-	7470A	PPB	09 Nov 2022 12:48:38	U	SMPL
N5512-07 - 1	AUD-1911-TOTE	-11	-0.0296	-	7470A	PPB	09 Nov 2022 12:50:55	U	SMPL
N5512-09 - 1	AUD-1896-TOTE	3376	0.7097	-	7470A	PPB	09 Nov 2022 12:53:12	U	SMPL
N5514-04 - 1	MW-6I	13	-0.0244	-	7470A	PPB	09 Nov 2022 12:55:27	U	SMPL
CCV87 - 1	CCV87	22059	4.7881	-	7470A	PPB	09 Nov 2022 12:57:43	U	SMPL
CCB87 - 1	CCB87	-102	-0.0495	-	7470A	PPB	09 Nov 2022 12:59:58	U	SMPL
N5514-05 - 1	MW-6I	69	-0.0122	-	7470A	PPB	09 Nov 2022 13:02:14	U	SMPL
N5514-08 - 1	MW-17	3	-0.0266	-	7470A	PPB	09 Nov 2022 13:04:29	U	SMPL
N5514-09 - 1	MW-17	-9	-0.0292	-	7470A	PPB	09 Nov 2022 13:06:45	U	SMPL
PB148914TB - 1	PB148849TB	-80	-0.0447	-	7470A	PPB	09 Nov 2022 13:09:00	U	SMPL
PB148886TB - 1	PB148886TB	-81	-0.0449	-	7470A	PPB	09 Nov 2022 13:11:16	U	SMPL
CCV88 - 1	CCV88	23277	5.0540	-	7470A	PPB	09 Nov 2022 13:19:42	U	SMPL
CCB88 - 1	CCB88	-17	-0.0310	-	7470A	PPB	09 Nov 2022 13:21:58	U	SMPL
PB148482TB - 1	PB148482TB	4	-0.0264	-	7470A	PPB	09 Nov 2022 13:24:15	U	SMPL
N5318-18LX5 - 1		7	-0.0257	-	7470A	PPB	09 Nov 2022 13:26:32	U	SMPL
N5318-18A - 1		19251	4.1752	-	7470A	PPB	09 Nov 2022 13:28:47	U	SMPL
N5504-02LX5 - 1		-70	-0.0425	-	7470A	PPB	09 Nov 2022 13:31:03	U	SMPL
N5504-02A - 1		18060	3.9152	-	7470A	PPB	09 Nov 2022 13:33:18	U	SMPL
N5504-01LX5 - 1		-70	-0.0425	-	7470A	PPB	09 Nov 2022 13:35:34	U	SMPL
N5504-01A - 1		13013	2.8134	-	7470A	PPB	09 Nov 2022 13:37:49	U	SMPL
CCV89 - 1	CCV89	23551	5.1138	-	7470A	PPB	09 Nov 2022 13:40:05	U	SMPL
CCB89 - 1	CCB89	-83	-0.0454	-	7470A	PPB	09 Nov 2022 13:42:20	U	SMPL