

7470A

Linear

 $\mu \text{ Abs.} :$

51112

A= 0.0000e+000

B= 1.9400e-004

C= 7.3152e-003

Rho= 0.9996532

Accept=Accepted

-Max. Conc:— 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.004	0.004	-18	0.000	-18				
0.2	0.200	0.190	-0.010	944	0.0 %	944				
2.5	2.500	2.550	0.050	13107	0.0 %	13107				
5.0	5.000	4.862	-0.138	25024	0.0 %	25024				
7.5	7.500	7.671	0.171	39501	0.0 %	39501				
10.0	10.000	9.923	-0.077	51112	0.0 %	51112				

LB122833 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Nov 2022 10:07:30

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-18	-	0.0000	7470A	PPB	10 Nov 2022 10:13:32	S	Std
0.2 - 1	50.2	944	-	0.2000	7470A	PPB	10 Nov 2022 10:18:01	S	Std
2.5 - 1	52.5	13107	-	2.5000	7470A	PPB	10 Nov 2022 10:20:18	S	Std
5.0 - 1	55	25024	-	5.0000	7470A	PPB	10 Nov 2022 10:22:35	S	Std
7.5 - 1	57.5	39501	-	7.5000	7470A	PPB	10 Nov 2022 10:27:25	S	Std
10.0 - 1	110	51112	-	10.0000	7470A	PPB	10 Nov 2022 10:32:16	S	Std
ICV15 - 1	ICV15	19208	3.7337	-	7470A	PPB	10 Nov 2022 10:35:05	U	SMPL
ICB15 - 1	ICB15	87	0.0242	-	7470A	PPB	10 Nov 2022 10:37:20	U	SMPL
CCV90 - 1	CCV90	25252	4.9062	-	7470A	PPB	10 Nov 2022 10:39:35	U	SMPL
CCB90 - 1	CCB90	-78	-0.0078	-	7470A	PPB	10 Nov 2022 10:41:50	U	SMPL
CRA - 1	CRA	893	0.1806	-	7470A	PPB	10 Nov 2022 10:44:05	U	SMPL
HighStd - 1	HighStd	51571	10.0121	-	7470A	PPB	10 Nov 2022 10:46:21	U	SMPL
ChkStd - 1	ChkStd	37088	7.2024	-	7470A	PPB	10 Nov 2022 10:48:36	U	SMPL
PB148926BL - 1	PBW	21	0.0114	-	7470A	PPB	10 Nov 2022 10:55:57	U	SMPL
PB148926BS - 1	LCSW	18909	3.6757	-	7470A	PPB	10 Nov 2022 10:58:12	U	SMPL
N5494-02 - 1	WC-19	305	0.0665	-	7470A	PPB	10 Nov 2022 11:05:18	U	SMPL
N5494-04 - 1	WC-18	428	0.0903	-	7470A	PPB	10 Nov 2022 11:07:34	U	SMPL
N5494-06 - 1	WC-17	331	0.0715	-	7470A	PPB	10 Nov 2022 11:09:51	U	SMPL
N5494-08 - 1	WC-16	455	0.0956	-	7470A	PPB	10 Nov 2022 11:12:07	U	SMPL
N5494-10 - 1	WC-15	444	0.0935	-	7470A	PPB	10 Nov 2022 11:14:23	U	SMPL
CCV91 - 1	CCV91	25202	4.8965	-	7470A	PPB	10 Nov 2022 11:16:38	U	SMPL
CCB91 - 1	CCB91	-41	-0.0006	-	7470A	PPB	10 Nov 2022 11:18:53	U	SMPL
N5494-12 - 1	WC-24	538	0.1117	-	7470A	PPB	10 Nov 2022 11:21:09	U	SMPL
N5494-14 - 1	WC-23	597	0.1231	-	7470A	PPB	10 Nov 2022 11:23:24	U	SMPL
N5494-16 - 1	WC-22	294	0.0644	-	7470A	PPB	10 Nov 2022 11:25:39	U	SMPL
N5494-18 - 1	WC-20	271	0.0599	-	7470A	PPB	10 Nov 2022 11:27:55	U	SMPL
N5494-20 - 1	WC-21	95	0.0257	-	7470A	PPB	10 Nov 2022 11:30:10	U	SMPL
N5495-03 - 1	WC-32	307	0.0669	-	7470A	PPB	10 Nov 2022 11:35:00	U	SMPL
N5495-06 - 1	WC-31	336	0.0725	-	7470A	PPB	10 Nov 2022 11:37:16	U	SMPL
N5495-09 - 1	WC-29	373	0.0797	-	7470A	PPB	10 Nov 2022 11:39:33	U	SMPL
N5495-12 - 1	WC-14	498	0.1039	-	7470A	PPB	10 Nov 2022 11:41:50	U	SMPL
N5502-02 - 1	2182	258	0.0574	-	7470A	PPB	10 Nov 2022 11:44:06	U	SMPL
CCV92 - 1	CCV92	24974	4.8523	-	7470A	PPB	10 Nov 2022 11:46:21	U	SMPL
CCB92 - 1	CCB92	-46	-0.0016	-	7470A	PPB	10 Nov 2022 11:48:37	U	SMPL
N5506-02 - 1	TP-5	214	0.0488	-	7470A	PPB	10 Nov 2022 11:53:33	U	SMPL
N5506-05 - 1	TP-6	373	0.0797	-	7470A	PPB	10 Nov 2022 11:55:49	U	SMPL
N5506-08 - 1	TP-8	312	0.0678	-	7470A	PPB	10 Nov 2022 11:58:04	U	SMPL
N5506-11 - 1	TP-9	380	0.0810	-	7470A	PPB	10 Nov 2022 12:00:20	U	SMPL
N5506-14 - 1	TP-10	426	0.0900	-	7470A	PPB	10 Nov 2022 12:02:36	U	SMPL
N5506-14DUP - 1	TP-10DUP	315	0.0684	-	7470A	PPB	10 Nov 2022 12:04:53	U	SMPL
N5506-14MS - 1	TP-10MS	18109	3.5205	-	7470A	PPB	10 Nov 2022 12:07:10	U	SMPL
N5506-14MSD - 1	TP-10MSD	19034	3.6999	-	7470A	PPB	10 Nov 2022 12:09:27	U	SMPL
CCV93 - 1	CCV93	25005	4.8583	-	7470A	PPB	10 Nov 2022 12:11:44	U	SMPL
CCB93 - 1	CCB93	6	0.0085	-	7470A	PPB	10 Nov 2022 12:14:00	U	SMPL
PB148907TB - 1	PB148907TB	418	0.0884	-	7470A	PPB	10 Nov 2022 12:16:15	U	SMPL
N5506-14LX5 - 1		-31	0.0013	-	7470A	PPB	10 Nov 2022 12:21:53	U	SMPL
N5506-14A - 1		19011	3.6955	-	7470A	PPB	10 Nov 2022 12:24:08	U	SMPL
CCV94 - 1	CCV94	28014	5.4421	-	7470A	PPB	10 Nov 2022 12:26:24	U	SMPL
CCB94 - 1	CCB94	-42	-0.0008	-	7470A	PPB	10 Nov 2022 12:28:40	U	SMPL