

LB122838


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

52358



A= 0.0000e+000

B= 1.9087e-004

C= -4.0707e-002

Rho= 0.9999205

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	U/D
0.0	0.000	-0.030	-0.030	55	0.000	55					
0.2	0.200	0.212	0.012	1325	0.0 %	1325					
2.5	2.500	2.530	0.030	13467	0.0 %	13467					
5.0	5.000	4.954	-0.046	26166	0.0 %	26166					
7.5	7.500	7.582	0.082	39935	0.0 %	39935					
10.0	10.000	9.953	-0.047	52358	0.0 %	52358					

U/D
6
-1
1
0

LB122838 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Nov 2022 14:26:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	55	-	0.0000	7471B	PPB	10 Nov 2022 14:42:02	S	Std
0.2 - 1	50.2	1325	-	0.2000	7471B	PPB	10 Nov 2022 14:44:19	S	Std
2.5 - 1	52.5	13467	-	2.5000	7471B	PPB	10 Nov 2022 14:46:36	S	Std
5.0 - 1	55	26166	-	5.0000	7471B	PPB	10 Nov 2022 14:48:53	S	Std
7.5 - 1	57.5	39935	-	7.5000	7471B	PPB	10 Nov 2022 14:51:09	S	Std
10.0 - 1	50	52358	-	10.0000	7471B	PPB	10 Nov 2022 14:53:25	S	Std
ICV16 - 1	ICV16	19916	3.7607	-	7471B	PPB	10 Nov 2022 14:56:13	U	SMPL
ICB16 - 1	ICB16	55	-0.0302	-	7471B	PPB	10 Nov 2022 14:58:28	U	SMPL
CCV01 - 1	CCV01	27021	5.1168	-	7471B	PPB	10 Nov 2022 15:00:43	U	SMPL
CCB01 - 1	CCB01	-27	-0.0459	-	7471B	PPB	10 Nov 2022 15:02:58	U	SMPL
CRA - 1	CRA	1351	0.2172	-	7471B	PPB	10 Nov 2022 15:05:13	U	SMPL
HighStd - 1	HighStd	52084	9.9006	-	7471B	PPB	10 Nov 2022 15:07:29	U	SMPL
ChkStd - 1	ChkStd	38361	7.2813	-	7471B	PPB	10 Nov 2022 15:09:44	U	SMPL
PB148957BL - 1	PBS	-75	-0.0550	-	7471B	PPB	10 Nov 2022 15:12:00	U	SMPL
PB148957BS - 1	LCSS	21422	4.0481	-	7471B	PPB	10 Nov 2022 15:14:16	U	SMPL
N5509-01 - 1	TP-7	787	0.1095	-	7471B	PPB	10 Nov 2022 15:16:32	U	SMPL
N5509-04 - 1	TP-13	4467	0.8119	-	7471B	PPB	10 Nov 2022 15:18:49	U	SMPL
N5509-07 - 1	TP-12	7496	1.3901	-	7471B	PPB	10 Nov 2022 15:21:05	U	SMPL
N5509-10 - 1	TP-11	3100	0.5510	-	7471B	PPB	10 Nov 2022 15:23:22	U	SMPL
N5518-01 - 1	PAT-02-110822	5041	0.9215	-	7471B	PPB	10 Nov 2022 15:25:37	U	SMPL
CCV02 - 1	CCV02	26720	5.0594	-	7471B	PPB	10 Nov 2022 15:27:53	U	SMPL
CCB02 - 1	CCB02	-100	-0.0598	-	7471B	PPB	10 Nov 2022 15:30:08	U	SMPL
N5518-03 - 1	PAT-03-110822	2783	0.4905	-	7471B	PPB	10 Nov 2022 15:32:23	U	SMPL
N5520-01 - 1	124035	3443	0.6165	-	7471B	PPB	10 Nov 2022 15:34:38	U	SMPL
N5520-03 - 1	124037	2056	0.3517	-	7471B	PPB	10 Nov 2022 15:36:54	U	SMPL
N5520-05 - 1	124042	2561	0.4481	-	7471B	PPB	10 Nov 2022 15:39:09	U	SMPL
N5520-07 - 1	R20-3	2458	0.4285	-	7471B	PPB	10 Nov 2022 15:41:25	U	SMPL
N5524-01 - 1	72-11995	5861	1.0780	-	7471B	PPB	10 Nov 2022 15:48:37	U	SMPL
N5524-01DUP - 1	72-11995DUP	5814	1.0690	-	7471B	PPB	10 Nov 2022 15:50:53	U	SMPL
N5524-01MS - 1	72-11995MS	28659	5.4295	-	7471B	PPB	10 Nov 2022 15:56:12	U	SMPL
N5524-01MSD - 1	72-11995MSD	26194	4.9590	-	7471B	PPB	10 Nov 2022 15:58:28	U	SMPL
N5528-01 - 1	SAMPLE-3	77361	14.7253	-	7471B	PPB	10 Nov 2022 16:00:44	U	SMPL
CCV03 - 1	CCV03	26798	5.0743	-	7471B	PPB	10 Nov 2022 16:02:59	U	SMPL
CCB03 - 1	CCB03	20	-0.0369	-	7471B	PPB	10 Nov 2022 16:05:16	U	SMPL
N5529-01 - 1	NMS-SOIL-1108	1304	0.2082	-	7471B	PPB	10 Nov 2022 16:07:32	U	SMPL
N5529-04 - 1	NMS-CONCRETE-1108	938	0.1383	-	7471B	PPB	10 Nov 2022 16:09:47	U	SMPL
N5537-01 - 1	360	3605	0.6474	-	7471B	PPB	10 Nov 2022 16:12:03	U	SMPL
N5554-01 - 1	RBR251374	2571	0.4500	-	7471B	PPB	10 Nov 2022 16:14:18	U	SMPL
N5554-03 - 1	RT3249	3267	0.5829	-	7471B	PPB	10 Nov 2022 16:16:34	U	SMPL
N5561-01 - 1	SB-25	2920	0.5166	-	7471B	PPB	10 Nov 2022 16:18:50	U	SMPL
N5561-03 - 1	SB-06	4816	0.8785	-	7471B	PPB	10 Nov 2022 16:21:06	U	SMPL
N5561-05 - 1	SB-05	7037	1.3025	-	7471B	PPB	10 Nov 2022 16:23:23	U	SMPL
PB148958BL - 1	PBS	-66	-0.0533	-	7471B	PPB	10 Nov 2022 16:25:40	U	SMPL
PB148958BS - 1	LCSS	19675	3.7147	-	7471B	PPB	10 Nov 2022 16:27:56	U	SMPL
CCV04 - 1	CCV04	26814	5.0773	-	7471B	PPB	10 Nov 2022 16:30:11	U	SMPL
CCB04 - 1	CCB04	-77	-0.0554	-	7471B	PPB	10 Nov 2022 16:32:27	U	SMPL
N5530-01 - 1	SB-13	1797	0.3023	-	7471B	PPB	10 Nov 2022 16:34:42	U	SMPL
N5530-03 - 1	SB-12	950	0.1406	-	7471B	PPB	10 Nov 2022 16:36:58	U	SMPL
N5530-05 - 1	SB-11	730	0.0986	-	7471B	PPB	10 Nov 2022 16:39:13	U	SMPL
N5530-07 - 1	SB-10	1310	0.2093	-	7471B	PPB	10 Nov 2022 16:41:29	U	SMPL
N5530-09 - 1	SB-09	1371	0.2210	-	7471B	PPB	10 Nov 2022 16:43:45	U	SMPL
N5531-01 - 1	DUP03	1399	0.2263	-	7471B	PPB	10 Nov 2022 16:46:01	U	SMPL
N5531-02 - 1	SB-08	854	0.1223	-	7471B	PPB	10 Nov 2022 16:48:18	U	SMPL
N5531-04 - 1	SB-07	1550	0.2551	-	7471B	PPB	10 Nov 2022 16:50:35	U	SMPL
N5531-06 - 1	SB-34	4799	0.8753	-	7471B	PPB	10 Nov 2022 16:52:52	U	SMPL
N5531-08 - 1	SB-33	8225	1.5292	-	7471B	PPB	10 Nov 2022 16:55:08	U	SMPL
CCV05 - 1	CCV05	26230	4.9658	-	7471B	PPB	10 Nov 2022 16:57:23	U	SMPL
CCB05 - 1	CCB05	-51	-0.0504	-	7471B	PPB	10 Nov 2022 16:59:39	U	SMPL
N5531-10 - 1	SB-30	4132	0.7480	-	7471B	PPB	10 Nov 2022 17:01:54	U	SMPL
N5531-12 - 1	SB-28A	6313	1.1643	-	7471B	PPB	10 Nov 2022 17:04:10	U	SMPL
N5531-13 - 1	SB-28B	607	0.0752	-	7471B	PPB	10 Nov 2022 17:06:25	U	SMPL
N5531-15 - 1	SB-27	6135	1.1303	-	7471B	PPB	10 Nov 2022 17:08:41	U	SMPL
N5531-17 - 1	SB-26	3935	0.7104	-	7471B	PPB	10 Nov 2022 17:10:56	U	SMPL
N5538-01 - 1	TP-14-A-WEST	5235	0.9585	-	7471B	PPB	10 Nov 2022 17:13:13	U	SMPL
N5538-03 - 1	TP-14-B-EAST	1385	0.2236	-	7471B	PPB	10 Nov 2022 17:15:29	U	SMPL
N5545-01 - 1	CL-1-110922	2642	0.4636	-	7471B	PPB	10 Nov 2022 17:17:46	U	SMPL
N5545-01DUP - 1	CL-1-110922DUP	2631	0.4615	-	7471B	PPB	10 Nov 2022 17:20:03	U	SMPL
N5545-01MS - 1	CL-1-110922MS	21150	3.9962	-	7471B	PPB	10 Nov 2022 17:22:18	U	SMPL
CCV06 - 1	CCV06	26274	4.9742	-	7471B	PPB	10 Nov 2022 17:24:34	U	SMPL
CCB06 - 1	CCB06	-65	-0.0531	-	7471B	PPB	10 Nov 2022 17:26:51	U	SMPL
N5545-01MSD - 1	CL-1-110922MSD	22463	4.2468	-	7471B	PPB	10 Nov 2022 17:29:07	U	SMPL
N5524-01LX5 - 1		1188	0.1860	-	7471B	PPB	10 Nov 2022 17:31:23	U	SMPL
N5524-01A - 1		24229	4.5839	-	7471B	PPB	10 Nov 2022 17:33:39	U	SMPL
N5545-01LX5 - 1		525	0.0595	-	7471B	PPB	10 Nov 2022 17:35:55	U	SMPL
N5545-01A - 1		21839	4.1277	-	7471B	PPB	10 Nov 2022 17:38:11	U	SMPL
N5528-01DLX5 - 1		16730	3.1526	-	7471B	PPB	10 Nov 2022 17:40:28	U	SMPL
CCV07 - 1	CCV07	26344	4.9876	-	7471B	PPB	10 Nov 2022 17:42:44	U	SMPL
CCB07 - 1	CCB07	-55	-0.0512	-	7471B	PPB	10 Nov 2022 17:45:01	U	SMPL