



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB122852

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	11/11/2022	08:30
2	N5572-03		1	12	100.00	>212.0	>212.0	11/11/2022	09:00
3	N5572-03DUP		1	11	100.00	>212.0	>212.0	11/11/2022	09:30
4	N5572-04		1	13	100.00	>212.0	>212.0	11/11/2022	10:00
5	N5572-05		1	14	100.00	>212.0	>212.0	11/11/2022	10:30
6	N5572-06		1	12	100.00	>212.0	>212.0	11/11/2022	11:00
7	N5572-07		1	13	100.00	>212.0	>212.0	11/11/2022	11:30
8	N5582-06		1	11	100.00	>212.0	>212.0	11/11/2022	12:00
9	N5582-07		1	13	100.00	>212.0	>212.0	11/11/2022	12:30
10	N5582-08		1	11	100.00	>212.0	>212.0	11/11/2022	13:00
11	N5582-09		1	12	100.00	>212.0	>212.0	11/11/2022	13:30
12	N5582-10		1	13	100.00	>212.0	>212.0	11/11/2022	14:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

Lb122852

WorkList Name : fp11-11

WorkList ID : 164933

Department : Wet-Chemistry

Date : 11-11-2022 08:21:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/15/2022	Water	N5572-03	Flash Point	Cool 4 deg C	PSEG03	G11	31914	11/10/2022	1010B
11/15/2022	Water	N5572-04	Flash Point	Cool 4 deg C	PSEG03	G11	31915	11/10/2022	1010B
11/15/2022	Water	N5572-05	Flash Point	Cool 4 deg C	PSEG03	G11	31910	11/10/2022	1010B
11/15/2022	Water	N5572-06	Flash Point	Cool 4 deg C	PSEG03	G11	31911	11/10/2022	1010B
11/15/2022	Water	N5572-07	Flash Point	Cool 4 deg C	PSEG03	G11	31912	11/10/2022	1010B
11/15/2022	Water	N5582-06	Flash Point	Cool 4 deg C	PSEG03	G11	31880	11/10/2022	1010B
11/15/2022	Water	N5582-07	Flash Point	Cool 4 deg C	PSEG03	G11	CLI-0659	11/10/2022	1010B
11/15/2022	Water	N5582-08	Flash Point	Cool 4 deg C	PSEG03	G11	CLI-0668	11/10/2022	1010B
11/15/2022	Water	N5582-09	Flash Point	Cool 4 deg C	PSEG03	G11	CLI-0669	11/10/2022	1010B
11/15/2022	Water	N5582-10	Flash Point	Cool 4 deg C	PSEG03	G11	CLI-0667	11/10/2022	1010B

Date/Time 11/11/22 08:25
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: SJ (SM)

Date/Time 11/11/22 14:30
 Raw Sample Received by: SJ (SM)
 Raw Sample Relinquished by: RM CWC