

LB122853


7470A

INSTRUMENT ID: LV2

Linear

 μ Abs.:

43429



A= 0.0000e+000

B= 2.3496e-004

C= 1.2042e-002

Rho= 0.9988140

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/σ 1)
0.0	0.000	0.025	0.025	57	0.000	57					-
0.2	0.200	0.196	-0.004	783	0.0 %	783					-2
2.5	2.500	2.710	0.210	11483	0.0 %	11483					8
5.0	5.000	4.809	-0.191	20415	0.0 %	20415					-4
7.5	7.500	7.244	-0.256	30779	0.0 %	30779					-3
10.0	10.000	10.216	0.216	43429	0.0 %	43429					2

LB122853 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Nov 2022 09:27:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	57	-	0.0000	7470A	PPB	11 Nov 2022 09:58:59	S	Std
0.2 - 1	51.2	783	-	0.2000	7470A	PPB	11 Nov 2022 10:01:16	S	Std
2.5 - 1	52.5	11483	-	2.5000	7470A	PPB	11 Nov 2022 10:03:34	S	Std
5.0 - 1	55	20415	-	5.0000	7470A	PPB	11 Nov 2022 10:05:51	S	Std
7.5 - 1	57.5	30779	-	7.5000	7470A	PPB	11 Nov 2022 10:08:07	S	Std
10.0 - 1	59.0	43429	-	10.0000	7470A	PPB	11 Nov 2022 10:12:59	S	Std
ICV17 - 1	ICV17	16351	3.8538	-	7470A	PPB	11 Nov 2022 10:18:25	U	SMPL
ICB17 - 1	ICB17	-37	0.0033	-	7470A	PPB	11 Nov 2022 10:20:39	U	SMPL
CCV08 - 1	CCV08	20904	4.9236	-	7470A	PPB	11 Nov 2022 10:33:47	U	SMPL
CCB08 - 1	CCB08	-127	-0.0178	-	7470A	PPB	11 Nov 2022 10:36:02	U	SMPL
CRA - 1	CRA	747	0.1876	-	7470A	PPB	11 Nov 2022 10:38:17	U	SMPL
HighStd - 1	HighStd	41599	9.7860	-	7470A	PPB	11 Nov 2022 10:40:32	U	SMPL
ChkStd - 1	ChkStd	29120	6.8540	-	7470A	PPB	11 Nov 2022 10:42:47	U	SMPL
PB148973BL - 1	PBW	-54	-0.0006	-	7470A	PPB	11 Nov 2022 10:45:04	U	SMPL
PB148973BS - 1	LCSW	16168	3.8108	-	7470A	PPB	11 Nov 2022 10:47:20	U	SMPL
N5509-02 - 1	TP-7	0	0.0120	-	7470A	PPB	11 Nov 2022 10:49:36	U	SMPL
N5509-02DUP - 1	TP-7DUP	-2	0.0116	-	7470A	PPB	11 Nov 2022 10:51:53	U	SMPL
N5509-02MS - 1	TP-7MS	15709	3.7030	-	7470A	PPB	11 Nov 2022 10:54:09	U	SMPL
N5509-02MSD - 1	TP-7MSD	15550	3.6656	-	7470A	PPB	11 Nov 2022 10:56:26	U	SMPL
N5509-05 - 1	TP-13	-29	0.0052	-	7470A	PPB	11 Nov 2022 10:58:41	U	SMPL
CCV09 - 1	CCV09	22634	5.3301	-	7470A	PPB	11 Nov 2022 11:05:46	U	SMPL
CCB09 - 1	CCB09	-9	0.0099	-	7470A	PPB	11 Nov 2022 11:08:01	U	SMPL
N5509-08 - 1	TP-12	64	0.0271	-	7470A	PPB	11 Nov 2022 11:10:16	U	SMPL
N5509-11 - 1	TP-11	40	0.0214	-	7470A	PPB	11 Nov 2022 11:12:31	U	SMPL
N5520-02 - 1	124035	1	0.0123	-	7470A	PPB	11 Nov 2022 11:14:47	U	SMPL
N5520-04 - 1	124037	-19	0.0076	-	7470A	PPB	11 Nov 2022 11:17:02	U	SMPL
N5520-06 - 1	124042	8	0.0139	-	7470A	PPB	11 Nov 2022 11:19:18	U	SMPL
N5520-08 - 1	R20-3	14	0.0153	-	7470A	PPB	11 Nov 2022 11:21:34	U	SMPL
N5524-03 - 1	2711	36	0.0205	-	7470A	PPB	11 Nov 2022 11:23:50	U	SMPL
N5529-02 - 1	NMS-SOIL-1108	21	0.0170	-	7470A	PPB	11 Nov 2022 11:26:07	U	SMPL
N5529-05 - 1	NMS-CONCRETE-1108	34	0.0200	-	7470A	PPB	11 Nov 2022 11:28:24	U	SMPL
N5537-02 - 1	360	26	0.0182	-	7470A	PPB	11 Nov 2022 11:30:40	U	SMPL
CCV10 - 1	CCV10	20784	4.8954	-	7470A	PPB	11 Nov 2022 11:36:16	U	SMPL
CCB10 - 1	CCB10	35	0.0203	-	7470A	PPB	11 Nov 2022 11:38:31	U	SMPL
N5538-02 - 1	TP-14-A-WEST	108	0.0374	-	7470A	PPB	11 Nov 2022 11:40:47	U	SMPL
N5538-04 - 1	TP-14-B-EAST	45	0.0226	-	7470A	PPB	11 Nov 2022 11:43:03	U	SMPL
PB148974BL - 1	PBW	4	0.0130	-	7470A	PPB	11 Nov 2022 11:45:18	U	SMPL
PB148974BS - 1	LCSW	14772	3.4828	-	7470A	PPB	11 Nov 2022 11:47:34	U	SMPL
N5524-04 - 1	2712	24	0.0177	-	7470A	PPB	11 Nov 2022 11:49:49	U	SMPL
N5550-02 - 1	D2782	96	0.0346	-	7470A	PPB	11 Nov 2022 11:52:05	U	SMPL
N5553-03 - 1	8249	25	0.0179	-	7470A	PPB	11 Nov 2022 11:54:21	U	SMPL
N5553-03DUP - 1	8249DUP	45	0.0226	-	7470A	PPB	11 Nov 2022 11:56:38	U	SMPL
N5553-03MS - 1	8249MS	16723	3.9412	-	7470A	PPB	11 Nov 2022 11:58:55	U	SMPL
N5553-03MSD - 1	8249MSD	16032	3.7789	-	7470A	PPB	11 Nov 2022 12:01:11	U	SMPL
CCV11 - 1	CCV11	21219	4.9976	-	7470A	PPB	11 Nov 2022 12:03:26	U	SMPL
CCB11 - 1	CCB11	-2	0.0116	-	7470A	PPB	11 Nov 2022 12:05:42	U	SMPL
N5553-04 - 1	TOTE-COMP	133	0.0433	-	7470A	PPB	11 Nov 2022 12:07:57	U	SMPL
PB148975BL - 1	PBW	60	0.0261	-	7470A	PPB	11 Nov 2022 12:10:13	U	SMPL
PB148975BS - 1	LCSW	15507	3.6555	-	7470A	PPB	11 Nov 2022 12:12:28	U	SMPL
N5515-01 - 1	FMI121	8	0.0139	-	7470A	PPB	11 Nov 2022 12:14:44	U	SMPL
N5532-01 - 1	B-3(OW)	9	0.0142	-	7470A	PPB	11 Nov 2022 12:17:00	U	SMPL
N5532-03 - 1	DUP-06	32	0.0196	-	7470A	PPB	11 Nov 2022 12:19:16	U	SMPL
N5532-05 - 1	FB-01	13	0.0151	-	7470A	PPB	11 Nov 2022 12:21:32	U	SMPL
N5535-02 - 1	COMP	208	0.0609	-	7470A	PPB	11 Nov 2022 12:23:49	U	SMPL
N5551-03 - 1	MW-12I	38	0.0210	-	7470A	PPB	11 Nov 2022 12:26:06	U	SMPL
N5551-03DUP - 1	MW-12IDUP	48	0.0233	-	7470A	PPB	11 Nov 2022 12:28:21	U	SMPL
CCV12 - 1	CCV12	20690	4.8733	-	7470A	PPB	11 Nov 2022 12:30:37	U	SMPL
CCB12 - 1	CCB12	-65	-0.0032	-	7470A	PPB	11 Nov 2022 12:32:52	U	SMPL
N5551-05 - 1	N551-03MS	13985	3.2979	-	7470A	PPB	11 Nov 2022 12:35:08	U	SMPL
N5551-06 - 1	N551-03MSD	14246	3.3592	-	7470A	PPB	11 Nov 2022 12:37:23	U	SMPL
N5551-13 - 1	MW-8I	437	0.1147	-	7470A	PPB	11 Nov 2022 12:39:39	U	SMPL
N5551-15 - 1	MW-21	412	0.1088	-	7470A	PPB	11 Nov 2022 12:41:54	U	SMPL
N5562-01 - 1	B-4C(OW)	89	0.0330	-	7470A	PPB	11 Nov 2022 12:44:10	U	SMPL
N5562-03 - 1	B-1(OW)	-14	0.0088	-	7470A	PPB	11 Nov 2022 12:46:26	U	SMPL
N5550-01 - 1	D2781	13	0.0151	-	7470A	PPB	11 Nov 2022 12:48:43	U	SMPL
PB148927TB - 1	PB148927TB	19	0.0165	-	7470A	PPB	11 Nov 2022 12:51:00	U	SMPL
PB148944TB - 1	PB148944TB	38	0.0210	-	7470A	PPB	11 Nov 2022 12:53:16	U	SMPL
CCV13 - 1	CCV13	20555	4.8416	-	7470A	PPB	11 Nov 2022 12:55:33	U	SMPL
CCB13 - 1	CCB13	6	0.0135	-	7470A	PPB	11 Nov 2022 12:57:49	U	SMPL
N5509-02LX5 - 1		11	0.0146	-	7470A	PPB	11 Nov 2022 13:00:05	U	SMPL
N5509-02A - 1		16008	3.7732	-	7470A	PPB	11 Nov 2022 13:02:21	U	SMPL
N5553-03LX5 - 1		-48	0.0008	-	7470A	PPB	11 Nov 2022 13:04:37	U	SMPL
N5553-03A - 1		16151	3.8068	-	7470A	PPB	11 Nov 2022 13:06:52	U	SMPL
N5551-03LX5 - 1		-38	0.0031	-	7470A	PPB	11 Nov 2022 13:09:09	U	SMPL
N5551-03A - 1		15735	3.7091	-	7470A	PPB	11 Nov 2022 13:11:25	U	SMPL
CCV14 - 1	CCV14	20479	4.8237	-	7470A	PPB	11 Nov 2022 13:13:41	U	SMPL
CCB14 - 1	CCB14	-49	0.0005	-	7470A	PPB	11 Nov 2022 13:15:58	U	SMPL