



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 11/14/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 11/11/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 11/12/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122855

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N5587-01	OF-2-20221110	1	0.96	8.99	9.95	7.42	71.9	
N5587-02	OF-3-20221110	2	0.91	8.83	9.74	6.16	59.5	
N5587-03	N5587-02MS	3	0.91	8.83	9.74	6.16	59.5	
N5587-04	N5587-02MSD	4	0.91	8.83	9.74	6.16	59.5	
N5587-05	OF-4-20221110	5	0.95	9.02	9.97	5.38	49.1	
N5587-06	DUP-02-20221110	6	0.91	9.04	9.95	6.5	61.8	
N5598-01	FRAC-TANK-0760533	21	0.91	9.05	9.96	5.66	52.5	SLUDGE SAMPLE
N5607-01	SB-01	7	0.91	9.04	9.95	8.34	82.2	
N5607-02	WC-01	8	0.93	8.76	9.69	8.21	83.1	
N5607-03	SB-36	9	0.89	8.68	9.57	8.64	89.3	
N5607-04	WC-36	10	0.96	8.74	9.7	9.00	92.0	
N5607-05	SB-37	11	0.9	8.68	9.58	8.56	88.2	
N5607-06	WC-37	12	0.91	8.75	9.66	8.85	90.7	
N5607-07	SB-38	13	0.93	9.03	9.96	9.62	96.2	
N5607-08	WC-38	14	0.85	8.96	9.81	9.00	91.0	
N5607-09	SB-39	15	0.95	9.01	9.96	8.45	83.2	
N5607-10	WC-39	16	0.89	8.72	9.61	7.92	80.6	
N5607-11	SB-40A	17	0.95	8.99	9.94	9.00	89.5	
N5607-12	WC-40	18	0.94	9.04	9.98	8.49	83.5	
N5607-13	SB-40B	19	0.93	8.84	9.77	7.07	69.5	
N5608-01	NP-1	20	0.91	8.93	9.84	9.23	93.2	
N5610-01	P-110922RB-01	22	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N5610-02	P-110922RB-02	23	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$